



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Transcat - St. Louis

**895 Bolger Court
Fenton, MO 63026**

Fulfills the requirements of

ISO/IEC 17025:2017

and the national standards

**ANSI/NCSL Z540-1-1994 (R2002) AND
ANSI/NCSL Z540.3-2006 (R2013)**

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 07 September 2025
Certificate Number: AC-2489.13



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

ANSI/NCSL Z540.3-2006 (R2013)

Transcat – St. Louis

895 Bolger Court

Fenton, MO 63026

Dennis Evans 636-349-7722

CALIBRATION

Valid to: **September 7, 2025**

Certificate Number: **AC-2489.13**

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|---|---|
| DC Current – Measure ¹ | Up to 100 μ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A | 33 μ A/A + 0.92 nA 29 μ A/A + 5.8 nA 29 μ A/A + 58 nA 46 μ A/A + 0.58 μ A 0.013 % of reading + 12 μ A | 8.5 Digit Multimeter |
| DC Current – Measure ¹ | (1 to 3) A | 0.096 % of reading + 0.47 mA | 6.5 Digit Multimeter |
| DC Current – Source ¹ | Up to 330 μ A (0.33 to 3.3) mA (3.3 to 33) mA (33 to 330) mA (0.33 to 1.1) A (1 to 3) A (3 to 11) A (11 to 20.5) A | 0.12 mA/A + 16 nA 82 μ A/A + 39 nA 98 μ A/A + 0.19 μ A 78 μ A/A + 1.9 μ A 0.16 mA/A + 31 μ A 0.3 mA/A + 31 μ A 0.51 mA/A + 0.39 mA 0.93 mA/A + 0.58 mA | Multiproduct Calibrator |
| DC Clamp-on Ammeter ¹ (Non-Toroidal Type) Hall Effect Sensor | (20 to 150) A (150 to 1 000) A | 0.51 % of reading + 0.14 A 0.51 % of reading + 0.5 A | Multiproduct Calibrator, 50-turn Coil |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-----------------------------------|---------------------------------|---|--|
| AC Current – Measure ¹ | Up to 100 μ A | | 8.5 Digit Multimeter |
| | (10 to 20) Hz | 0.46 % of reading + 35 nA | |
| | (20 to 45) Hz | 0.17 % of reading + 35 nA | |
| | (45 to 100) Hz | 0.072 % of reading + 35 nA | |
| | 100 Hz to 1 kHz | 0.072 % of reading + 35 nA | |
| | (0.1 to 1) mA | | |
| | (10 to 20) Hz | 0.46 % of reading + 0.23 μ A | |
| | (20 to 45) Hz | 0.17 % of reading + 0.23 μ A | |
| | (45 to 100) Hz | 0.07 % of reading + 0.23 μ A | |
| | 100 Hz to 5 kHz | 0.038 % of reading + 0.23 μ A | |
| | (1 to 10) mA | | |
| | (10 to 20) Hz | 0.46 % of reading + 2.3 μ A | |
| | (20 to 45) Hz | 0.17 % of reading + 2.3 μ A | |
| | (45 to 100) Hz | 0.071 % of reading + 2.3 μ A | |
| | 100 Hz to 5 kHz | 0.038 % of reading + 2.3 μ A | |
| | (10 to 100) mA | | |
| (10 to 20) Hz | 0.46 % of reading + 23 μ A | | |
| (20 to 45) Hz | 0.17 % of reading + 23 μ A | | |
| (45 to 100) Hz | 0.071 % of reading + 23 μ A | | |
| 100 Hz to 5 kHz | 0.037 % of reading + 23 μ A | | |
| 100 mA to 1 A | | | |
| (10 to 20) Hz | 0.46 % of reading + 0.23 mA | | |
| (20 to 45) Hz | 0.19 % of reading + 0.23 mA | | |
| (45 to 100) Hz | 0.097 % of reading + 0.23 mA | | |
| 100 Hz to 5 kHz | 0.12 % of reading + 0.23 mA | | |
| AC Current – Source ¹ | (29 to 330) μ A | | Multiproduct Calibrator |
| | (10 to 20) Hz | 0.16 % of reading + 80 nA | |
| | (20 to 45) Hz | 0.12 % of reading + 80 nA | |
| | 45 Hz to 1 kHz | 0.097 % of reading + 80 nA | |
| | (1 to 5) kHz | 0.23 % of reading + 0.12 μ A | |
| | (5 to 10) kHz | 0.62 % of reading + 0.16 μ A | |
| | (10 to 30) kHz | 1.2 % of reading + 0.31 μ A | |
| | 330 μ A to 3.3 mA | | |
| | (10 to 20) Hz | 0.16 % of reading + 0.12 μ A | |
| | (20 to 45) Hz | 0.097 % of reading + 0.12 μ A | |
| | 45 Hz to 1 kHz | 0.078 % of reading + 0.12 μ A | |
| | (1 to 5) kHz | 0.16 % of reading + 0.16 μ A | |
| | (5 to 10) kHz | 0.39 % of reading + 0.23 μ A | |
| | (10 to 30) kHz | 0.78 % of reading + 0.47 μ A | |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---------------------------------|---|--|
| AC Current – Source ¹ | (3.3 to 33) mA | | Multiproduct Calibrator |
| | (10 to 20) Hz | 0.14 % of reading + 1.6 μ A | |
| | (20 to 45) Hz | 0.071 % of reading + 1.6 μ A | |
| | 45 Hz to 1 kHz | 0.035 % of reading + 1.6 μ A | |
| | (1 to 5) kHz | 0.064 % of reading + 1.6 μ A | |
| | (5 to 10) kHz | 0.16 % of reading + 2.3 μ A | |
| | (10 to 30) kHz | 0.31 % of reading + 3.1 μ A | |
| | (33 to 330) mA | | |
| | (10 to 20) Hz | 0.14 % of reading + 16 μ A | |
| | (20 to 45) Hz | 0.071 % of reading + 16 μ A | |
| | 45 Hz to 1 kHz | 0.033 % of reading + 16 μ A | |
| | (1 to 5) kHz | 0.078 % of reading + 39 μ A | |
| | (5 to 10) kHz | 0.16 % of reading + 78 μ A | |
| | (10 to 30) kHz | 0.31 % of reading + 0.16 mA | |
| | 330 mA to 1.1 A | | |
| | (10 to 20) Hz | 0.14 % of reading + 78 μ A | |
| | 45 Hz to 1 kHz | 0.04 % of reading + 78 μ A | |
| | (1 to 5) kHz | 0.47 % of reading + 0.78 mA | |
| | (5 to 10) kHz | 1.9 % of reading + 3.9 mA | |
| | (1.1 to 3) A | | |
| (10 to 20) Hz | 0.14 % of reading + 78 μ A | | |
| 45 Hz to 1 kHz | 0.049 % of reading + 78 μ A | | |
| (1 to 5) kHz | 0.47 % of reading + 0.78 mA | | |
| (5 to 10) kHz | 1.9 % of reading + 3.9 mA | | |
| (3 to 11) A | | | |
| (10 to 100) Hz | 0.049 % of reading + 1.6 mA | | |
| 100 Hz to 1 kHz | 0.079 % of reading + 1.6 mA | | |
| (1 to 5) kHz | 2.3 % of reading + 1.6 mA | | |
| (11 to 20.5) A | | | |
| (10 to 100) Hz | 0.095 % of reading + 3.9 mA | | |
| 100 Hz to 1 kHz | 0.12 % of reading + 3.9 mA | | |
| (1 to 5) kHz | 2.3 % of reading + 3.9 mA | | |
| AC Clamp-on Ammeters ¹ (Toroidal Type) Transformer Type Sensor | (20 to 150) A | | Multiproduct Calibrator, 50-turn Coil |
| | (45 to 65) Hz | 0.3 % of reading + 26 mA | |
| | (65 to 440) Hz | 0.83 % of reading + 47 mA | |
| | (150 to 1 000) A | | |
| | (45 to 65) Hz | 0.35 % of reading + 0.12 A | |
| | (65 to 440) Hz | 1.1 % of reading + 0.22 A | |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|---|--|
| AC Clamp-on Ammeters ¹ (Non-Toroidal Type) Hall Effect Sensor | (20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz | 0.57 % of reading + 0.25 A 1 % of reading + 0.25 A 0.6 % of reading + 0.9 A 1.3 % of reading + 0.92 A | Multiproduct Calibrator, 50-turn Coil |
| DC Resistance – Source/Measure ¹ (Variable Artifact) | Up to 10 Ω (10 to 100) Ω (0.1 to 1) kΩ (1 to 10) kΩ (10 to 100) kΩ (0.1 to 1) MΩ (1 to 10) MΩ (10 to 100) MΩ (0.1 to 1) GΩ | 18 μΩ/Ω + 58 μΩ 15 μΩ/Ω + 0.58 mΩ 13 μΩ/Ω + 0.58 mΩ 12 μΩ/Ω + 5.8 mΩ 13 μΩ/Ω + 58 mΩ 21 μΩ/Ω + 2.3 Ω 62 μΩ/Ω + 0.12 kΩ 0.059 % of reading + 1.2 kΩ 0.82 % of reading + 12 kΩ | 8.5 Digit Multimeter, Decade Resistor |
| DC Resistance – Source ¹ (Variable Artifact) | (100 to 1 000) kΩ (1 to 10) MΩ (10 to 100) MΩ (100 to 1 000) MΩ (1 to 10) GΩ (10 to 100) GΩ (100 to 1 000) GΩ | 0.037% of reading 0.037 % of reading + 1.2 μΩ/Ω/V 0.12 % of reading + 1.2 μΩ/Ω/V 0.23 % of reading + 1.2 μΩ/Ω/V 0.59 % of reading + 1.2 μΩ/Ω/V 1.2 % of reading + 1.2 μΩ/Ω/V 1.2 % of reading + 1.2 μΩ/Ω/V | High Accuracy Decade Resistor |
| DC Voltage – Measure ¹ | Up to 100 mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 500) V (500 to 800) V (800 to 1 000) V | 8.3 μV/V + 0.58 μV 5.3 μV/V + 0.58 μV 5.3 μV/V + 0.58 μV 7.7 μV/V + 35 μV 15 μV/V + 0.12 mV 18 μV/V + 0.12 mV 21 μV/V + 0.12 mV | 8.5 Digit Multimeter |
| DC High Voltage – Measure ¹ | (1 to 10) kV (10 to 20) kV (20 to 30) kV (30 to 40) kV (40 to 50) kV (50 to 60) kV (60 to 70) kV (70 to 80) kV (80 to 90) kV (90 to 100) kV | 0.039 % of reading + 92 mV 0.038 % of reading + 2.4 V 0.041 % of reading + 2.4 V 0.047 % of reading + 2.4 V 0.056 % of reading + 2.4 V 0.071 % of reading + 2.4 V 0.089 % of reading + 2.4 V 0.12 % of reading + 2.5 V 0.15 % of reading + 2.5 V 0.17 % of reading + 2.5 V | Digital HV Meter, Associated High Voltage Probes |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|-----------------------------------|---|---|--|
| DC Voltage – Source ¹ | (0 to 330) mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (300 to 1 000) V | 16 μV/V + 0.78 μV 9 μV/V + 1.6 μV 10 μV/V + 16 μV 15 μV/V + 0.12 mV 14 μV/V + 1.2 mV | Multiproduct Calibrator |
| AC Voltage – Measure ¹ | Up to 10 mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz 1 MHz to 4 MHz (10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (0.1 to 1) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz | 0.04 % of reading + 3.5 μV 0.03 % of reading + 1.2 μV 0.04 % of reading + 1.2 μV 0.15 % of reading + 1.2 μV 0.59 % of reading + 1.2 μV 4.6 % of reading + 2.3 μV 1.5 % of reading + 5.8 μV 8.1 % of reading + 8.1 μV 0.013 % of reading + 4.6 μV 0.009 7 % of reading + 2.3 μV 0.017 % of reading + 2.3 μV 0.038 % of reading + 2.3 μV 0.093 % of reading + 2.3 μV 0.36 % of reading + 12 μV 1.2 % of reading + 12 μV 1.8 % of reading + 12 μV 4.7 % of reading + 81 μV 4.7 % of reading + 92 μV 17 % of reading + 0.12 mV 0.008 8 % of reading + 46 μV 0.008 3 % of reading + 23 μV 0.017 % of reading + 23 μV 0.036 % of reading + 23 μV 0.093 % of reading + 23 μV 0.35 % of reading + 0.12 mV 1.2 % of reading + 0.12 mV 1.8 % of reading + 0.12 mV 4.6 % of reading + 0.81 mV 4.6 % of reading + 0.92 mV 17 % of reading + 1.2 mV | 8.5 Digit Multimeter |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---|--|--|
| AC Voltage – Measure ¹ | (1 to 10) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (10 to 100) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (100 to 700) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz | 0.009 5 % of reading + 0.46 mV 0.023 % of reading + 0.23 mV 0.017 % of reading + 0.23 mV 0.036 % of reading + 0.23 mV 0.093 % of reading + 0.23 mV 0.35 % of reading + 1.2 mV 1.2 % of reading + 1.2 mV 1.8 % of reading + 1.2 mV 4.6 % of reading + 8.1 mV 4.6 % of reading + 9.2 mV 17 % of reading + 12 mV 0.024 % of reading + 4.6 mV 0.024 % of reading + 2.3 mV 0.024 % of reading + 2.3 mV 0.041 % of reading + 2.3 mV 0.14 % of reading + 2.3 mV 0.46 % of reading + 12 mV 1.7 % of reading + 12 mV 0.048 % of reading + 46 mV 0.048 % of reading + 23 mV 0.071 % of reading + 23 mV 0.19 % of reading + 23 mV 0.35 % of reading + 23 mV | 8.5 Digit Multimeter |
| AC High Voltage – Measure ¹ | (0.7 to 5) kV 10 mHz to 10 Hz (10 to 30) Hz (30 to 50) Hz (50 to 70) Hz (70 to 100) Hz (100 to 200) Hz (200 to 450) Hz (450 to 600) Hz | 0.14 % of reading + 0.17 V 0.12 % of reading + 0.29 V 0.099 % of reading + 0.37 V 0.068 % of reading + 0.37 V 0.099 % of reading + 0.37 V 0.099 % of reading + 0.37 V 0.48 % of reading + 0.17 V 0.47 % of reading + 0.17 V | Digital HV Meter, Associated High Voltage Probes |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|----------------------------|---|--|
| AC High Voltage – Measure ¹ | (5 to 30) kV | | Digital HV Meter, Associated High Voltage Probes |
| | 10 mHz to 10 Hz | 0.19 % of reading + 2.4 V | |
| | (10 to 30) Hz | 0.13 % of reading + 2.4 V | |
| | (30 to 50) Hz | 0.11 % of reading + 2.4 V | |
| | (50 to 70) Hz | 0.077 % of reading + 2.4 V | |
| | (70 to 100) Hz | 0.11 % of reading + 2.4 V | |
| | (100 to 200) Hz | 0.11 % of reading + 2.4 V | |
| | (200 to 450) Hz | 0.7 % of reading + 2.4 V | |
| | (450 to 600) Hz | 1.4 % of reading + 2.4 V | |
| | (30 to 50) kV | | |
| | 10 mHz to 10 Hz | 0.24 % of reading + 2.5 V | |
| | (10 to 30) Hz | 0.18 % of reading + 2.5 V | |
| | (30 to 50) Hz | 0.13 % of reading + 2.5 V | |
| | (50 to 70) Hz | 0.1 % of reading + 2.5 V | |
| | (70 to 100) Hz | 0.13 % of reading + 2.5 V | |
| | (100 to 200) Hz | 0.69 % of reading + 2.5 V | |
| | (200 to 450) Hz | 2.9 % of reading + 2.5 V | |
| | (50 to 70) kV | | |
| 10 mHz to 10 Hz | 0.37 % of reading + 2.6 V | | |
| (10 to 30) Hz | 0.26 % of reading + 2.6 V | | |
| (30 to 50) Hz | 0.16 % of reading + 2.6 V | | |
| (50 to 70) Hz | 0.16 % of reading + 2.6 V | | |
| (70 to 100) Hz | 1.2 % of reading + 2.6 V | | |
| (100 to 200) Hz | 1.2 % of reading + 2.6 V | | |
| (200 to 450) Hz | 17 % of reading + 2.6 V | | |
| AC Voltage – Source ¹ | (1 to 33) mV | | Multiproduct Calibrator |
| | 10 Hz to 45 Hz | 0.065 % of reading + 4.7 μV | |
| | 45 Hz to 10 kHz | 0.014 % of reading + 4.7 μV | |
| | (10 to 20) kHz | 0.017 % of reading + 4.7 μV | |
| | (20 to 50) kHz | 0.079 % of reading + 4.7 μV | |
| | (50 to 100) kHz | 0.27 % of reading + 9.3 μV | |
| | (100 to 500) kHz | 0.62 % of reading + 39 μV | |
| | (33 to 330) mV | | |
| | 10 Hz to 45 Hz | 0.027 % of reading + 6.2 μV | |
| | 45 Hz to 10 kHz | 0.012 % of reading + 6.2 μV | |
| | (10 to 20) kHz | 0.013 % of reading + 6.2 μV | |
| | (20 to 50) kHz | 0.027 % of reading + 6.2 μV | |
| (50 to 100) kHz | 0.062 % of reading + 25 μV | | |
| (100 to 500) kHz | 0.16 % of reading + 54 μV | | |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|-----------------------------|---|--|
| AC Voltage – Source ¹ | (0.33 to 3.3) V | | Multiproduct Calibrator |
| | 10 Hz to 45 Hz | 0.027 % of reading + 39 μV | |
| | 45 Hz to 10 kHz | 0.012 % of reading + 47 μV | |
| | (10 to 20) kHz | 0.015 % of reading + 47 μV | |
| | (20 to 50) kHz | 0.024 % of reading + 39 μV | |
| | (50 to 100) kHz | 0.055 % of reading + 49 μV | |
| | (100 to 500) kHz | 0.19 % of reading + 0.47 mV | |
| | (3.3 to 33) V | | |
| | 10 Hz to 45 Hz | 0.027 % of reading + 0.50 mV | |
| | 45 Hz to 10 kHz | 0.012 % of reading + 0.47 mV | |
| | (10 to 20) kHz | 0.019 % of reading + 0.47 mV | |
| | (20 to 50) kHz | 0.027 % of reading + 0.47 mV | |
| | (50 to 100) kHz | 0.7 % of reading + 1.2 mV | |
| | (33 to 330) V | | |
| | 45 Hz to 1 kHz | 0.015 % of reading + 1.6 mV | |
| 1 kHz to 10 kHz | 0.016 % of reading + 4.7 mV | | |
| (10 to 20) kHz | 0.020 % of reading + 4.7 mV | | |
| (20 to 50) kHz | 0.025 % of reading + 4.7 mV | | |
| (50 to 100) kHz | 0.16 % of reading + 39 mV | | |
| (330 to 1020) V | | | |
| 45 Hz to 1 kHz | 0.023 % of reading + 7.8 mV | | |
| 1 kHz to 5 kHz | 0.02 % of reading + 7.8 mV | | |
| (5 to 10) kHz | 0.023 % of reading + 7.8 mV | | |
| Capacitance – Source ¹ (Simulation) | 10 Hz to 10 kHz | | Multiproduct Calibrator |
| | (0.22 to 0.399 9) nF | 0.39 % of reading + 7.8 pF | |
| | (0.4 to 1.09 99) nF | 0.39 % of reading + 7.8 pF | |
| | 10 Hz to 3 kHz | | |
| | (1.1 to 3.299 9) nF | 0.39 % of reading + 7.8 pF | |
| | 10 Hz to 1 kHz | | |
| | (3.3 to 10.999 9) nF | 0.21 % of reading + 7.8 pF | |
| | (11 to 32.999 9) nF | 0.21 % of reading + 78 pF | |
| | (33 to 109.999) nF | 0.21 % of reading + 78 pF | |
| | (110 to 329.999) nF | 0.21 % of reading + 0.23 nF | |
| | 10 Hz to 600 Hz | | |
| (0.33 to 1.099 9) μF | 0.21 % of reading + 0.78 nF | | |
| 10 Hz to 300 Hz | | | |
| (1.1 to 3.299 99) μF | 0.21 % of reading + 2.3 nF | | |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|--|--|
| Capacitance – Source ¹ (Simulation) | 10 Hz to 150 Hz (3.3 to 10.999 9) μ F 10 Hz to 120 Hz (11 to 32.999 9) μ F 10 Hz to 80 Hz (33 to 109.999) μ F DC to 50 Hz (110 to 329.999) μ F DC to 20 Hz (0.33 to 10.999 9) mF DC to 6 Hz (1.1 to 3.299 99) mF DC to 2 Hz (3.3 to 10.999 9) mF DC to 0.6 Hz (11 to 32.999 9) mF DC to 0.2 Hz (33 to 110) mF | 0.21 % of reading + 7.8 nF 0.32 % of reading + 23 nF 0.35 % of reading + 78 nF 0.35 % of reading + 0.23 μ F 0.35 % of reading + 0.78 μ F 0.35 % of reading + 2.3 μ F 0.35 % of reading + 7.8 μ F 0.58 % of reading + 23 μ F 0.85 % of reading + 78 μ F | Multiproduct Calibrator |
| Capacitance – Measure ¹ (1 kHz) | 1 pF to 1 nF (1 to 10) nF (10 to 100) nF (0.1 to 1) μ F (1 to 10) μ F (10 to 100) μ F (0.1 to 1) mF (1 to 10) mF (10 to 100) mF | 1.9% of reading + 19 pF 0.83 % of reading + 39 pF 0.83 % of Reading + 0.39 nF 0.83 % of reading + 3.9 nF 0.82 % of Reading + 39 nF 0.9 % of reading + 0.39 μ F 0.89 % of reading + 3.9 μ F 0.89 % of reading + 39 μ F 3.2 % of reading + 0.16 mF | 6.5 Digit Multimeter |
| Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹ | Type B (250 to 350) $^{\circ}$ C (350 to 445) $^{\circ}$ C (445 to 580) $^{\circ}$ C (580 to 750) $^{\circ}$ C (750 to 1 000) $^{\circ}$ C (1 000 to 1 820) $^{\circ}$ C | 1.2 $^{\circ}$ C 0.9 $^{\circ}$ C 0.71 $^{\circ}$ C 0.55 $^{\circ}$ C 0.45 $^{\circ}$ C 0.35 $^{\circ}$ C | Ectron 1140A Thermocouple Calibrator/Simulator |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---------------------|---|--|
| Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹ | Type E | | Ectron 1140A Thermocouple Calibrator/Simulator |
| | (-270 to -245) °C | 1.6 °C | |
| | (-245 to -195) °C | 0.24 °C | |
| | (-195 to -155) °C | 0.12 °C | |
| | (-155 to -90) °C | 0.095 °C | |
| | (-90 to 0) °C | 0.08 °C | |
| | (0 to 15) °C | 0.076 °C | |
| | (15 to 890) °C | 0.064 °C | |
| | (890 to 1 000) °C | 0.074 °C | |
| | Type J | | |
| | (-210 to -180) °C | 0.15 °C | |
| | (-180 to -120) °C | 0.12 °C | |
| | (-120 to -50) °C | 0.093 °C | |
| | (-50 to 990) °C | 0.08 °C | |
| | (990 to 1 200) °C | 0.094 °C | |
| | Type K | | |
| | (-270 to -255) °C | 2.5 °C | |
| | (-255 to -195) °C | 0.85 °C | |
| | (-195 to -115) °C | 0.16 °C | |
| | (-115 to -55) °C | 0.12 °C | |
| | (-55 to 1 000) °C | 0.087 °C | |
| | (1 000 to 1 372) °C | 0.096 °C | |
| | Type N | | |
| | (-270 to -260) °C | 5.4 °C | |
| (-260 to -200) °C | 1.5 °C | | |
| (-200 to -140) °C | 0.29 °C | | |
| (-140 to -70) °C | 0.18 °C | | |
| (-70 to 25) °C | 0.14 °C | | |
| (25 to 160) °C | 0.12 °C | | |
| (160 to 1 300) °C | 0.11 °C | | |
| Type R | | | |
| (-50 to -30) °C | 0.8 °C | | |
| (-30 to 45) °C | 0.69 °C | | |
| (45 to 160) °C | 0.49 °C | | |
| (160 to 380) °C | 0.35 °C | | |
| (380 to 775) °C | 0.3 °C | | |
| (775 to 1 768) °C | 0.26 °C | | |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment | |
|--|--|---|--|---|
| Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹ | Type S | | Ectron 1140A Thermocouple Calibrator/Simulator | |
| | (-50 to -30) °C | 0.76 °C | | |
| | (-30 to 45) °C | 0.68 °C | | |
| | (45 to 105) °C | 0.49 °C | | |
| | (105 to 310) °C | 0.41 °C | | |
| | (310 to 615) °C | 0.35 °C | | |
| | (615 to 1 768) °C | 0.31 °C | | |
| | Type T | | | |
| | (-270 to -255) °C | 1.9 °C | | |
| | (-255 to -240) °C | 0.6 °C | | |
| | (-240 to -210) °C | 0.36 °C | | |
| | (-210 to -150) °C | 0.22 °C | | |
| | (-150 to -40) °C | 0.15 °C | | |
| (-40 to 100) °C | 0.095 °C | | | |
| (100 to 400) °C | 0.08 °C | | | |
| Oscilloscopes ^{1,2} | Amplitude – DC Voltage into 50 Ω load into 1 MΩ load | (-6 to 6) V | 0.22 % of reading + 31 μV | Multiproduct Calibrator with 1.1 GHz Scope Option |
| | | (-130 to 130) V | 0.12 % of reading + 31 μV | |
| | Amplitude – Square Wave into 50 Ω load | 10 Hz to 10 kHz 1 mV p-p to 6.6 Vp-p | 22 % of reading + 31 μV | |
| | | into 1 MΩ load | 10 Hz to 1 kHz 1 mV p-p to 130 Vp-p | |
| | (1 kHz to 10) kHz 1 mV p-p to 130 Vp-p | | 0.22 % of reading + 31 μV | |
| | Time Markers into 50 Ω load | 1 ns to 20 ms | 0.000 22 % of reading | |
| | | 50 ms | 0.005 9 % of reading | |
| | | 0.1 s | 0.009 8 % of reading | |
| | | 0.2 s | 0.018 % of reading | |
| | | 0.5 s | 0.041 % of reading | |
| 1 s | | 0.08 % of reading | | |
| 2 s | | 0.16 % of reading | | |
| 5 s | | 0.39 % of reading | | |



ANSI National Accreditation Board

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|---|--|--|
| Oscilloscopes ^{1,2} Rise Time – Source into 50 Ω load Rate: 1 kHz to 2 MHz Rate: 2 MHz to 10 MHz | 5 mVp-p to 2.5 Vp-p (200 to 300) ps (250 to 350) ps | 50 ps 50 ps | Multiproduct Calibrator with 1.1 GHz Scope Option |
| Leveled Sine Wave into 50 Ω load | 5 mVp-p to 5 Vp-p 50 kHz 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mVp-p to 3.5 Vp-p 600 MHz to 1.1 GHz | 1.8 % of reading + 0.23 mV 2.8 % of reading + 0.23 mV 3.2 % of reading + 0.23 mV 4.7 % of reading + 0.23 mV 5.5 % of reading + 0.23 mV | |
| Bandwidth/Flatness (50 kHz Reference) into 50 Ω load | 5 mVp-p to 5.5 Vp-p 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mVp-p to 3.5 Vp-p (600 to 1 100) MHz | 1.4 % of reading + 78 μV 1.8 % of reading + 78 μV 3.2 % of reading + 78 μV 3.9 % of reading + 78 μV | |
| Input Impedance – Measure | (40 to 60) Ω 500 kΩ to 1.5 MΩ | 0.082 % of reading 0.081 % of reading | |
| Input Capacitance – Measure | (5 to 50) pF | 3.9 % of reading + 0.39 pF | |
| Waveform Generator (Sine, Square, Triangle) Amplitude into 50 Ω load into 1 MΩ load | 10 Hz to 10 kHz 1.8 mVp-p to 2.5 Vp-p 1.8 mVp-p to 55 Vp-p | 2.3 % of reading + 78 μV 2.3 % of reading + 78 μV | |
| Frequency | 10 Hz to 10 kHz | 0.001 9 % of reading + 12 mHz | |

Electrical – DC/Low Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--|---|--|
| LF Phase – Source ¹ | Up to 180 ° (10 to 65) Hz (65 to 500) Hz 500 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 20) kHz | 0.11 ° 0.21 ° 0.39 ° 1.9 ° 3.9 ° 7.8 ° | Multiproduct Calibrator |
| DC Power – Source ¹ (0.33 to 330) mA (0.33 to 3) A (3 to 20.5) A | (11 to 330) μW 11 W to 3 kW 99 mW to 20.9 kW | 0.018 % of reading 0.017 % of reading 0.054 % of reading | Multiproduct Calibrator |
| AC Power – Source ^{1,3} PF = 1 3.3 mA to 3 A 3.3 mA to 20.5 A 33 mA to 3 A 33 mA to 20.5 A (3 to 20.5) A | (10 to 45) Hz 0.11 mW to 99 W (45 to 65) Hz 0.11 mW to 20.9 kW (65 to 500) Hz 11 mW to 3.06 kW 500 Hz to 1 kHz 11 mW to 20.9 kW (65 to 500) Hz 9.9 W to 20.9 kW | 0.18 % of reading 0.14 % of reading 0.16 % of reading 0.16 % of reading 0.16 % of reading | Multiproduct Calibrator |

Length – Dimensional Metrology

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|---|--------------------------------|---|---|
| Micrometers and Calipers ^{1,4} (Outside, Inside, Depth, Step) | (0.05 to 1) in (1 to 12) in | (13 + 1L) μin (7 + 5L) μin | Gage Blocks, Long Gage Blocks |
| Anvil Flatness ¹ | Up to 1 in | 6.3 μin | Optical Flats |
| Indicators ^{1,4} (Dial and Digital) | Up to 1 in (1 to 6) in | (10 + 2L) μin (4 + 10L) μin | Gage Blocks |
| Distance Measuring Equipment ⁴ | Up to 99 999 ft | (0.05 + 0.000 5D) ft | Cylinder with Square Ends with Incremental Counter |

Mass and Mass Related

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|--|---|---|
| Balances and Scales ^{1,5} | Up to 5 g (5 to 10) g (10 to 20) g (20 to 50) g (50 to 100) g (100 to 200) g (200 to 500) g 500 g to 1 kg (1 to 2) kg (2 to 3) kg | 40 µg 59 µg 89 µg 0.15 mg 0.31 mg 0.9 mg 1.5 mg 3.1 mg 4.4 mg 4.8 mg | ASTM E617 Class 1 weights and internal calibration procedure utilized for the calibration of the weighing system. |
| Torque – Measure ¹ (Dial, Digital, Click Wrenches) | (4 to 500) lbf·in (30 to 400) lbf·in (80 to 1 000) lbf·in (20 to 250) lbf·ft (60 to 600) lbf·ft | 0.61 % of reading 0.61 % of reading 0.61 % of reading 0.61 % of reading 0.71 % of reading | Torque Calibration System |
| Absolute Pressure Measuring Devices ¹ (Pneumatic) | (0 to 14.7) psia (14.7 to 39.7) psia (39.7 to 514.7) psia | 0.002 5 psi 0.000 71 % of reading + 0.002 3 psi 0.006 5 % of reading | Pneumatic Pressure Controller/Calibrator |
| Gauge Pressure Measuring Devices ¹ (Pneumatic) | (-14.2 to < 0) psig (> 0 to 25) psig (25 to 500) psig (-36 to -22) inH ₂ O (-22 to 22) inH ₂ O (22 to 60) inH ₂ O (60 to 72) inH ₂ O (72 to 804) inH ₂ O | 0.000 64 % of reading + 0.001 3 psi 0.001 4 % of reading + 0.001 3 psi 0.006 7 % of reading 0.009 % of reading + 0.000 15 inH ₂ O 0.002 inH ₂ O 0.009 % of reading + 0.000 15 inH ₂ O 0.006 5 inH ₂ O 0.009 % of reading + 0.000 15 inH ₂ O | Pneumatic Pressure Controller/Calibrator |
| Gauge Pressure Measuring Devices, Pressure Controllers/Calibrators ¹ (Hydraulic) | (5 to 150) psig (150 to 1 500) psig (50 to 1 500) psig (1 500 to 15 000) psig | 0.032 psi 0.008 % of reading + 0.02 psi 0.14 psi 0.008 % of reading + 0.02 psi | Deadweight Tester |

Thermodynamic

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---|---|--|
| Drywell Calibrators, Liquid Baths ¹ | (-195 to 0) °C (0 to 160) °C (160 to 420) °C (420 to 660) °C | 0.015 °C 0.015 °C 0.02 °C 0.032 °C | SPRT, Digital Temperature Indicator |

Thermodynamic

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|--|---|---|--|
| Digital/Mechanical Thermometers, RTD Probes, Thermocouple Probes, Thermistors ¹ | (-30 to -20) °C | 0.041 °C | Micro-Bath, SPRT, Digital Temperature Indicator |
| | (-20 to 25) °C (25 to 150) °C | 0.018 °C 0.021 °C | Liquid Bath, SPRT, Digital Temperature Indicator |
| | (150 to 160) °C (160 to 300) °C (300 to 600) °C | 0.061 °C 0.14 °C 0.52 °C | Dry-well, SPRT, Digital Temperature Indicator |

Time and Frequency

| Parameter/Equipment | Range | Expanded Uncertainty of Measurement (+/-) | Reference Standard, Method, and/or Equipment |
|----------------------------------|--|--|---|
| Frequency – Source/Measure | 10 MHz | 5.9 nHz/Hz | Rubidium Frequency Standard |
| Frequency – Measure ¹ | 1 to 10 kHz 10 kHz to 10 MHz 10 MHz to 225 MHz | 640 pHz/Hz + 4.5 µHz 640 pHz/Hz + 5.0 µHz 640 pHz/Hz | Frequency Counter, Rubidium Frequency Standard |
| Frequency – Source ¹ | 1 Hz to 20 MHz | 58 nHz/Hz | Function Generator, Rubidium Frequency Standard |

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

- Notes:
1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
 2. The stated uncertainty is the laboratory's ability to source a fast rise pulse that is approximately 250 ps. In the typical application of measuring rise time of an oscilloscope, this value is one of the contributing factors, but other factors are derived from the DUT.
 3. The uncertainties shown are for the most favorable conditions. There is an increase in uncertainty that corresponds to the laboratory's AC voltage and current uncertainties at different frequencies other than the ones shown. Power factors (PF) other than the one shown contribute to the power uncertainty. PF is related to the cosine of phase. Therefore, uncertainties track the laboratory's phase uncertainty closely at PF near one but are magnified heavily as PF approaches zero. The lab may also report reactive power, apparent power, and power factor under this accreditation. If needed, contact the laboratory for more information regarding uncertainties at frequency and power factor combinations other than the ones shown.
 4. L = length in inches; D = Distance in feet.
 5. The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The CMC presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
 6. The legal entity for this client is Transcat, Inc.
 7. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2489.13.



Jason Stine, Vice President

