

# Manual Supplement

Manual Title:	113 Calibration Information	Supplement Issue:	<b>3</b>
Print Date:	April 2008	Issue Date:	2/14
Revision/Date:	1, 6/11	Page Count:	1


---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1, 63065, 63359


On page 3, in the **Symbols** Table add:


	Conforms to relevant South Korean EMC Standards
CAT II	Measurement Category II is applicable to test and measuring circuits connected directly to utilization points (socket outlets and similar points) of the low-voltage MAINS installation.
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
CAT IV	Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.

In the **General Specifications**, delete **Shock**, **Safety Compliance**, **EMI Regulations**, and **Certifications** and add:

**IP Rating** ..... IEC 60529: IP42  
**Safety** ..... IEC 61010-1: 600 V CAT IV, Pollution Degree 2  
**Electromagnetic Environment** ..... IEC 61326-1: Portable  
**Electromagnetic Compatibility** ..... Applies to use in Korea only: Class A Equipment (Industrial Broadcasting & Communication Equipment) <sup>[1]</sup>  
 [1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.

## Change #2, 49160, 67402

On page 4, under the **Accuracy Specifications** table, under **Function** ➔, add footnote [4]. In  Check, add footnote [5]:

- [4] Range specified above 110 counts.  
 [5] Minimum measureable input in  Chek autorange is 3 V dc. For measurements below 3 V dc, push Range button twice to get to 6 V dc manual range.