

JCB\_F\_12.02 2012-02



## CERTIFICATE

No. Z1 13 06 14544 097

**Holder of Certificate:** Fluke Corporation

> 6920 Seaway Boulevard Everett, WA 98203-9090

USA

Production Facility(ies):

56588

**Certification Mark:** 



**Product:** 

**Multimeters** 

(Insulation Resistance Multimeter)

Model(s):

1503, 1507

Parameters:

Rated Input Voltage:

6VDC (4 x 1.5V AA batteries IEC LR6)

Measurement Category:

600 V, CAT III

**Protection Class:** 

**IP40** 

Degree of Protection: Temperature, Ambient:

55°C

Tested according to:

EN 61010-1:2010 EN 61010-2-030:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate

to third parties. See also notes overleaf.

Test report no.:

PI1305384109-000

Date, 2013-06-24

Page 1 of 1





