

# Manual Supplement

Manual Title:	190 Series II Service	Supplement Issue:	1
Print Date:	February 2013	Issue Date:	4/15
Revision/Date:	2, 2/15	Page Count:	1

---

This supplement contains information necessary to ensure the accuracy of the above manual.

**Change #1, 134**

On page 4-19, replace Table 4-7 with:

**Input A, B, C, D Frequency Measurement Accuracy Test**

Model	Time base	5502A-SC... MODE	Voltage	Frequency	Input A, B Reading
All	20 ms/div	wavegen, sine	600 mVpp	16 Hz	15.90 to 16.10
190-062	20 ns/div	levsine	600 mVpp	60 MHz	59.68 to 60.32
190-104 190-102	20 ns/div	levsine	600 mVpp	100 MHz	99.3 to 100.7
190-204 190-202	20 ns/div	levsine	600 mVpp	200 MHz	198.8 to 201.2
190-502 190-504 <sup>[1]</sup>	10 ns/div	levsine	600 mVpp	500 MHz	497.3 to 502.7
[1] Due to the decreased sampling rate with more than 2 inputs on, the 500 MHz check of a 190-504 must be done with only 2 channels on. Check with A & B on, C & D off, then with A & B off, C & D on.					