



# ANSI-ASQ National Accreditation Board

## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

### United Scale & Engineering Corporation<sup>3</sup> A TRASCAT COMPANY

16725 W. Victor Road, New Berlin, WI 53151  
4123 Terminal Dr., McFarland, WI 53558  
1322 Russett Court, Green Bay, WI 54313  
Dan Christianson Phone: 800-236-1733  
www.unitedscale.com dchristi@unitedscale.com

#### CALIBRATION

Valid to: June 19, 2017

Certificate Number: AC-1148

#### I. Mechanical


PARAMETER / EQUIPMENT	RANGE	CALIBRATION AND MEASUREMENT CAPABILITY [EXPRESSED AS UNCERTAINTY(±)]	REFERENCE STANDARD OR EQUIPMENT	METHOD(S)
Class I Balances	Up to 100 g Up to 230 g Up to 610 g	0.11 mg 0.20 mg 0.52 mg	ASTM Class F1 weights	OEM Manuals
Class I Balances	Up to 100 g Up to 230 g Up to 610 g	0.15 mg 0.29 mg 0.72 mg	ASTM Class 0 weights	OEM Manuals
Class II Balances	Up to 610 g Up to 6 100 g	1.4 mg 12.5 mg	ASTM Class F1 weights	OEM Manuals
Class II Balances	Up to 32 kg Up to 34 kg	0.21 g 0.60 g	ASTM Class 2 weights	OEM Manuals
Class II Balances	Up to 6 400 g Up to 32 kg Up to 34 kg Up to 64 kg Up to 100 kg Up to 200 kg	0.007 g 0.12 g 0.65 g 7.1 g 12 g 24 g	ASTM Class F weights	OEM Manuals



PARAMETER / EQUIPMENT	RANGE	CALIBRATION AND MEASUREMENT CAPABILITY [EXPRESSED AS UNCERTAINTY(±)]	REFERENCE STANDARD OR EQUIPMENT	METHOD(S)
Class III Light capacity Scales	Up to 2 lb Up to 5 lb Up to 10 lb Up to 20 lb Up to 50 lb Up to 100 lb Up to 200 lb	0.00062 lb 0.0016 lb 0.0033 lb 0.0062 lb 0.013 lb 0.026 lb 0.062 lb	ASTM Class F weights	NIST Handbook 44
Class III Medium Capacity Scales	Up to 500 lb Up to 1 000 lb Up to 2 000 lb Up to 5 000 lb Up to 10 000 lb Up to 20 000 lb	0.13 lb 0.26 lb 0.62 lb 1.2 lb 2.3 lb 5.8 lb	ASTM Class F weights	NIST Handbook 44
Class III Heavy Capacity Scales	Up to 50 000 lb Up to 100 000 lb Up to 200 000 lb	12 lb 23 lb 23 lb	ASTM Class F weights	NIST Handbook 44

*Notes:*

1. Calibration and Measurement Uncertainties (Expanded Uncertainty) are based on approximately a 95% confidence interval, using a coverage of  $k=2$ .
2. The uncertainty associated when calibrating a balance/scale is dependent on local conditions, such as the resolution of the unit being calibrated and the environment in which the balance/scale is operating. The uncertainty listed in the scope here represents the best uncertainty for a balance/scale which the organization typically calibrates in its lab. Since field (on-site) conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected in the field (on-site) than what is reported on the accredited scope.
3. This organization maintains satellite organization(s). The accredited corporate site with the above address is also accredited for satellite site(s).
4. This organization maintains the following satellite sites: 4123 Terminal Dr., McFarland, WI 53558; 1322 Russett Court, Green Bay, WI 54313.
5. This scope is formatted as part of a single document including the Certificate of Accreditation No. AC-1148.



Vice-President

