



# Certificate of Compliance

**Certificate:** 1161956

**Master Contract:** 162029

**Project:** 2589519

**Date Issued:** December 29, 2012

**Issued to:** Fluke Corporation

6920 Seaway Blvd  
Everett, WA 98203  
USA

Attention: Gary L. Allen

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Paul Tanaka*

**Issued by:** Paul Tanaka, ASCT.

## **PRODUCTS**

**CLASS 3631 06** - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

**CLASS 3631 86** - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Digital Multimeters, 77 IV, 175, 177, 179; battery powered, 9V Alkaline, IEC 6LR61.

## **Notes:**

The above models are Equipment Class 2, Pollution Degree 2, Measurement Category: CAT III 1000V, CAT IV 600V.

## **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-30-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-030 Particular requirements for testing and measuring circuits

UL Std. No. 61010-1 (3rd ed.) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



**Certificate:** 1161956

**Master Contract:** 162029

**Project:** 2589519

**Date Issued:** December 29, 2012

---

UL Std. No. 61010-2-30 (3rd ed.) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-030 Particular requirements for testing and measuring circuits

### **Reference Standards**

IEC 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

IEC 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2-030 Particular requirements for testing and measuring circuits