



Certificate of Compliance

Certificate: 2022012 (LR 44340C)

Master Contract: 162029

Project: 2594418

Date Issued: March 22, 2013

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA
Attention: Mike Meisner

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cindy Johnson

Issued by: Cindy Johnson, ASCT.

PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Digital Multimeter, Models 113, 114, 115, 115C, 116, 116C, 117 and 117C; internally powered by 9 V dc Alkaline NEDA 1604A/IEC 6LR61.

Note: The 11x Family are Equipment Class II, Pollution Degree 2, Measurement Category 600 V CAT III, except the 113 which is rated 600V CAT IV. The 115, and 117 are also rated for 10 A ac/dc.



Certificate: 2022012 (LR 44340C)

Master Contract: 162029

Project: 2594418

Date Issued: March 22, 2013

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

UL Std. No. 61010-1 (3rd ed) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st ed) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

Reference Standards

IEC 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

IEC 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits