



# CERTIFICATE

No. Z1 12 04 14544 079

**Holder of Certificate:** Fluke Corporation

6920 Seaway Boulevard  
Everett, WA 98203-9090  
USA

**Production  
Facility(ies):**

14544, 56588

**Certification Mark:**



**Product:**

**Multimeters  
( Handheld Digital Multimeter )**

**Model(s):**

**28 II EX**

**Parameters:**

Rated Input Voltage:  
4.5 VDC (3 x 1.5V AAA batteries IECLR03)  
Measurement Category: 600V CAT IV / 1000V CAT III  
Measuring Current: 10A max  
Measuring Voltage: 1000V max  
Protection Class: II  
IP Protection: IP67  
Temperature Ambient: -15 to +50°C

**Tested according to:** EN 61010-1:2001

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:**

PI1202591-000

**Date,** 2012-04-03

Page 1 of 1

