



America

CERTIFICATE

No. Z1 13 06 14544 094

Holder of Certificate: **Fluke Corporation**

6920 Seaway Boulevard
Everett, WA 98203-9090
USA

Production Facility(ies): 47951, 84610

Certification Mark:



Product: **Multimeters
(Handheld Digital Clamp Meters)**

Model(s): **323, 324, 325**

Parameters:

Rated Input Voltage:	3 VDC (2 x 1.5V AAA batteries IEC LR03)
Measurement Category:	600V CAT III / 300V CAT IV
Measuring Current:	400A max (At the jaw)
Measuring Voltage:	600V max
Protection Class:	II
IP Protection:	IP30 (not marked)
Temperature Ambient:	-10 to +50°C

Tested according to: EN 61010-1:2001
EN 61010-2-032:2002

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: PI1207512-100

Date, 2013-06-14

Page 1 of 1

