

# Manual Supplement

Manual Title: 323/324/325 Users  
Part Number: 4045153  
Print Date: May 2012  
Revision/Date:

Supplement Issue: **3**  
Issue Date: 7/13  
Page Count: 2

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1, 62269

On page 3, add the following Warning:

- **Do not use a current measurement as an indication that a circuit is safe to touch. A voltage measurement is necessary to know if a circuit is hazardous.**

On page 10, under **Specifications**, *AC Current (Jaw)*, change Accuracy for the 324:

Change: Accuracy.....2.0 %  $\pm 5$  digits (45 Hz - 65 Hz)

2.5 %  $\pm 5$  digits (65 Hz - 400 Hz)

To: Accuracy.....1.5 %  $\pm 5$  digits (45 Hz to 400 Hz)

Add the following Note:

*Note*

*Add 2 % for position sensitivity.*

## Change #2, 63287

On page 11, change:

From: *Frequency (325)*

To: *Frequency with Jaw (325)*

### Change #3, 66148

On page 7, add the following to the **Symbols** table:

	Conforms to relevant South Korean EMC Standards.
---	--

On page 13, under **EMI, EMC** add the following:

Electromagnetic Compatibility ..... Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment) <sup>[1]</sup>

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.