



C E R T I F I C A T E

No. Z1 13 06 14544 093

Holder of Certificate: **Fluke Corporation**

6920 Seaway Boulevard
Everett, WA 98203-9090
USA

**Production
Facility(ies):**

47951, 84610

Certification Mark:



Product:

**Multimeters
(Remote Display TRMS Clamp Meter)**

Model(s):

381

Parameters:

Rated Input Voltage:
Remote Display: 3.0 VDC (2 x 1.5V AAA batteries IEC LR03)
Meter/Clamp base: 4.5 VDC (3 x 1.5V AA batteries IEC LR6)
Protection Class: II
Temperature Ambient: -10 to +50°C
Measurement category: 600V CAT IV, 1000V CAT III

Tested according to:

EN 61010-1:2001
EN 61010-2-032:2002

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

231-1007592-100

Date, 2013-06-14

A handwritten signature in black ink, likely of the certification holder or representative.

Page 1 of 1

