

# **Advanced Validation System**

# Kaye Validator AVS



Turn the page to the Next Generation...


### **Advanced Validation Technology**

The Kaye Validator AVS (Advanced Validation System) is a state of the art validation system design that meets industry and regulatory (FDA / GAMP) requirements for thermal validation.

The Validator AVS combines high accuracy measurements, automated sensor calibration, intuitive metro style user interface, and extensive reporting to simplify the complete validation process.

The Validator AVS is the successor of the widely recognized Kaye Validator 2000, the accepted standard in wired validations systems for over 15 years.

- Hardened, Dedicated Validation Console
- Asset Centric Data Management Concept
- Intuitive Metro-Style User Interface
- Backward Compatible to Existing Validator 2000, ValProbe and RF ValProbe



- Enhanced Connectivity
- Increased Scan Speed

### Lifting Validation to the Next Level

The Kaye Validator AVS System is a unique design and concept combining a stand-alone Validator AVS along with a Validator AVS Console. The AVS console is a rugged hardened console dedicated to interfacing with your Kaye AVS and legacy Kaye hardware. It is pre-loaded with all Kaye software and dedicated to Validation tasks only. This concept greatly simplifies software validation and dependency on continuously changing PC's, Operating Systems, and core loads.

The Kaye Validator AVS offers easy, dedicated and reliable validation. The AVS is intuitive, efficient, and easy to operate - allowing you to focus on the validation, not the technology.



### **Applications - Challenges - Solutions**

### **Applications**




- Steam Sterilizers (Autoclaves)
- Dry Heat Sterilizers
- Steam in Place (SIP)
- Water Cascade/Fall Sterilizers
- Incubators
- Stability Chambers
- Freezers
- Freeze Dryer/Lyophilisation
- Vessels





### Challenges

- Pharmaceutical industries are faced
- IT Environment
  - Increased IT security and lock down
  - Continually changing operation sy - Hardware compatibility
    - Complex software operation
- Validation
  - Diverse evolution of technologies
    Data backward compatibility
  - Complex and time consuming data
    Cost and time of validation and



### Solutions



- Kaye Validator AVS Console dedicated for validation
- OS and Hardware controlled and validated by Kaye
- Common Software tools for Kaye Hardware
- Backward compatible to existing Kaye Products
- Eliminates IT control
- Intuitive metro-style touch screen interface
- Simplified Validation
- Asset Centric Data Management concept
- 21 CFR part 11 compliant





with increasing operational challenges

wn on portable data vstems

in validation

a organization re-validation

## **Kaye Validator AVS**

#### **AVS System**

A Kaye Validator AVS system consists of the Validator AVS and the Validation Console. The console can be docked directly to the Validator AVS and is used as the operator interface to the Validator AVS.

Selectable input capacity (1 to 4 SIMs) up to 48 total inputs.




#### **Robust Design**

- Robust industrial design with two handles
- IPL55 rating, chemical resistant ABS housing
- Dedicated Validation Console for improved user interface
- On-board docking station for Kaye Validation Console
- Battery backup with field replaceable battery pack (3 hours)

#### Data Security via Smart Redundancy Concept

- Standalone operation of Validator AVS console connection not needed
- Validator AVS Internal Memory
- Second independent mirrored memory card for data redundancy
- Data download to validation console
- Forced download of study and audit data to USB
- Backup and restore synchronization of console data with server and other consoles





#### Hardware Connectivity

The Kaye Validator AVS comes complete with improved robust connections for IRTD and Calibration Baths. The Validator AVS is backward compatible with all existing IRTD and Kaye Baths for Automatic Calibration. Two relay outputs are also available to be activated via Qualification events.






The Validator has 4 slots on the back of the unit for easy plug-in of SIMs

#### **Sensor Inputs**

- Up to 4 SIMs 48 channel capacity
- Scan speed of 36 channels per second / 48 channels 2 seconds
- SIMs for TCs, 4-20mA, 0-10V and RTDs
- Improved Sensor Connectivity (quick-fix & lock connectors)
- Accepts a wide range of thermocouple types (T, T premium, J, K, E, B, R, N, S)







### Kaye Validation Console A New Flexible Approach to Validation

The Kaye Validator AVS Console is a state-of-the-art, portable and rugged console dedicated to the programming, displaying, reporting, and storage of Validator AVS data. The Console comes pre-loaded and configured with the suite of Kaye AVS and legacy software and is customized to specific Validations tasks.

The Console offers direct docking and Wi-Fi connectivity with the Validator AVS; it brings about a new approach to tackling your Software Validation.



### Validation Console Specifications

#### Processor

Intel<sup>®</sup> Core<sup>™</sup> Processor

#### **Durability**

#### **IP54** rated

- Water, Dust and Splash resistant
- Durable Gorilla® Glass
- Rubberized for shock dampening
- Molded Handle to prevent drops

#### **Display**

10.4" Display with Projective Capacitive Dual Touch Gorilla Glass

#### **System Storage**

mSATA Solid State Drive (SSD)

#### **Integrated Communications**

Intel® Centrino® Advanced-N 6235 IEEE 802.11a/b/g/n Wi-Fi<sup>®</sup>

#### Separate Docking Station Available

#### I/O Ports

**Docking Connector** 

USB 3.0 port with rubberized cover

#### Embedded I/O

On-Board Camera capability of taking pictures with Console

#### **Dimensions / Weight**

10.0" x 10.0" x .95" (256mm x 256mm x 24.3mm) 3.3 lbs (1,50 kg) <sup>1</sup>

#### **Battery**

Battery life up to 6 hours <sup>2</sup>

#### **Backwards Compatibility**


Can run with Kaye Validator, Kaye ValProbe Software and IRTDWin

1. Weight represents approximate system weight measured with a 40WHr battery. Actual system weight may vary depending on component and manufacturing variability.

2. Battery life varies by configuration, applications in use, utilized features and operating conditions. Maximum battery capacity decreases with time and use.



# Two ways to Connect the Validation Console to the Kaye Validator AVS



The Validator AVS offers a fully functional docking station with direct access to the ports located on rear of the unit.

Console battery is charged while docked.

2 Network Mode

The Validator AVS and the console can connect to a local network by using Ethernet or Wi-Fi connection.

One validation console can handle multiple Kaye Validator AVS's simultaneously.





The Kaye Validator AVS system can establish wireless connections\* by utilizing any kind of available Wi-Fi infrastructure like in-house Wi-Fi access points or simply set up a smartphone as a hotspot. This feature simplifies your daily routine work. You can access the live data wirelessly on the console screen while the Validator is wired on the other side of the autoclave. You can start or stop studies and read the live data from a Kaye Validator AVS in a cleanroom without entering the room.


\* This feature is not available in some countries. Please contact your local Kaye support for details.

# **Kaye Validator AVS Software**

#### Asset Centric Data Management

The Kaye Validator AVS includes an intuitive Asset Centric Data Management concept (patent pending) which allows you to store and access your data faster and more efficiently.

Each individual process that you validate whether an autoclave or freezer etc. can be setup and defined as an asset. All files and data related to this asset, like setups, calibrations, or study files,



are organized and accessed in one single screen around the basic asset data. It is even possible to upload additional documents like standard operation procedures or certificates and associate it with the asset. Assets can be sorted and searched by type, location, manufacturer etc. for easy access.


- Organized Study Data
- Simplified Data Search

#### **Equipment Assets**

The Kaye Validator AVS also allows you to define assets for each piece of Kaye Validation equipment. Data such as serial number, calibration due dates can be defined. The software will automatically notify user when calibrations are due.

The equipment search function uses the Kaye serial number, that is automatically retrieved as part of the study file\*, to find related files. With just one fingertip you get a list of qualification studies, where the equipment asset was used.

\* not for Temperature bath product line



#### **Sensor Calibration**

Kaye the original creator of the Automatic Sensor Calibration feature has included enhancements eliminating manual methods of sensor calibration resulting in better accuracy. The Kaye Validator AVS is backward compatible to existing Kaye IRTD and Calibration baths. The Automatic Calibration feature minimizes training and ensure accurate, and repeatable calibrations optimized for your Kaye calibration equipment



Select only the sensors you want to calibrate. Defining a calibration set lets you calibrate any number of sensors among all those wired to the system.



Set the criteria for a sensor calibration - low, high and check point for the actual calibration.


		-	V	1		-		and C	
sanai	fore:	Sault	1044	Sala	"hep"	THEORY	the .	1 deal	Clarified
8m 1				day 1				641	Time Stepant. 20164
fair(1)	8411	10.00	-	10011	NOT T	11	-	1447	V
14111	-		( 990)	10010	1000	+	-	140	the second second
14111	-	100	144	10010	1007	-	100	Sec.	Called Street
14174	-	-	in.	100.12	10411	-	-	Sector	Sector states a
See 14	100	-	Ac.	144.14	1000.0	-	-	Sain	Second Second
14111	8.071	-	100	1948-17	10.95	-	-	ler1	Strengtheast
Set 14		10.000	Her:	20010	Note T	-		20010	Constanting of the
	-	-	-	1-0.10		-	-	1001	
	***	10.00	100	10010	10.0 %	-	-	141	
	100.0	-	100	loge to	10415	-	-	1407	
140714	-	100	and the	instant.	take to	-	-	1000	

The Console shows the entire calibration process on one screen. Data fields change color to show the progress of stability and deviation for each sensor. A status screen lists each step and indicates where the system is in the process.

#### Hardware

The Kaye Validation console can connect to a Validator AVS directly through the docking port of the AVS or any network based connection available. It enables the user to transfer setups, start studies and monitor live data or read finished studies. After the start a study, the Validator AVS runs the tasks independently. One Kaye Validation console can control several Validator AVS in parallel and one Validator AVS can handle several consoles. That means you can disconnect and work with a different Validator AVS.

While connected to the Validator AVS the user can see the live data in List or Graphical view, group based calculation and event messages. Any connected hardware is displayed with serial number. Simple color code informs the operator of the calibration status of his sensors and sensor input modules.



# **Kaye AVS Reporting Tool**

#### **Document Critical Validation Studies**

The Kaye Validator AVS Console includes an extensive and flexible Reporting Tool used to analyze and document your critical Validation studies. The AVS Reporting Tool is a separate application which is seamlessly integrated into the AVS software. It can be used to analyze and document not only Validator AVS files but also legacy Validator 2000 files.

While offering several new features and enhancements the Reporting Tool was designed to ensure that the proven and accepted formats of the Validator 2000 Summary, Detailed, and Calibration formats were maintained.

Enhancements to Graphing reports, Set-up reports, as well as new reports such as AVS Wiring Layout and Advanced Analytics provide faster and more detailed ways of analyzing your data. Reports can be previewed, printed, saved as a PDF or exported in CSV format.

#### **Configuration Choices**

Prior to generating reports the AVS Reporting Tool provides a host of configuration choices:

- Sensors Included in Report
- Sensors Separated by Groups
- Sensor Placement and Description
- Define Cycles (Qualification, Exposure, etc)
- Calculations (Statistical, Lethality, Saturation, MKT etc)
- Header / Footers
- Graphing
- Templates

These features provide maximum flexibility to ensure you get the data and calculations you need in the format you need to meet your Validation reporting needs.

#### Wiring Diagram

#### Validator AVS Wiring Layout



#### Detailed Lethality Report


Hads been	. Calman	summing in	-					10.00	Property li		Distantial of	-		_
									TUMP .					
Linuty 1	-													
	7001	Tare 15	1944.00	7440.74	20410	764170	144.11	740.78	1.744-19	100110	144711	744710	-	
0.0.14	-	-												
DOM: N														
1000	111	-	110	1.0	1.0		100	100	1.0	1.0	1.00		100	1944
minut .	8.08	-		-		-	1.00							Taxa II
-	+++				100			-		1.00				Sect
	-				12	-	-	12		-				-
Carlor 1		Calacitic I	Case (1)	(watch)	100	100	-	<b>Martin</b>	144010	100	10.	100	100	1000
	10	1.0	1.10	1.0	110		1.0	1.0		1.0	1.17	1.0	1.0	las 1
-	10.00	1.00	100	871	-	10	14	18	140	1.80	187	14	10	144.11
-	100	1.00	14	1.00	1.00	ALC: N	1.00		1.00	4.00	1.86	100	1.0	-
			- C			. T.			- T	-				
	14	hes .	100	1.0	244	4.00	1.4	10	1.0	1.00	1.44	440	1.00	-
100001	-	NAME OF	<b>Gastic</b>	CANESD!	1.0	100	100	In a lot	1000	Date: Dist	841	***	1660	Charles
													-	
in start	100	1.41	100.00	-100	844		1.0	1.00	140	1.00	1.00		4.0	144.71
-	18.14		-	4.0	414	4.0	1.04	110		100	1.19	***	140	1447
-			40			-		+**	-	1.00		10	1.	1.00
			***	-	4.0	***			441	1.00		iii	1.86	100
state()	447	(AND)	1944(11)	10	844	8.01	1847	CHARGE ST	Design 1	1414	100	ALC: NO	0440	(Sec.)
matrial.	4.40	14.00	12.18	-		10	-	4.17	1144	10.01	16.00	-	1.10	See 2
the second	-	HART -	10.84	-	10.00	10.00		44.00	-	10.14	-	10.00	1.00	i fairte
-	10.01	10.00	-	18.00	100.000	1116		11.00	10.00	114	-	11.00	-10	141
-	12.8	16.70	-	18.00	10.44	12:00	-	51.48	-	10.00	10.00	13.86	11.00	Tax 10
alasia() (	NM00	1440010	1000	line (	Test I	14.011	0440	CMARCE:	1440	144		444	1440	Chiefe I
-	-0.01	-	11.00	10.00	-	1918	-	11.00	-	1110		11.00	-	her h
week.	1146	18.00	1948	11.00	1144	1146	-	- 1947	min.	18.18	10.04	-9.46		Take To
-	1647	nie	-	4.00	10.00	18.05	-	14.05	-	48,58	16.41	**	**	101.11
-				-										
-				-										

#### **Calibration Report**

								Pump
p=				Calibratio	oc.ort 19-Ja	-3816.1	1.38.13.8	y Volkeri.
Low	Calibratio	n Point 10	3*8	056557	0.5			
Durt Tento	ana 11.28	ation of Un 13 relate 95.007 relate sharps	TE.		of plating 11	37.04		Eser
-	Sec.	De la	100	North St.	Dy	10	Area 1	10.5
-	10.1012	1000	1.07		84870	1.00	80.5	101
5.46	****	1011	6.01	484110	8876	1.12	10.02	849.12
Deel	stan Erst	untion of U		and Sames				
18.44	-3216 113	17:30			Temperat	re Danié	NO 10 101	¢
<b>Lini</b>	Tang: 1	Die	Lat.	Terry	100000000	100	Cheral Cl	
141	3.0.6	2.85	142	991.2	49.4	143	W (H C	2.84
148	****	0.0010	147	2.68	1412	1.00	412-0	2112
- 4	0.014	491.42	1.01	0.01%	3.03	1.10	1000	98.6
Diam'r	without Floor	ALC: NO	California	on Tempe	Temperation		en de sour-	e
Lini:	Tang	1000	Lat	Tang	COM COM	Let.	They	Der
141	10.00 10	-441-10	148	81.01 %	348.40	1.49	10.03	-241.0
1.08	30.99.02	0.0910	647	8111.0	31986	148	1014.0	2,604
14	80.0	-0.91 %C	8.44	8181-6	311.0	8.42	964 °C	2-18-E
16.00		10.31			Temperat	re there		E
Det:	Sera	344	LIN	Sere	Der	6.00	See.	Der
141	10.00 12	2112	1.02	61.01.2	3194	140	0.01.0	2.00
-	10.00	8.01%	1.07	810.2	1945	1.00	0.00 %	1012
14	8.0.0	8.00 %	1.11	0.01.0	100.0	6.49	min.c	18.2
19-10	+20H 111	99.08			Temperat	re Direit	H H 552'	e
ENC:	Tatla	Pet	1.01	. Temp	Det	646	Thing	244
181	80.00	100.0	198	4111 2	311.0	149	000	2.86
	ware.	194.2	sar.	8111.0	184.4	1.04	40.4	100.0
14	844	-141.0	1.00	41.01	-941.6	1.12	88.5	246.0
18.54	-3818 111	9.30			Temperat	es Timé	and \$6,000*	E
Lie?	Tana	Det	Lac	Tere	Des Color	Let.	Terral 12	des
241	37.30.08	-8.81 10	1-00	30.01 (C	3.88	149	MI 51*C	2,49.6
148	80.0	0.00.00	1.07	88.6	98.2	1.08	80.0310	98.C
14	84.6	-201-12	8.84	2.18.19	-881.16	1.12	100.0	2,816
19.40	- 2111 121	00:04			Temperate	es fiend	we 10.577	£
1043	Same	dire	Lie	See.	der in	1.645	Samp	der
141	8.0.2	1.00	1.05	81872	461.0	140	WED-02	3.8.1
144	新建定	1012	147	相相工	881%	149	603	1.00
1.00	10.0210	-0.04 12	1.11	4101 2	8.8415	1.12	10.01	2.00



#### Graph Report



#### Qualification Report

buly Name Pade	gel saleh	d lost .	-		504	Pedano		50F #W	ene h	tuda-e
					AL	75.64				
amperature D	alarti		Chine I	0			5.0	-010-C	-	
beinet signe Div	a contra		Descore	Cycle	_			Participant in the local distribution of the	Cycle	-
	- Mar	Man	Arg	ð.m	Max-Mir	Mo	Mar	Arg	6.49	Mar Mt
1100,8(%)	21.54	121.58	88.22	3.74	100.05	121.55	122.01	121.89	27.01	0.48
New 128 (101)	21.21	100.71	81.87	2.66	88.40	122.58	101.34	101.11	22.56	4.76
(21) BCT #6(	21.38	120.79	95.71	2.00	88.40	120.66	121.32	121.10	22.50	0.67
type tarries	2138	120.65	81.15	2.00	99.30	125.62	121.30	121.00	12.46	6.08
(27) 8CT engl	21.22	119.81	81.12	2 22		100.00	121.10	120.99	21.04	1.16
type (120 (12)	21,28	120.11	82.14	2.47	-	111.55	121.36	120.85	21.14	1.01
Type T30 (*C)	21.28	120.45		2.00	10.42	100.64	01.78	121.28	214	1.05
Type 131 (*C)	21.38	100.00	-	134	-	100.78	121.40	121.11		
Type T32(*C)	0.00			122	1.22					
	21.54	120.99	89.01	3.38	88.45	100.84	121.38	101.12	32.62	0.98
type 133 (*C)	31.38	121.08		340	813	100.61	121.40	101.20	33.01	0.00
type T24 (*C)	21.28	121.10	36.H	3.30	98.77	120.84	121.44	121.16	32.85	0.60
(2) HET my	21.38	121.13	88.00	340	88.75	120.86	121.52	121.21	23.07	0.88
reperature Se	-		-			-	- 16	-		
Cycle Skart Cycle Duration	11-04-1	01810.24 )	04			21-0(6-2)	010.10.40	116		
den of Mile-		018 10 24		- Fare	31.				Tare	126
fees of Maar Tana	ut		LNI	PTH	U.	102.0	018.10.40 1 Bil 118.10.81	κ.	PTIO	UF.
tes Marge	10.000			- 14 -		1124				
Ant Spread Time	12.8		1	10.2 PT12		200		100	70.4	
ter Aletti	1.1		1.04	PTO		171		14	111	
up of the put	843	1			241	121.				-05-1
Marchine Merci	100		in .	PTH				in.	Table	
	- 10						_		1964	148

### Reporting

- AVS Wiring Layout
- Setup Report
- Calibration Report
- Graph Report
- Detailed Report:
  - Statistical
  - Lethality
  - Saturation
  - MKT
- Summary Report
- Verification Report
- Super Impose Report
- Audit Trail Report



### **Flexible and Compliant**

#### **Electronic Signature**

The Kaye Validator AVS is specifically designed to enable compliance with FDA 21 CFR Part 11. All recorded data, including calibration offsets, set-up parameters, and administrative tasks are saved in secure, encrypted, tamper-proof electronic records in a format accessible only through the system software.


In addition to pre-configured privilege levels, it is possible to explicitly set permissions for each user.

With data synchronization to a shared folder it is possible to exchange configuration and data files like your assets, setups and study files with other Kaye Validation consoles. It also allows to synchronize the user database but also merging the audit trails of several consoles enabling sorting, searching and printing of department-wide audit trails, for example, a list of all failed login attempts within a specified time period across all synchronized Kaye Validation consoles. Every console has a unique but customizable machine ID for identification.

#### **User Management**

Preferences Polici	User Managem	nent		
+ New User	New User			1
And the second second	User Information			
5-m	Voler Letote		ther E	
Area Masterba			Conditio Farmand	
Takitadon Engr	Validation Man	9° v		Lorenter
And Maltperia Validation Engle	Proces.			1
No. of Concession, name	Citable User Actioned			
Infek Dice	User Privileges			
C. Constanting	🖬 Salas Charles	<b>N</b> Out	Rates :	Collection .
Raf 2009kh	🖬 Reports	Martine and American	in Orient States	🖬 Audit half
Quality Manage	🖬 Adman.	<b>.</b>		🛃 vlana Capitan
When Luthin	Carden deserts	🖬 kili A		Colora Roat

#### Policies



#### Audit Trail Report

_		per of terminal	1965	T10010 0140251104				
Ave Tel								
-	Start Started	Berte Bare	in .	Germante				
	(trian)	-	1000104-0000	And S. Y. Inc. Start States, open-				
	1004	-	100000000000	And A first possible to \$ 1.7 per				
	-	-	Nov.77547587	The second secon				
	canal.	and extension	*****	the Start State				
	instant.	-	*******	The Party of the P				
1	-	10.17	10-014101	Int of 7 passed in fact 5 T last				
	-10041	-	-	RAL DESCRIPTION				
1	1000	Sectors.		And A Annual Control of the				
1	-	1011	-	The Local Division of				
		2022	******	Service of Local States of the				
	-	100 10 (mmm)	Same and	No. 5 You have been been and a feature of the second secon				
	-	-	*****	Sect of sectors in a 7 sec.				
		2020		Can of W. Line Honey Tradeol Togeth & Solar Antonio State and Antonio State Recognitional Antonio				
	index.	-		And instance where being hims a printing of the second sec				
1	-	-	10000000000	Their instance allowing fallers in them is a range over which its party fallers in the				
-		-	******	Self Transition				
1	-	-	******	Sect Traphic				
1	-		5-0-0-1214	Des of A. Des place				
1	-	and an instance.		ter C. Y. Sorthers "Schot" agents the Papers' generation bags filters"				
		-	******	period. A line period of second of				
	1000	-	Name and Address	And W. Y. Hardbook Solids' Apparts.				

### Referencing

#### High Accuracy Referencing

Amphenol's temperature calibration equipment is designed specifically to maximize overall system accuracy. Calibration equipment includes temperature references with superior uniformity for sensors, traceable intelligent RTD standards, and validation software to communicate with the hardware.

#### Intelligent RTD Standard

The IRTD Temperature Standard (IRTD-400) is a NIST-traceable instrument that is calibrated over the range of -195°C to 420°C. It is accurate to  $\pm 0.025$ °C over the entire operating range.

The IRTD-400 is a completely self-contained measurement system, containing the electronics for calibration and temperature conversion.

Communicating directly with the Validator software, the IRTD-400 eliminates the potential for human error, assuring accurate and traceable measurements.



	CTR-40	CTR-80
Temperature Range	-40 to 150°C	-80 to 100°C
Temperature Stability	±0.005°C to -40°C (ethanol) ±0.005°C to 25°C (eau) ±0.007°C to 150°C (huile 5012)	±0.03°C
Temperature Uniformity	±0.01°C	±0.03°C

#### Fast/Accurate References

One temperature reference covers the temperature range for the high and low calibration point used for a typical validation study. Choose the model that best fits your need from the chart below.

Temperature dry wells employ unique inserts that minimize cooling of the thermocouple tips due to stem conduction. Without proper inserts, transfer uncertainty in excess of 0.5°C can occur with 22 gage, type T thermocouples. Amphenol units provide an uncertainty of 0.1°C.

Calibrating over a more limited range will also increase accuracy. Since regulations require calibration for the temperature range of a process, an autoclave, for example, can be calibrated from  $90^{\circ}$ C to  $125^{\circ}$ C. This method reduces the error from thermocouple characteristics to less than  $0.05^{\circ}$ C—a two-fold improvement over a calibration at  $0^{\circ}$ C and  $125^{\circ}$ C.

The temperature bath, CTR-80, provides very fast response (90 minutes from ambient to -80°C) and quiet operation. Rugged casters allow this unit to be moved with little effort. A specially designed cover supports two IRTD standards and all thermocouples from the Kaye Validator AVS.





	HTR 400	LTR -25/140	LTR-40/140	LTR-90
Temperature Range	25°C above ambient to 400°C	–25°C to 140°C	–40°C to 140°C	–95°C to 140°C
Ambient Operating Range	5°C to 50°C	5°C to 50°C	5°C to 50°C	5°C to 50°C
Set-Point Accuracy	0.2°C to 300°C 0.3°C to 400°C	0.2°C	0.2°C	0.2°C
Temperature Stability	0.02°C to 300°C 0.05°C to 400°C	0.02°C	0.02°C	0.02°C

### Accessories

The Kaye product range including moisture meters, testers and sensors are designed to meet the most demanding industrial requirements for process improvement, thermal validation and reporting. Specializing in providing turnkey system solutions and supporting them with unmatched technical service, we offer a complete range of temperature standards, baths, thermocouples and fittings, all designed to provide the most accurate process measurement available.

The Kaye product range is relied upon by the world's leading pharmaceutical and biotechnology companies to validate and monitor critical sterilization processes as required by governing regulatory bodies.

- Thermocouples for Autoclaves
- Thermocouples: Stainless Steel
- Thermocouples for Dry Heat Tunnels
- Thermocouples with Stainless Steel Tip

#### Thermocouples

Kaye thermocouple wire is manufactured with the highest purity and uniformity available to the industry. Quality control and testing of every wire spool and thermocouple probe ensures consistent measurement results. Each spool of wire includes a Certificate of Conformance — your guarantee that it meets the accuracy specifications. Each Teflon<sup>®</sup> Thermocouple is leakage vacuum tested.





#### Feed-thru for Autoclave Applications

Easy way to seal the autoclave port when introducing thermocouples into the chamber. Standard 1.5" TRI-CLAMP<sup>®</sup> process connection. Installation is simple with out the need of any tools, fitted with safety release mechanism.

#### Pressure Transducer for Autoclaves

Comply with current standards to measure pressure in parallel to temperature when qualifying autoclaves. The pressure sensor is optimized to work with autoclaves and the Validator<sup>®</sup> AVS.





#### Shipping Case

Protect your Validator AVS during transfer and shipping and store it safely when not being used.

Cables



### **System Documentation**

#### **Quality Control Documents**

Amphenol's quality policy, the ISO 9001 implementation and certificate, and document control standard operating procedures (SOPs)

#### **Development Procedures**

Design control and project management SOPs, and functional specifications

#### **Quality Assurance Procedures**

Test plan and test case procedures

#### **Release Documents**

Quality assurance certification and product release notices

#### **Quality Assurance Test Documentation**

Quality assurance test plan and test cases

#### **IQ/OQ Protocol**

The Installation Qualification/Operational Qualification Protocol defines a set of procedures to ensure that the Kaye Validator AVS system is properly installed and operated according to Amphenol recommendations, and is adequately documented and controlled according to cGMP requirements. The documents are provided in hard copy and on CD, allowing users to modify the documentation to suit specific organizational requirements.

The IQ/OQ Protocol includes the following:

- Installation Qualification document
- Operational Qualification document AVS
- Operational Qualification document AVS Report
- Standard Operating Procedures document

If you prefer to have IQ/OQ executed by qualified Kaye technicians we also provide Validation IQ/OQ On-Site Execution.

#### Validation Reference

The Kaye Validator AVS system is supported with documentation that verifies a fully validated system, including software, hardware and firmware. The Validation Reference Binder provides a comprehensive overview of the Amphenol Quality Policy, description of ISO 9001 implementation and support procedures, and standards for the development, testing, and maintenance of hardware and software. Quality Control documents, Development procedures, Quality Assurance procedures, Release documents, and Quality Assurance test documentation are all included.

The Validation Reference is a serialized document, ensuring that registered users automatically receive notification and updates to keep documentation current. The result is a summary of information you would obtain by conducting an audit at Amphenol's facility—complete, well organized, neatly packaged, and immediately accessible.

Advanced Validation Technology | 17

### **Additional Services**

- Factory / On-Site System Calibration
- Annual Service Contract
- Rentals

## **System Specifications**

#### **Total System Specifications**

When you use specifications to compare equipment, be sure to establish an error budget that accounts for all possible measurement uncertainty. Sensor calibration is an integral part of validation, and total system accuracy should include potential error from the recorder, as well as the temperature reference and traceable standard.

Since all component errors are additive to the total system, every potential error is significant. A summary of the error budget for an Amphenol validation system after sensor calibration with type T thermocouples, used at steam and dry heat, is listed below. These specifications are guaranteed under worst case conditions. Under typical operating conditions, you can expect significantly better accuracy.

Kaye Validator AVS (resolution and short term stability)	0.017°C	k=1
IRTD Temperature Standard	0.01°C	k=1
Temperature Reference	0.051°C	k=1
Total System Uncertainty	0.078°C	k=1



# **Kaye Validation Specifications**

Analog Input	Up to 48
Thermocouples	Type T, J, K,E,B,R,N,S: 0.1°C; T+ limited range 0.01°C resolution
Scanning Speed	36 channel in 1 sec and 48 channels in 2 Sec
Internal Memory	4 gb for data collection
Input Impedance	10K $\Omega$ . Source greater than 10K $\Omega$ produces open circuit indication
Common Mode Rejection	160 db (8 inputs/sec) @ line frequency 145 db (12 inputs/sec) @ line frequency 140 db @ DC
Max. Common Mode Voltage	100V pk ch-to-ch350V pk ch-to ch to frame ground
Normal Mode Rejection	82 db @ 60 Hz (8 inputs/sec)69 db @ 60 Hz (12 inputs/sec)
Voltage Input	0 to 10 VDC
Resolution	1:72,000
Voltage Input Accuracy	30 days: ±(0.003% of reading + 2 counts + 4 microvolts) 1 year: ±(0.006% of reading + 2 counts + 4 microvolts)
Sensitivity	0.5 microvolts/count on most sensitive range
Voltage Temp. Coef.	±(0.1 microvolts + 0.001% reading)/°C
Compensator Temp. Coef.	±0.01°C per °C
Input Terminal Temperature Non-uniformity	±0.1°C from calibrated terminal
Input Ranges	-6 to 30mV, -12 to 60mV, -60 to 300mV, -2 to 10V
Environmental	Temperature: 0 to 50°C (32 to 122°F) Relative humidity: 95% non-condensing
Power	90 to 250 VAC, 50/60 Hz
Fuse Rating	4A Slow Blow
Size	190H X 411W X 381 mm D (457 mm with SIM) 7.5 in H x 16.2 in W x 15 in D (18 in with SIM)
Weight	10.60 kg (23.4 lbs)
Battery	Lithium ion with minimum 3 hours of battery backup

#### **Europe/Asia**

Amphenol Advanced Sensors Germany GmbH Sinsheimerstr 6 75179 Pforzheim Germany

T: +49 (0) 7231 14335 0 F: +49 (0) 7212-14335 29

#### USA

Amphenol Thermometrics Inc. 967 Windfall Rd St. Marys, PA 15857

T: +1 814-834-9140 F: +1 814-781-7969

#### www.kaye-validator-avs.com

#### www.amphenol-sensors.com

 $\ensuremath{\textcircled{O}}$  2016 Amphenol Corporation. All Rights Reserved. Specifications are subject to change without notice.

Other company names and product names used in this document are the registered trademarks of their respective owners.

