



Standards Council of Canada  
Conseil canadien des normes

Program for Accreditation of Laboratories – Canada  
Programme d'accréditation des laboratoires – Canada

## CERTIFICATE OF ACCREDITATION



National Research  
Council Canada

Conseil national  
de recherches Canada

Calibration Laboratory Assessment Service  
Service d'évaluation de laboratoires d'étalonnage

## CERTIFICAT D'ACCRÉDITATION

**Transcat Canada - Montréal**  
9900 Côte-de-Liesse, Montreal, QC H8T 1A1

having been assessed by the National Research Council of Canada (NRC), under the authority of the Standards Council of Canada (SCC), and found to conform with the requirements of ISO/IEC 17025:2017 and conditions established by SCC, and the NRC Calibration Laboratory Assessment Service (CLAS), and having demonstrated the capability of calibrating measurement instruments and standards and providing verified traceability to the national measurement standards of Canada, in specified fields and specified uncertainty limits, is hereby recognized as an

**ACCREDITED CALIBRATION LABORATORY**  
For specific measurement capabilities which are hereby CERTIFIED  
by CLAS




**LABORATOIRE D'ÉTALONNAGE ACCRÉDITÉ**  
CERTIFIÉ par le CLAS pour des capacités précises de mesurage


as listed in the Directory of the Canadian Calibration Network maintained by NRC and approved by SCC. The national measurement standards of Canada are realized, maintained and disseminated by NRC under the authority of the *National Research Council Act* and the *Weights and Measurements Act*. Bilateral agreements recognizing the equivalence of national measurement standards exist between NRC and other national metrology institutes. Copies of these agreements are available from NRC.

indiquées dans le Répertoire du réseau canadien d'étalonnage établi par le CNRC et approuvé par le CCN. Les étalons nationaux du Canada sont établis, maintenus et émis par le CNRC en vertu de la *Loi sur le Conseil national de recherches* et de la *Loi sur les poids et mesures*. Il existe entre le CNRC et d'autres instituts nationaux de métrologie des accords bilatéraux qui reconnaissent l'équivalence des étalons nationaux de mesure. Le CNRC tient à la disposition du public des exemplaires de ces accords.



  
Chief Metrologist (NRC) / Métrologue en chef (CNRC)

Accredited laboratory number: / Numéro de laboratoire accrédité : 668  
SCC file number: / Dossier du CCN n° : 15845  
NRC CLAS Certificate No. / Numéro du certificat CNRC CLAS : 2010-01  
Initial accreditation date: / Date de la première accréditation : 2010-03-01

  
Vice-President – Accreditation Services / Vice-président – Services d'accréditation

Issued on: / Délivré le : 2020-03-01

The validity of this certificate, including the date of last re-accreditation and its expiry can be confirmed by the accompanying Scope of Accreditation document in the Directory of Accredited Laboratories on the SCC website at [www.scc.ca](http://www.scc.ca).

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. The accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF communiqué dated April 2017).

Pour vérifier la validité du présent certificat, y compris la date de la dernière réaccréditation et la date d'expiration du certificat, consulter la portée d'accréditation qui se trouve dans le répertoire des laboratoires accrédités dans le site Web du CCN au [www.ccn.ca](http://www.ccn.ca).

Ce laboratoire est accrédité conformément à la Norme internationale reconnue ISO/IEC 17025:2017. Cette accréditation démontre la compétence technique d'un organisme pour une portée définie et l'exploitation d'un système de management de la qualité de laboratoire (cf. communiqué conjoint ISO-ILAC-IAF date de avril 2017).



9900 Cote-de-Liesse, Montreal, QC H8T 1A1 • Phone: 514.631.6653 • Fax: 514.631.6122 • [Transcat.ca](http://Transcat.ca)

---

January 18, 2021

Dear Valued Customer,

We are pleased to provide this letter of clarification in reference to the information on our current ISO 17025 certificates of accreditation.

Our laboratories have been audited and assessed by the National Research Council of Canada (NRC) through its Calibration Laboratory Assessment Service (CLAS). Based on the successful completion of our audits, we are confirmed as accredited by the Standards Council of Canada (SCC) who provides our ISO 17025 certificates of accreditation. These frameable paper certificates are formally issued by the Standards Council of Canada and indicate our SCC laboratory number, our CLAS certificate number, the date of our initial accreditation and the date on which the paper certificate was printed.

For the period of validity of our **current scope of accreditation**, including the **date on which the scope was issued and the expiry date**, these are published on the SCC website at [scc.ca](http://scc.ca).

For the list of our full measurement capabilities, our detailed scope of accreditation is outlined on the NRC website at [nrc.canada.ca](http://nrc.canada.ca) in the directory of accredited calibration laboratories.

If you have any questions, please do not hesitate to contact our Quality Manager, David Llorens via email at [david.llorens@transcat.ca](mailto:david.llorens@transcat.ca) or via telephone at 1-800-828-1470 extension 7232.

We appreciate being your trusted calibration partner, and we thank you for your business!

Yours very truly,

A handwritten signature in blue ink that reads "Ingrid M. Ulrich". The signature is fluid and cursive, with the first letters of the first and last names being capitalized.

Ingrid M. Ulrich, CA, CPA, MBA  
Vice-President, Operations & Administration  
Transcat Canada Inc.



## TESTING AND CALIBRATION LABORATORY ACCREDITATION PROGRAM (LAP)

### Scope of Accreditation

Accredited Laboratory No. 668

**Legal Name of Accredited Laboratory:** Transcat Canada - Montréal

Contact Name: David Llorens

Address: 9900 Côte-de-Liesse,  
Montréal, QC H8T 1A1

Telephone: +1 514 631 6653

Fax: +1 514 631 6122

Website: [www.dispersion.ca/public/en/index.php](http://www.dispersion.ca/public/en/index.php)

Email: [David.Llorens@transcat.ca](mailto:David.Llorens@transcat.ca)

<b>SCC File Number:</b>	15845
<b>Provider:</b>	NRC-CLAS
<b>Provider File Number:</b>	504
<b>Accreditation Standards:</b>	ISO/IEC 17025:2017
<b>Clients Served:</b>	All interested parties Some calibration services are available on-site.
<b>Field of Calibration:</b>	Mass Volume
<b>Program Specialty Area:</b>	Calibration
<b>Initial Accreditation:</b>	2010-03-02
<b>Most Recent Accreditation:</b>	2020-03-01
<b>Accreditation Valid to:</b>	2022-03-02

*Remarque: La présente portée d'accréditation existe également en français, sous la forme d'un document distinct.*  
*Note: This scope of accreditation is also available in French as a separately issued document.*



## **CALIBRATION OF MEASURING AND TEST EQUIPMENT**

For calibration measurement capability, please refer to the Canadian Calibration Network web page at the National Research Council of Canada. This laboratory is accredited by the Standards Council of Canada as part of the Calibration Laboratory Assessment Service (CLAS) program and is listed at [nrc.canada.ca](http://nrc.canada.ca).

This document forms part of the Certificate of Accreditation issued by the Standards Council of Canada (SCC). The original version is available in the Directory of Accredited Laboratories on the SCC website at [www.scc.ca](http://www.scc.ca).

---

Elias Rafoul  
Vice-President, Accreditation Services  
Publication on: 2020-03-05

# CLAS Certificate Number 2010-01

**From:** National Research Council Canada

<b>Company name</b>	<b>Transcat Canada – Montreal</b>
<b>Company address</b>	9900 Côte-de-Liesse Montréal, Québec H8T 1A1
<b>Contact</b>	<b>David Llorens</b> <b>Telephone:</b> 514-631-6653 <b>Fax:</b> 514-631-6122 <b>Email:</b> <a href="mailto:David.Llorens@transcat.ca">David.Llorens@transcat.ca</a>
<b>Clients served</b>	<ul style="list-style-type: none"><li>• All interested parties</li><li>• Some calibration services are available on-site. These services are indicated in the "Remarks" column of the following pages</li></ul>
<b>Fields of calibration</b>	<ul style="list-style-type: none"><li>• Mass</li><li>• Volume</li></ul>
<b>SCC accreditation (ISO/IEC 17025)</b>	<ul style="list-style-type: none"><li>• Accredited Laboratory N° (number) 668</li><li>• First issued 2010-03-02</li><li>• Issue 7.3e 2020-08-28</li></ul>

**i** This scope of calibration capabilities is published by the CLAS program of the National Research Council of Canada (NRC) in close co-operation with the laboratory accreditation program of the

Standards Council of Canada (SCC), Canada's accreditation body for calibration and testing laboratories. The SCC accredits the capability of the named laboratory for being able to perform the listed calibrations at the given Calibration Measurement Capability (see Supplementary **note C** and **note D**) with traceability to the International System of Units (SI) or to standards acceptable to the CLAS program.

Measured Quantity and Range or Instrument	Calibration and Measurement Capability expressed as an Uncertainty ( $\pm$ ) (See <u>supplementary notes</u> )	Type of Service	Remarks
<b>Mass</b>			
25 kg	30 mg	I	One to one comparison. For the calibration of masses with traceability to the International System of Units (SI) through the primary standards of NRC and of Measurement Canada (MR-1). Calibration results are reported in
20 kg	20 mg		
10 kg	10 mg		
5 kg	2.5 mg		
3 kg	2.4 mg		
2 kg	2.0 mg		
1 kg	0.5 mg		
500 g	0.24 mg		

300 g	0.15 mg
200 g	0.10 mg
100 g	0.080 mg
50 g	0.040 mg
30 g	0.050 mg
20 g	0.033 mg
10 g	0.017 mg
5 g	0.010 mg
3 g	0.008 mg
2 g	0.007 mg
1 g	0.007 mg
500 mg	0.004 mg
300 mg	0.003 mg
200 mg	0.002 mg
100 mg	0.002 mg
50 mg	0.002 mg
30 mg	0.002 mg
20 mg	0.002 mg
10 mg	0.002 mg
5 mg	0.002 mg
3 mg	0.002 mg

terms of conventional mass as defined by the Organisation internationale de métrologie légale (OIML): "The conventional value of the value of the result of weighing a body in air is equal to mass of a standard, of conventionally chosen density ( $8000 \text{ kg/m}^3$ ), at a conventionally chosen temperature ( $20^\circ\text{C}$ ), which balances this body at this reference temperature in air of conventionally chosen density ( $1.2 \text{ kg/m}^3$ )."

The Calibration Measurement Capability listed relates to measurements on integral conventional standard weights. It can be achieved only if the weights being

2 mg	0.002 mg		calibrated are suitable for such a measurement. Calibrations can be given in other units as required.  This service is not available on-site.
1 mg	0.002 mg		
10 kg	40 mg	I	One to one comparison. On-site calibration available. Procedure : PDL-09-MG-030
20 kg	60 mg		
25 kg	70 mg		

### **Balances (Electronic, non-automatic) - Specific values**

1 mg, 2 mg and 5 mg	0.003 mg	II	For the calibration of balances. The Calibration Measurement Capability listed can be achieved only if the balances being calibrated are suitable for such a measurement. The Calibration Measurement Capability is based upon the use of ASTM Class 1 reference weights. The laboratory corrects for known
10 mg, 20 mg and 50 mg	0.003 mg		
100 mg and 200 mg	0.003 mg		
500 mg	0.005 mg		
1 g and 2 g	0.007 mg		
5 g	0.010 mg		
10 g	0.019 mg		



offsets assigned to its reference weights so that it can calibrate with uncertainties that are better than the ASTM tolerances of the weights used, in some cases. The Calibration Measurement Capability reflects the uncertainty contribution of the weights including co-variances. It also includes the repeatability and readability of the best balances that the laboratory calibrates routinely in the given measurement ranges. The Calibration Measurement Capabilities for measurements between 500 g and 30 kg are dominated by the performances of the best balances calibrated by the laboratory, rather

than by the laboratory's reference standards and measurement processes. The uncertainty stated on the calibration report will include the uncertainty contribution of the balances that were calibrated. All values are expressed in conventional mass, as defined above.

This service is available on-site.

### Balances (Electronic, non-automatic) - Other values

>10 g to 20 g	0.035 mg	II	For the calibration of balances. The Calibration Measurement Capability listed can be achieved only if the balances being calibrated are suitable for such a measurement. The Calibration Measurement Capability is based upon the use of ASTM Class 1
>20 g to 50 g	0.10 mg		
>50 g to 100 g	0.10 mg		
>100 g to 200 g	0.15 mg		
>200 g to 300 g	0.25 mg		
>300 g to 500 g	0.35 mg		
>500 g to 1.2 kg	2.0 mg		
>1.2 kg to 8 kg	40 mg		
>8 kg to 30 kg	80 mg		

reference weights. The laboratory corrects for known offsets assigned to its reference weights so that it can calibrate with uncertainties that are better than the ASTM tolerances of the weights used, in some cases. The Calibration Measurement Capability reflects the uncertainty contribution of the weights including co-variances. It also includes the repeatability and readability of the best balances that the laboratory calibrates routinely in the given measurement ranges. The Calibration Measurement Capabilities for measurements between 500 g and 30 kg are dominated by the performances

of the best balances calibrated by the laboratory, rather than by the laboratory's reference standards and measurement processes. The uncertainty stated on the calibration report will include the uncertainty contribution of the balances that were calibrated. All values are expressed in conventional mass, as defined above.

This service is available on-site.

0 kg to 60 kg	300 mg	II	On-site calibration available.  Procedure : PDL-09-MG-010
60 kg to 150 kg	3000 mg		
150 kg to 500 kg	30000 mg		
500 kg to 1000 kg	60000 mg		

### **Piston Pipette**

0.2 µL to 2 µL	0.017 µL	II	For the calibration of single channel piston pipettes in accordance with the laboratory's
> 2 µL to 10 µL	0.021 µL		
> 10 µL to 20 µL	0.025 µL		

> 20 µL to 100 µL	0.10 µL		calibration procedure PDL-09-MG-050.  This calibration service is available on-site.
> 100 µL to 200 µL	0.14 µL		
> 200 µL to 1000 µL	0.76 µL		
> 1000 µL to 5000 µL	4.4 µL		
> 5000 µL to 10000 µL	5.9 µL		
> 10000 µL to 50000 µL	7.6 µL		
2 µL to 20 µL	0.31 µL	II	For the calibration of multi-channel piston pipettes in accordance with the laboratory's calibration procedure PDL-09-MG-050.  This calibration service is only available in the laboratory.
> 20 µL to 200 µL	0.36 µL		
> 200 µL to 1000 µL	0.76 µL		

**Date modified:**

2020-03-04