



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

### United Scale & Engineering Corporation

A TRANSCAT COMPANY

16725 W. Victor Road

New Berlin WI 53151

(including satellite location listed on the scope)

Fulfills the requirements of

## ISO/IEC 17025:2017

and the national standards

## ANSI/NCSL Z540-1-1994 (R2002)

In the field of

## CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 07 September 2021

Certificate Number: AC-2489.16



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
AND ANSI/NCSL Z540-1-1994 (R2002)**

**United Scale & Engineering Corporation**  
A TRANSCAT COMPANY  
16725 W. Victor Road  
New Berlin, WI 53151  
Dan Christianson  
800-236-1733

**CALIBRATION**

Valid to: **September 7, 2021**

Certificate Number: **AC-2489.16**

**Mass and Mass Related**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Class I Balances	Up to 100 g	0.29 mg	ASTM Class 1 weights
	Up to 230 g	0.59 mg	
	Up to 610 g	1.5 mg	
Class II Balances	Up to 610 g	3.2 mg	OIML Class F1 weights
	Up to 6 100 g	32 mg	
Class II Balances	Up to 6 400 g	1.9 g	Class F weights
	Up to 32 kg	2.9 g	
	Up to 34 kg	3.9 g	
	Up to 64 kg	7.3 g	
	Up to 100 kg	12 g	
	Up to 200 kg	24 g	
Class III Light capacity Scales	Up to 2 lb	0.000 72 lb	Class F weights
	Up to 5 lb	0.005 9 lb	
	Up to 10 lb	0.002 6 lb	
	Up to 20 lb	0.006 2 lb	
	Up to 50 lb	0.013 lb	
	Up to 100 lb	0.026 lb	
	Up to 200 lb	0.062 lb	
Class III Medium Capacity Scales	Up to 500 lb	0.13 lb	Class F weights
	Up to 1 000 lb	0.26 lb	
	Up to 2 000 lb	0.62 lb	
	Up to 5 000 lb	1.3 lb	
	Up to 10 000 lb	2.8 lb	
	Up to 20 000 lb	6.9 lb	

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Class III Medium Capacity Scales	Up to 400 kg	0.14 kg	Class F weights
	Up to 600 kg	0.15 kg	
	Up to 1 000 kg	0.28 kg	
	Up to 2 500 kg	0.48 kg	
	Up to 5 000 kg	0.87 kg	
	Up to 9 000 kg	1.2 kg	
Class III L Heavy Capacity Scales	Up to 50 000 lb	14 lb	Class F weights
	Up to 100 000 lb	26 lb	
	Up to 200 000 lb	26 lb	

**Services performed at satellite laboratory**

1322 Russett Court  
Green Bay, WI 54313

**Mass and Mass Related**

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Class I Balances	Up to 100 g	0.29 mg	ASTM Class 1 weights
	Up to 230 g	0.59 mg	
	Up to 610 g	1.5 mg	
Class II Balances	Up to 610 g	3.2 mg	OIML Class F1 weights
	Up to 6 100 g	32 mg	
Class II Balances	Up to 6 400 g	1.9 g	Class F weights
	Up to 32 kg	2.9 g	
	Up to 34 kg	3.9 g	
	Up to 64 kg	7.3 g	
	Up to 100 kg	12 g	
	Up to 200 kg	24 g	
Class III Light capacity Scales	Up to 2 lb	0.000 72 lb	Class F weights
	Up to 5 lb	0.005 9 lb	
	Up to 10 lb	0.002 6 lb	
	Up to 20 lb	0.006 2 lb	
	Up to 50 lb	0.013 lb	
	Up to 100 lb	0.026 lb	
	Up to 200 lb	0.062 lb	



ANSI National Accreditation Board

### Mass and Mass Related

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Class III Medium Capacity Scales	Up to 500 lb	0.13 lb	Class F weights
	Up to 1 000 lb	0.26 lb	
	Up to 2 000 lb	0.62 lb	
	Up to 5 000 lb	1.3 lb	
	Up to 10 000 lb	2.8 lb	
	Up to 20 000 lb	6.9 lb	
Class III Medium Capacity Scales	Up to 400 kg	0.14 kg	Class F weights
	Up to 600 kg	0.15 kg	
	Up to 1 000 kg	0.28 kg	
	Up to 2 500 kg	0.48 kg	
	Up to 5 000 kg	0.87 kg	
	Up to 9 000 kg	1.2 kg	
Class III L Heavy Capacity Scales	Up to 50 000 lb	14 lb	Class F weights
	Up to 100 000 lb	26 lb	
	Up to 200 000 lb	26 lb	

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2489.16.

R. Douglas Leonard Jr., VP, PILR SBU