



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Transcat – San Diego
7726 Arjons Drive
San Diego, CA 92126

Fulfills the requirements of

ISO/IEC 17025:2017

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to read 'R. Douglas Leonard Jr.', is positioned above a horizontal line.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 01 September 2024

Certificate Number: L2214



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
AND
ANSI/NCSL Z540-1-1994 (R2002)**

Transcat – San Diego
7726 Arjons Drive
San Diego, CA 92126
Martin Bakker 858-621-2630

CALIBRATION

Valid to: **September 1, 2024**

Certificate Number: **L2214**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Source ²	Up to 220 μ A (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A	45 μ A/A + 6.9 nA 39 μ A/A + 8.1 nA 39 μ A/A + 46nA 58 μ A/A + 0.7 μ A 0.24 mA/A + 12 μ A	Fluke 5720A Multiproduct Calibrator
DC Voltage – Source ²	Up to 220 mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 000) V	9.1 μ V/V + 0.4 μ V 5.7 μ V/V + 0.7 μ V 4.4 μ V/V + 2.5 μ V 4 μ V/V + 4 μ V 6.3 μ V/V + 40 μ V 7.6 μ V + 0.4 mV	Fluke 5720A Multiproduct Calibrator
AC Current – Source ²	Up to 220 μ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.032 % of reading + 16 nA 0.019 % of reading + 10 nA 0.014 % of reading + 8 nA 0.026 % of reading + 10 nA 0.11 % of reading + 65 nA 0.031 % of reading + 40 nA 0.019 % of reading + 35 nA 0.014 % of reading + 35 nA 0.026 % of reading + 0.11 μ A 0.11 % of reading + 0.65 μ A	Fluke 5720A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ²	(2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.033 % of reading + 0.4 μA 0.02 % of reading + 0.35 μA 0.015 % of reading + 0.35 μA 0.022 % of reading + 0.55 μA 0.11 % of reading + 5 μA 0.033 % of reading + 4 μA 0.018 % of reading + 3.5 μA 0.014 % of reading + 2.5 μA 0.021 % of reading + 3.5 μA 0.11 % of reading + 10 μA 0.047 % of reading + 0.17 mA 0.095 % of reading + 0.38 mA 0.36 % of reading + 0.16 mA	Fluke 5720A Multiproduct Calibrator
AC Voltage – Source ²	Up to 2.2 mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.16 % of reading + 4 μV 0.1 % of reading + 4 μV 0.077 % of reading + 4 μV 0.12 % of reading + 4 μV 0.17 % of reading + 5 μV 0.33 % of reading + 10 μV 0.47 % of reading + 20 μV 0.58 % of reading + 20 μV 0.044 % of reading + 4 μV 0.031 % of reading + 4 μV 0.015 % of reading + 4 μV 0.031 % of reading + 4 μV 0.059 % of reading + 5 μV 0.12 % of reading + 10 μV 0.16 % of reading + 20 μV 0.3 % of reading + 20 μV	Fluke 5720A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ²	(22 to 220) mV		Fluke 5720A Multiproduct Calibrator
	(10 to 20) Hz	0.028 % of reading + 12 μ V	
	(20 to 40) Hz	0.011 % of reading + 7 μ V	
	40 Hz to 20 kHz	0.009 % of reading + 7 μ V	
	(20 to 50) kHz	0.021 % of reading + 7 μ V	
	(50 to 100) kHz	0.047 % of reading + 17 μ V	
	(100 to 300) kHz	0.092 % of reading + 20 μ V	
	(300 to 500) kHz	0.14 % of reading + 25 μ V	
	500 kHz to 1 MHz	0.28 % of reading + 45 μ V	
	(0.22 to 2.2) V		
	(10 to 20) Hz	0.028 % of reading + 40 μ V	
	(20 to 40) Hz	0.01 % of reading + 15 μ V	
	40 Hz to 20 kHz	0.005 % of reading + 8 μ V	
	(20 to 50) kHz	0.008 % of reading + 10 μ V	
	(50 to 100) kHz	0.012 % of reading + 30 μ V	
	(100 to 300) kHz	0.043 % of reading + 80 μ V	
	(300 to 500) kHz	0.1 % of reading + 0.2 mV	
	500 kHz to 1 MHz	0.18 % of reading + 0.3 mV	
	(2.2 to 22) V		
	(10 to 20) Hz	0.028 % of reading + 0.4 mV	
	(20 to 40) Hz	0.01% of reading + 0.15 mV	
	40 Hz to 20 kHz	0.005 % of reading + 0.05 mV	
	(20 to 50) kHz	0.008 % of reading + 0.1 mV	
	(50 to 100) kHz	0.011 % of reading + 0.2 mV	
(100 to 300) kHz	0.03 % of reading + 0.6 mV		
(300 to 500) kHz	0.1 % of reading + 2 mV		
500 kHz to 1 MHz	0.17 % of reading + 3.2 mV		
(22 to 220) V			
(10 to 20) Hz	0.028 % of reading + 4 mV		
(20 to 40) Hz	0.01% of reading + 1.5 mV		
40 Hz to 20 kHz	0.006 % of reading + 0.6 mV		
(20 to 50) kHz	0.009 % of reading + 1 mV		
(50 to 100) kHz	0.016 % of reading + 2.5 mV		
(100 to 300) kHz	0.09 % of reading + 16 mV		
(220 to 750) V			
(30 to 50) kHz	0.061 % of reading + 11 mV		
(50 to 100) kHz	0.23 % of reading + 45 mV		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Resistance – Source ² (Fixed-Simulated Values)	1 Ω	98 μΩ/Ω	Fluke 5720A Multiproduct Calibrator
	1.9 Ω	96 μΩ/Ω	
	10 Ω	24 μΩ/Ω	
	19 Ω	25 μΩ/Ω	
	100 Ω	11 μΩ/Ω	
	190 Ω	11 μΩ/Ω	
	1 kΩ	9.4 mΩ/kΩ	
	1.9 kΩ	10 mΩ/kΩ	
	10 kΩ	10 mΩ/kΩ	
	19 kΩ	10 mΩ/kΩ	
	100 kΩ	12.6 mΩ/kΩ	
	190 kΩ	29.5 mΩ/kΩ	
	1 MΩ	22 Ω/MΩ	
	1.9 MΩ	125 Ω/MΩ	
	10 MΩ	74 Ω/MΩ	
19 MΩ	0.65 kΩ/MΩ		
100 MΩ	0.6 kΩ/MΩ		

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. Based on an accredited calibration by the manufacturer, used at the temperature in which the Multiproduct Calibrator was calibrated ($t_{cal} = \pm 5^\circ\text{C}$) and assuming the instrument is zeroed at least every seven days or when the ambient temperature changes more than 5°C .
3. This scope is formatted as part of a single document including Certificate of Accreditation No. L2214.



R. Douglas Leonard Jr., VP, PILR SBU