



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**Transcat - Chesapeake**  
928 Canal Drive  
Chesapeake, VA 23323

Fulfills the requirements of

**ISO/IEC 17025:2017**

and national standards

**ANSI/NCSL Z540-1-1994 (R2002) AND  
ANSI/NCSL Z540.3-2006 (R2013)**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 07 September 2025  
Certificate Number: AC-2489.21



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**AND**

**ANSI/NCSL Z540-1-1994 (R2002)  
ANSI/NCSL Z540.3-2006 (R2013)**

**Transcat – Chesapeake**  
928 Canal Drive  
Chesapeake, VA 23323  
Dante Daneri 757-558-2550

**CALIBRATION**

Valid to: **September 7, 2025**

Certificate Number: **AC-2489.21**

**Acoustics and Vibration**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Vibration – Measure Voltage Sensitivity	100 Hz 10 mV/g	1.5 % of reading	Reference Accelerometer w/ Calibrator
Frequency Response	(0.8 to 20) g (5 to 20) Hz (20 to 100) Hz (100 to 2 500) Hz (2 500 to 10 000) Hz  (0.8 to 20) g pk (7 to 10) Hz (10 to 30) Hz (300 to 2 000) Hz (2 000 to 10 000) Hz	2.1 % of reading 1.9 % of reading 1.5 % of reading 2.8 % of reading  7 % of reading 6 % of reading 4.5 % of reading 7 % of reading	Vibration Calibrator
Sound – Generate	1 kHz 110 dB	0.42 dB	SPL Calibrator
Sound – Measure	20 Hz to 10 kHz (50 to 120) dB	0.5 dB	Sound Level Meter



ANSI National Accreditation Board

**Chemical Quantities**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters	4 pH	0.38 % of reading	Accredited pH Solutions
	7 pH	0.21 % of reading	
	10 pH	0.15 % of reading	
Conductivity Meters	25 $\mu$ S/cm	7.4 % of reading	Accredited Conductivity Solutions
	75 $\mu$ S/cm	1.8 % of reading	
	1 015 $\mu$ S/cm	0.54 % of reading	
	1 408 $\mu$ S/cm	0.53 % of reading	
Gas Detection Equipment <sup>1</sup> CO (Carbon Monoxide)	0.002 % CO	0.000 088 % CO	Certified Gas Mixtures
	0.006 % CO	0.000 16 % CO	
	0.01 % CO	0.000 23 % CO	
	0.1 % CO	0.000 84 % CO	
	0.5 % CO	0.003 2 % CO	
CH <sub>4</sub> (Methane LEL)	50 % LEL	1.1 % LEL	
C <sub>5</sub> H <sub>12</sub> (Pentane LEL)	58 % LEL	1.5 % LEL	
H <sub>2</sub> S (Hydrogen Sulfide)	0.002 5 % H <sub>2</sub> S	0.000 042 % H <sub>2</sub> S	
	0.002 % H <sub>2</sub> S	0.000 052 % H <sub>2</sub> S	
O <sub>2</sub> (Oxygen)	0.5 % O <sub>2</sub>	0.01 % O <sub>2</sub>	
	5 % O <sub>2</sub>	0.14 % O <sub>2</sub>	
	15 % O <sub>2</sub>	0.17 % O <sub>2</sub>	
	18 % O <sub>2</sub>	0.2 % O <sub>2</sub>	
C <sub>4</sub> H <sub>8</sub> (Isobutylene)	0.01 % C <sub>4</sub> H <sub>8</sub>	0.000 23 % C <sub>4</sub> H <sub>8</sub>	
CO <sub>2</sub> (Carbon Dioxide)	0.5 % CO <sub>2</sub>	0.008 1 % CO <sub>2</sub>	
	5 % CO <sub>2</sub>	0.043 % CO <sub>2</sub>	

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source <sup>1</sup>	(0 to 220) mV	7.5 $\mu$ V/V + 0.4 $\mu$ V	Fluke 5730A/03 Multiproduct Calibrator
	(0.22 to 2.2) V	5 $\mu$ V/V + 0.7 $\mu$ V	
	(2.2 to 11) V	3.5 $\mu$ V/V + 2.5 $\mu$ V	
	(11 to 22) V	3.5 $\mu$ V/V + 4 $\mu$ V	
	(22 to 220) V	5 $\mu$ V/V + 40 $\mu$ V	
	(220 to 1 100) V	6.5 $\mu$ V/V + 0.4 mV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Measure <sup>1</sup>	Up to 200 mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1 000) V	5.2 $\mu\text{V/V} + 90 \text{ nV}$ 3.6 $\mu\text{V/V} + 0.39 \mu\text{V}$ 3.5 $\mu\text{V/V} + 3.9 \mu\text{V}$ 5.5 $\mu\text{V/V} + 39 \mu\text{V}$ 5.5 $\mu\text{V/V} + 0.47 \text{ mV}$	Fluke 8508A opt 001 8.5 Digit Multimeter
DC High Voltage – Measure <sup>1</sup>	(1 to 10) kV	0.039 % of reading + 92 mV	Vitrek 4700 High Voltage Meter
DC High Voltage – Measure <sup>1</sup>	Up to 40 kV	2 % of reading	Fluke 80K40HV Probe, Fluke 8508A opt 001 8.5 Digit Multimeter
DC Current – Source <sup>1</sup>	(0 to 220) $\mu\text{A}$ (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A	40 $\mu\text{A/A} + 6 \text{ nA}$ 35 $\mu\text{A/A} + 7 \text{ nA}$ 35 $\mu\text{A/A} + 40 \text{ nA}$ 45 $\mu\text{A/A} + 0.7 \mu\text{A}$ 80 $\mu\text{A/A} + 12 \mu\text{A}$	Fluke 5730A/03 Multiproduct Calibrator
DC Current – Source <sup>1</sup>	(2.2 to 3) A (3 to 11) A (11 to 20.5) A	0.3 mA/A + 31 $\mu\text{A}$ 0.51 mA/A + 0.39 mA 0.93 mA/A + 0.58 mA	Fluke 5522A/11 Multiproduct Calibrator
DC Current – Source <sup>1</sup>	(20 to 100) A	0.04 % of reading	Current Source, SR-100 Current Shunt, Fluke 8508 8.5 Digit Multimeter
DC Clamp-on Ammeter (Non-Toroidal Type) Hall Effect Sensor <sup>1</sup>	(20 to 150) A (150 to 1 000) A	0.51 % of reading + 0.14 A 0.52 % of reading + 0.5 A	Fluke 5522A/11 Multiproduct Calibrator, 50-turn Coil
DC Current – Measure <sup>1</sup>	Up to 200 $\mu\text{A}$ (0.2 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A (2 to 20) A	13 $\mu\text{A/A} + 0.31 \text{ nA}$ 13 $\mu\text{A/A} + 3.1 \text{ nA}$ 14 $\mu\text{A/A} + 31 \text{ nA}$ 47 $\mu\text{A/A} + 0.62 \mu\text{A}$ 0.18 mA/A + 12 $\mu\text{A}$ 0.39 mA/A + 0.31 mA	Fluke 8508A opt 001 8.5 Digit Multimeter
DC Current – Measure <sup>1</sup>	(20 to 100) A	0.04 % of reading + 24 $\mu\text{A}$	SR-100 Current Shunt, Fluke 8508A opt 001 8.5 Digit Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source <sup>1</sup> (Simulated)	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω (0.33 to 1.1) kΩ (1.1 to 3.3) kΩ (3.3 to 11) kΩ (11 to 33) kΩ (33 to 110) kΩ (110 to 330) kΩ (0.33 to 1.1) MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ (0.33 to 1.1) GΩ	32 μΩ/Ω + 0.78 mΩ 24 μΩ/Ω + 1.2 mΩ 22 μΩ/Ω + 1.1 mΩ 22 μΩ/Ω + 1.6 mΩ 22 μΩ/Ω + 1.6 mΩ 22 μΩ/Ω + 1.6 mΩ 22 μΩ/Ω + 0.16 Ω 22 μΩ/Ω + 0.16 Ω 27 μΩ/Ω + 1.6 Ω 26 μΩ/Ω + 1.6 Ω 66 μΩ/Ω + 23 Ω 0.1 mΩ/Ω + 39 Ω 0.19 mΩ/Ω + 1.9 kΩ 0.41 mΩ/Ω + 2.3 kΩ 0.23 % of reading + 78 kΩ 1.2 % of reading + 0.39 MΩ	Fluke 5522A/11 Multiproduct Calibrator
Resistance – Source <sup>1</sup> (Simulated)	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	40 μΩ 95 μΩ/Ω 95 μΩ/Ω 23 μΩ/Ω 23 μΩ/Ω 10 μΩ/Ω 10 μΩ/Ω 6.5 μΩ/Ω 6.5 μΩ/Ω 6.5 μΩ/Ω 6.5 μΩ/Ω 8.5 μΩ/Ω 8.5 μΩ/Ω 13 μΩ/Ω 18 μΩ/Ω 40 μΩ/Ω 47 μΩ/Ω 0.1 mΩ/Ω	Fluke 5730A/03 Multiproduct Calibrator
Resistance – Source <sup>1</sup> (Artifact)	1 Ω 10 kΩ	7.1 μΩ 43 mΩ	Fluke 742A-1 Fluke 742A-10k Resistance Standards
High Resistance – Source <sup>1</sup> (Artifact)	(1 to 1 000) MΩ (1 to 100) GΩ	0.2 % of reading 1 % of reading	Biddle 72-6346-1 Decade Resistor



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Resistance – Measure <sup>1</sup> Normal Mode – 4W	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω (0.2 to 2) kΩ (2 to 20) kΩ (20 to 200) kΩ (0.2 to 2) MΩ (2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ	17 μΩ/Ω + 3.9 μΩ 9.5 μΩ/Ω + 14 μΩ 7.8 μΩ/Ω + 47 μΩ 8.1 μΩ/Ω + 0.47 mΩ 7.8 μΩ/Ω + 4.7 mΩ 8.6 μΩ/Ω + 47 mΩ 9.5 μΩ/Ω + 93 mΩ 27 μΩ/Ω + 9.3 Ω 0.12 mΩ/Ω + 0.93 kΩ 0.14 % of reading + 93 kΩ	Fluke 8508A opt 001 8.5 Digit Multimeter
DC Resistance – Measure <sup>1</sup> Low Current Mode – 4W	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω (0.2 to 2) kΩ (2 to 20) kΩ (20 to 200) kΩ (0.2 to 2) MΩ (2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ	17 μΩ/Ω + 3.9 μΩ 9.4 μΩ/Ω + 14 μΩ 8 μΩ/Ω + 0.14 mΩ 8.2 μΩ/Ω + 1.4 mΩ 7.9 μΩ/Ω + 14 mΩ 8.6 μΩ/Ω + 93 mΩ 21 μΩ/Ω + 0.93 Ω 88 μΩ/Ω + 93 Ω 0.14 % of reading + 93 kΩ 0.14 % of reading + 0.93 MΩ	Fluke 8508A opt 001 8.5 Digit Multimeter
DC Resistance – Measure <sup>1</sup> High Voltage Mode – 4W	(2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ (2 to 20) GΩ	25 μΩ/Ω + 9.3 Ω 70 μΩ/Ω + 0.93 kΩ 0.19 mΩ/Ω + 93 kΩ 0.14 % of reading + 9.3 MΩ	Fluke 8508A opt 001 8.5 Digit Multimeter
Capacitance – Source <sup>1</sup>	(0.22 to 0.4) nF (0.4 to 1.1) nF (1.1 to 3.3) nF (3.3 to 11) nF (11 to 33) nF (33 to 110) nF (110 to 330) nF (0.33 to 1.1) μF (1.1 to 3.3) μF (3.3 to 11) μF (11 to 33) μF (33 to 110) μF (110 to 330) μF	0.4 % of reading + 7.8 pF 0.4 % of reading + 7.8 pF 0.4 % of reading + 7.8 pF 0.21 % of reading + 7.8 pF 0.2 % of reading + 78 pF 0.21 % of reading + 78 pF 0.2 % of reading + 0.23 nF 0.21 % of reading + 0.78 nF 0.21 % of reading + 2.3 nF 0.2 % of reading + 7.8 nF 0.32 % of reading + 23 nF 0.37 % of reading + 78 nF 0.38 % of reading + 0.23 μF	Fluke 5522A/11 Multiproduct Calibrator





ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source <sup>1</sup>	(0.33 to 1.1) mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF	0.35 % of reading + 0.78 μF 0.35 % of reading + 2.3 μF 0.35 % of reading + 7.8 μF 0.58 % of reading + 23 μF 0.85 % of reading + 78 μF	Fluke 5522A/11 Multiproduct Calibrator
Capacitance – Measure <sup>1,7</sup>	1 kHz 400 pF to 25 μF (25 to 100) μF (100 to 500) μF (0.5 to 1) mF	0.02 % of reading 0.05 % of reading 0.05 % of reading 0.09 % of reading	General Radio 1689 RLC Digibridge
Inductance – Measure <sup>1,7</sup>	1 kHz 1 mH to 10 H	0.02 % of reading	General Radio 1689 RLC Digibridge
AC Voltage – Source <sup>1</sup>	Up to 2.2 mV (10 to 20) Hz (20 to 40) Hz (0.04 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz (0.04 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz (22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.024 % of reading + 4 μV 0.009 % of reading + 4 μV 0.008 % of reading + 4 μV 0.02 % of reading + 4 μV 0.05 % of reading + 5 μV 0.11 % of reading + 10 μV 0.14 % of reading + 20 μV 0.27 % of reading + 20 μV 0.024 % of reading + 4 μV 0.009 % of reading + 4 μV 0.008 % of reading + 4 μV 0.02 % of reading + 4 μV 0.05 % of reading + 5 μV 0.11 % of reading + 10 μV 0.14 % of reading + 20 μV 0.27 % of reading + 20 μV 0.024 % of reading + 12 μV 0.009 % of reading + 7 μV 0.0057 % of reading + 7 μV 0.012 % of reading + 7 μV 0.031 % of reading + 17 μV 0.066 % of reading + 20 μV 0.14 % of reading + 25 μV 0.27 % of reading + 45 μV	Fluke 5730A/03 Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup>	(0.22 to 2.2) V		Fluke 5730A/03 Multiproduct Calibrator
	(10 to 20) Hz	0.024 % of reading + 40 μV	
	(20 to 40) Hz	0.009 % of reading + 15 μV	
	40 Hz to 20 kHz	0.004 2 % of reading + 8 μV	
	(20 to 50) kHz	0.006 7 % of reading + 10 μV	
	(50 to 100) kHz	0.008 5 % of reading + 30 μV	
	(100 to 300) kHz	0.034 % of reading + 80 μV	
	(300 to 500) kHz	0.1 % of reading + 0.2 mV	
	500 kHz to 1 MHz	0.17 % of reading + 0.3 mV	
	(2.2 to 22) V		
	(10 to 20) Hz	0.024 % of reading + 0.4 mV	
	(20 to 40) Hz	0.009 % of reading + 0.15 mV	
	40 Hz to 20 kHz	0.004 2 % of reading + 0.05 mV	
	(20 to 50) kHz	0.006 7 % of reading + 0.1 mV	
	(50 to 100) kHz	0.008 3 % of reading + 0.2 mV	
	(100 to 300) kHz	0.026 % of reading + 0.6 mV	
	(300 to 500) kHz	0.1 % of reading + 2 mV	
	500 kHz to 1 MHz	0.15 % of reading + 3.2 mV	
	(22 to 220) V		
	(10 to 20) Hz	0.024 % of reading + 4 mV	
(20 to 40) Hz	0.009 % of reading + 1.5 mV		
40 Hz to 20 kHz	0.005 2 % of reading + 0.6 mV		
(20 to 50) kHz	0.008 % of reading + 1 mV		
(50 to 100) kHz	0.015 % of reading + 2.5 mV		
(100 to 300) kHz	0.09 % of reading + 16 mV		
(300 to 500) kHz	0.44 % of reading + 40 mV		
500 kHz to 1 MHz	0.8 % of reading + 80 mV		
(220 to 250) V			
(15 to 50) Hz	0.03 % of reading + 16 mV		
(250 to 1 100) V			
50 Hz to 1 kHz	0.007 % of reading + 3.5 mV		
AC Voltage – Source <sup>1</sup>	(220 to 330) V		Fluke 5522A/11 Multiproduct Calibrator
	(1 to 10) kHz	0.016 % of reading + 4.7 mV	
	(10 to 20) kHz	0.02 % of reading + 4.7 mV	
	(20 to 50) kHz	0.025 % of reading + 4.7 mV	
	(50 to 100) kHz	0.16 % of reading + 39 mV	
	(330 to 1 020) V		
	(1 to 5) kHz	0.02 % of reading + 7.8 mV	
(5 to 10) kHz	0.023 % of reading + 7.8 mV		





ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup> (Wideband Amplitude)	30 Hz to 500 kHz (0.3 to 1.1) mV (1.1 to 3) mV (3 to 11) mV (11 to 33) mV (33 to 110) mV (110 to 330) mV (0.33 to 1.1) V (1.1 to 3.5) V	0.62 % of reading + 0.78 μV 0.54 % of reading + 1.2 μV 0.54 % of reading + 3.1 μV 0.47 % of reading + 6.2 μV 0.47 % of reading + 16 μV 0.39 % of reading + 39 μV 0.39 % of reading + 0.16 mV 0.31 % of reading + 0.19 mV	Fluke 5730A/03 Multiproduct Calibrator
Wideband Amplitude Flatness – Source <sup>1</sup> (1 kHz reference)	Up to 1.1 mV (10 to 30) Hz (30 to 119.99) Hz 120 Hz to 1.199 9 kHz (1.2 to 11.999) kHz (12 to 119.99) kHz 120 kHz to 1.199 9 MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (1.1 to 3.3) mV (10 to 30) Hz (30 to 119.99) Hz 120 Hz to 1.199 9 kHz (1.2 to 11.999) kHz (12 to 119.99) kHz 120 kHz to 1.199 9 MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.23 % of reading 0.078 % of reading 0.078 % of reading 0.078 % of reading 0.078 % of reading 0.16 % of reading + 1.2 μV 0.16 % of reading + 1.2 μV 0.31 % of reading + 1.2 μV 0.47 % of reading + 1.2 μV 2.1 % of reading + 5.8 μV 0.23 % of reading 0.078 % of reading 0.078 % of reading 0.078 % of reading 0.078 % of reading 0.078 % of reading + 1.2 μV 0.078 % of reading + 1.2 μV 0.23 % of reading + 1.2 μV 0.39 % of reading + 1.2 μV 1.2 % of reading + 1.2 μV	Fluke 5730A/03 Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Wideband Amplitude Flatness – Source <sup>1</sup> (1 kHz reference)	(3.3 to 11) mV		Fluke 5730A/03 Multiproduct Calibrator
	(10 to 30) Hz	0.23 % of reading	
	(30 to 119.99) Hz	0.078 % of reading	
	120 Hz to 1.199 9 kHz	0.078 % of reading	
	(1.2 to 11.999) kHz	0.078 % of reading	
	(12 to 119.99) kHz	0.078 % of reading	
	120 kHz to 1.199 9 MHz	0.078 % of reading + 1.2 μV	
	(1.2 to 2) MHz	0.078 % of reading + 1.2 μV	
	(2 to 12) MHz	0.16 % of reading + 1.2 μV	
	(12 to 20) MHz	0.31 % of reading + 1.2 μV	
	(20 to 30) MHz	0.78 % of reading + 1.2 μV	
	(11 to 33) mV		
	(10 to 30) Hz	0.23 % of reading	
	(30 to 119.99) Hz	0.078 % of reading	
	120 Hz to 1.199 9 kHz	0.078 % of reading	
	(1.2 to 11.999) kHz	0.078 % of reading	
	(12 to 119.99) kHz	0.078 % of reading	
	120 kHz to 1.199 9 MHz	0.078 % of reading + 1.2 μV	
	(11 to 16.5) mV		
	(1.2 to 2) MHz	0.16 % reading + 1.2 μV	
	(2 to 11.9) MHz	0.23 % reading + 1.2 μV	
	(12 to 20) MHz	0.39 % reading + 1.2 μV	
	(20 to 30) MHz	0.85 % reading + 1.2 μV	
	(16.5 to 33) mV		
	(1.2 to 2) MHz	0.078 % reading + 1.2 μV	
	(2 to 11.9) MHz	0.16 % reading + 1.2 μV	
	(12 to 20) MHz	0.31 % reading + 1.2 μV	
	(20 to 30) MHz	0.78 % reading + 1.2 μV	
	(33 to 110) mV		
	(10 to 30) Hz	0.23 % of reading	
(30 to 119.99) Hz	0.078 % of reading		
120 Hz to 1.199 9 kHz	0.078 % of reading		
(1.2 to 11.999) kHz	0.078 % of reading		
(12 to 119.99) kHz	0.078 % of reading		
120 kHz to 1.199 9 MHz	0.078 % of reading + 1.2 μV		
(33 to 55) mV			
(1.2 to 2) MHz	.16 % reading + 1.2 μV		
(2 to 11.9) MHz	0.23 % reading + 1.2 μV		
(12 to 20) MHz	0.39 % reading + 1.2 μV		
(20 to 30) MHz	0.85 % reading + 1.2 μV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Wideband Amplitude Flatness – Source <sup>1</sup> (1 kHz reference)	(55 to 110) mV		Fluke 5730A/03 Multiproduct Calibrator
	(1.2 to 2) MHz	0.078 % reading + 1.2 μV	
	(2 to 11.9) MHz	0.16 % reading + 1.2 μV	
	(12 to 20) MHz	0.31 % reading + 1.2 μV	
	(20 to 30) MHz	0.78 % reading + 1.2 μV	
	(110 to 330) mV		
	(10 to 30) Hz	0.23 % of reading	
	(30 to 119.99) Hz	0.078 % of reading	
	120 Hz to 1.199 9 kHz	0.078 % of reading	
	(1.2 to 11.999) kHz	0.078 % of reading	
	(12 to 119.99) kHz	0.078 % of reading	
	120 kHz to 1.199 9 MHz	0.078 % of reading + 1.2 μV	
	(110 to 165) mV		
	(1.2 to 2) MHz	0.16 % reading + 1.2 μV	
	(2 to 11.9) MHz	0.23 % reading + 1.2 μV	
	(12 to 20) MHz	0.39 % reading + 1.2 μV	
	(20 to 30) MHz	0.85 % reading + 1.2 μV	
	(165 to 330) mV		
	(1.2 to 2) MHz	0.078 % reading + 1.2 μV	
	(2 to 11.9) MHz	0.16 % reading + 1.2 μV	
	(12 to 20) MHz	0.31 % reading + 1.2 μV	
	(20 to 30) MHz	0.78 % reading + 1.2 μV	
	(0.33 to 1.1) V		
	(10 to 30) Hz	0.23 % of reading	
	(30 to 119.99) Hz	0.078 % of reading	
	120 Hz to 1.199 9 kHz	0.078 % of reading	
	(1.2 to 11.999) kHz	0.078 % of reading	
	(12 to 119.99) kHz	0.078 % of reading	
120 Hz to 1.199 9 MHz	0.078 % of reading + 1.2 μV		
(0.33 to 0.55) V			
(1.2 to 2) MHz	0.16 % reading + 1.2 μV		
(2 to 11.9) MHz	0.23 % reading + 1.2 μV		
(12 to 20) MHz	0.39 % reading + 1.2 μV		
(20 to 30) MHz	0.85 % reading + 1.2 μV		
(0.55 to 1.1) V			
(1.2 to 2) MHz	0.078 % reading + 1.2 μV		
(2 to 11.9) MHz	0.16 % reading + 1.2 μV		
(12 to 20) MHz	0.31 % reading + 1.2 μV		
(20 to 30) MHz	0.78 % reading + 1.2 μV		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Wideband Amplitude Flatness – Source <sup>1</sup> (1 kHz reference)	(1.1 to 3.5) V		Fluke 5730A/03 Multiproduct Calibrator
	(10 to 30) Hz	0.23 % of reading	
	(30 to 119.99) Hz	0.078 % of reading	
	120 Hz to 1.199 9 kHz	0.078 % of reading	
	(1.2 to 11.999) kHz	0.078 % of reading	
	(12 to 119.99) kHz	0.078 % of reading	
	120 kHz to 1.199 9 MHz	0.078 % of reading + 1.2 μV	
	(1.1 to 1.75) V		
	(1.2 to 2) MHz	0.16 % reading + 1.2 μV	
	(2 to 11.9) MHz	0.23 % reading + 1.2 μV	
	(12 to 20) MHz	0.39 % reading + 1.2 μV	
	(20 to 30) MHz	0.85 % reading + 1.2 μV	
AC Voltage – Measure <sup>1</sup>	(1.75 to 3.5) V		Fluke 8508A opt 001 8.5 Digit Multimeter
	(1.2 to 2) MHz	0.078 % reading + 1.2 μV	
	(2 to 11.9) MHz	0.16 % reading + 1.2 μV	
	(12 to 20) MHz	0.31 % reading + 1.2 μV	
	(20 to 30) MHz	0.78 % reading + 1.2 μV	
	Up to 200 mV		
	(1 to 10) Hz	0.017 % of reading + 70 μV	
	(0.2 to 200) V		
	(1 to 10) Hz	0.015 % of reading + 60 μV	
	(200 to 1 000) V		
	(1 to 10) Hz	0.015 % of reading + 70 μV	
	AC Voltage – Measure <sup>1</sup>	(0.6 to 2.2) mV	
(10 to 20) Hz		0.17 % of reading + 1.3 μV	
(20 to 40) Hz		0.074 % of reading + 1.3 μV	
40 Hz to 20 kHz		0.042 % of reading + 1.3 μV	
(20 to 50) kHz		0.081 % of reading + 2 μV	
(50 to 100) kHz		0.12 % of reading + 2.5 μV	
(100 to 300) kHz		0.23 % of reading + 4 μV	
(300 to 500) kHz		0.24 % of reading + 8 μV	
500 kHz to 1 MHz		0.35 % of reading + 8 μV	
(2.2 to 7) mV			
(10 to 20) Hz		0.085 % of reading + 1.3 μV	
(20 to 40) Hz		0.037 % of reading + 1.3 μV	
40 Hz to 20 kHz		0.021 % of reading + 1.3 μV	
(20 to 50) kHz		0.04 % of reading + 2 μV	
(50 to 100) kHz		0.06 % of reading + 2.5 μV	
(100 to 300) kHz		0.12 % of reading + 4 μV	
(300 to 500) kHz		0.13 % of reading + 8 μV	
500 kHz to 1 MHz		0.23% of reading + 8 μV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(7 to 22) mV		Fluke 5790A-03 AC Measurement Standard
	(10 to 20) Hz	0.029 % of reading + 1.3 μV	
	(20 to 40) Hz	0.019 % of reading + 1.3 μV	
	40 Hz to 20 kHz	0.011 % of reading + 1.3 μV	
	(20 to 50) kHz	0.021 % of reading + 2 μV	
	(50 to 100) kHz	0.031 % of reading + 2.5 μV	
	(100 to 300) kHz	0.081 % of reading + 4 μV	
	(300 to 500) kHz	0.089 % of reading + 8 μV	
	500 kHz to 1 MHz	0.17 % of reading + 8 μV	
	(22 to 70) mV		
	(10 to 20) Hz	0.024 % of reading + 1.5 μV	
	(20 to 40) Hz	0.012 % of reading + 1.5 μV	
	40 Hz to 20 kHz	65 μV/V + 1.5 μV	
	(20 to 50) kHz	0.013 % of reading + 2 μV	
	(50 to 100) kHz	0.026 % of reading + 2.5 μV	
	(100 to 300) kHz	0.051 % of reading + 4 μV	
	(300 to 500) kHz	0.067 % of reading + 8 μV	
	500 kHz to 1 MHz	0.11 % of reading + 8 μV	
	(70 to 220) mV		
	(10 to 20) Hz	0.021 % of reading + 1.5 μV	
	(20 to 40) Hz	85 μV/V + 1.5 μV	
	40 Hz to 20 kHz	38 μV/V + 1.5 μV	
	(20 to 50) kHz	69 μV/V + 2 μV	
	(50 to 100) kHz	0.016 % of reading + 2.5 μV	
	(100 to 300) kHz	0.025 % of reading + 4 μV	
	(300 to 500) kHz	0.038 % of reading + 8 μV	
	500 kHz to 1 MHz	0.1 % of reading + 8 μV	
	(220 to 700) mV		
	(10 to 20) Hz	0.021 % of reading + 1.5 μV	
	(20 to 40) Hz	76 μV/V + 1.5 μV	
40 Hz to 20 kHz	33 μV/V + 1.5 μV		
(20 to 50) kHz	51 μV/V + 2 μV		
(50 to 100) kHz	79 μV/V + 2.5 μV		
(100 to 300) kHz	0.018 % of reading + 4 μV		
(300 to 500) kHz	0.03 % of reading + 8 μV		
500 kHz to 1 MHz	0.096 % of reading + 8 μV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(0.7 to 2.2) V		Fluke 5790A-03 AC Measurement Standard
	(10 to 20) Hz	0.021 % of reading + 1.5 μV	
	(20 to 40) Hz	76 μV/V + 1.5 μV	
	40 Hz to 20 kHz	33 μV/V + 1.5 μV	
	(20 to 50) kHz	51 μV/V + 2 μV	
	(50 to 100) kHz	79 μV/V + 2.5 μV	
	(100 to 300) kHz	0.018 % of reading + 4 μV	
	(300 to 500) kHz	0.03 % of reading + 8 μV	
	500 kHz to 1 MHz	0.096 % of reading + 8 μV	
	(2.2 to 7) V		
	(10 to 20) Hz	0.02 % of reading	
	(20 to 40) Hz	67 μV/V	
	40 Hz to 20 kHz	24 μV/V	
	(20 to 50) kHz	48 μV/V	
	(50 to 100) kHz	81 μV/V	
	(100 to 300) kHz	0.02 % of reading	
	(300 to 500) kHz	0.04 % of reading	
	500 kHz to 1 MHz	0.12 % of reading	
	(7 to 22) V		
	(10 to 20) Hz	0.02 % of reading	
	(20 to 40) Hz	67 μV/V	
	40 Hz to 20 kHz	27 μV/V	
	(20 to 50) kHz	48 μV/V	
	(50 to 100) kHz	81 μV/V	
	(100 to 300) kHz	0.02 % of reading	
	(300 to 500) kHz	0.04 % of reading	
	500 kHz to 1 MHz	0.12 % of reading	
	(22 to 70) V		
(10 to 20) Hz	0.02 % of reading		
(20 to 40) Hz	68 μV/V		
40 Hz to 20 kHz	32 μV/V		
(20 to 50) kHz	57 μV/V		
(50 to 100) kHz	94 μV/V		
(100 to 300) kHz	0.02 % of reading		
(300 to 500) kHz	0.04 % of reading		
500 kHz to 1 MHz	0.12 % of reading		





ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(70 to 220) V		Fluke 5790A-03 AC Measurement Standard
	(10 to 20) Hz	0.02 % of reading	
	(20 to 40) Hz	68 μV/V	
	40 Hz to 20 kHz	31 μV/V	
	(20 to 50) kHz	69 μV/V	
	(50 to 100) kHz	98 μV/V	
	(100 to 300) kHz	0.02 % of reading	
	(300 to 500) kHz	0.05 % of reading	
	(220 to 700) V		
	(10 to 20) Hz	0.02 % of reading	
	(20 to 40) Hz	99 μV/V	
	40 Hz to 20 kHz	41 μV/V	
	(20 to 50) kHz	0.01 % of reading	
	(50 to 100) kHz	0.05 % of reading	
Wideband Amplitude Flatness – Measure <sup>1</sup> (1 kHz reference)	(0.6 to 2.2) mV		Fluke 5790A-03 AC Measurement Standard
	50 kHz to 1.2 MHz	0.07 % of reading + 1 μV	
	(1.2 to 2) MHz	0.07 % of reading + 1 μV	
	(2 to 10) MHz	0.17 % of reading + 1 μV	
	(10 to 20) MHz	0.3 % of reading + 1 μV	
	(20 to 30) MHz	0.7 % of reading + 2 μV	
	(2.2 to 7) mV		
	50 kHz to 1.2 MHz	0.07 % of reading + 1 μV	
	(1.2 to 2) MHz	0.07 % of reading + 1 μV	
	(2 to 10) MHz	0.1 % of reading + 1 μV	
	(10 to 20) MHz	0.17 % of reading + 1 μV	
	(20 to 30) MHz	0.37 % of reading + 1 μV	
	(7 to 22) mV		
	50 kHz to 1.2 MHz	0.07 % of reading	
(1.2 to 2) MHz	0.07 % of reading		
(2 to 10) MHz	0.1 % of reading		
(10 to 20) MHz	0.17 % of reading		
(20 to 30) MHz	0.37 % of reading		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Wideband Amplitude Flatness – Measure <sup>1</sup> (1 kHz reference)	(22 to 70) mV		Fluke 5790A-03 AC Measurement Standard
	50 kHz to 1.2 MHz	0.05 % of reading	
	(1.2 to 2) MHz	0.05 % of reading	
	(2 to 10) MHz	0.1 % of reading	
	(10 to 20) MHz	0.15 % of reading	
	(20 to 30) MHz	0.35 % of reading	
	(70 to 220) mV		
	50 kHz to 1.2 MHz	0.05 % of reading	
	(1.2 to 2) MHz	0.05 % of reading	
	(2 to 10) MHz	0.1 % of reading	
	(10 to 20) MHz	0.15 % of reading	
	(20 to 30) MHz	0.35 % of reading	
	(220 to 700) mV		
	50 kHz to 1.2 MHz	0.05 % of reading	
	(1.2 to 2) MHz	0.05 % of reading	
	(2 to 10) MHz	0.1 % of reading	
	(10 to 20) MHz	0.15 % of reading	
	(20 to 30) MHz	0.35 % of reading	
(0.7 to 2.2) V			
50 kHz to 1.2 MHz	0.05 % of reading		
(1.2 to 2) MHz	0.05 % of reading		
(2 to 10) MHz	0.1 % of reading		
(10 to 20) MHz	0.15 % of reading		
(20 to 30) MHz	0.35 % of reading		
(2.2 to 7) V			
50 kHz to 1.2 MHz	0.05 % of reading		
(1.2 to 2) MHz	0.05 % of reading		
(2 to 10) MHz	0.1 % of reading		
(10 to 20) MHz	0.15 % of reading		
(20 to 30) MHz	0.35 % of reading		
AC High Voltage – Measure <sup>1</sup>	(0.7 to 9) kV		Vitrek 4700 High Voltage Meter
	10 mHz to 10 Hz	0.15 % reading + 0.17 V	
	(10 to 30) Hz	0.15 % reading + 0.17 V	
	(30 to 50) Hz	0.14 % reading + 0.17 V	
	(50 to 70) Hz	0.14 % reading + 0.17 V	
	(70 to 100) Hz	0.14 % reading + 0.17 V	
	(100 to 200) Hz	0.14 % reading + 0.17 V	
	(200 to 450) Hz	0.48 % reading + 0.17 V	
(450 to 600) Hz	0.87 % reading + 0.17 V		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC High Voltage – Measure <sup>1</sup>	(9 to 10) kV 10 mHz to 10 Hz (10 to 30) Hz (30 to 50) Hz (50 to 70) Hz (70 to 100) Hz (100 to 200) Hz (200 to 450) Hz (450 to 600) Hz	0.15 % reading + 0.69 V 0.15 % reading + 0.69 V 0.14 % reading + 0.69 V 0.14 % reading + 0.69 V 0.14 % reading + 0.69 V 0.14 % reading + 0.69 V 0.48 % reading + 0.69 V 0.87 % reading + 0.69 V	Vitretek 4700 High Voltage Meter
AC Current – Source <sup>1</sup>	Up to 220 µA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.025 % of reading + 16 nA 0.016 % of reading + 10 nA 0.011 % of reading + 8 nA 0.028 % of reading + 12 nA 0.11 % of reading + 65 nA 0.025 % of reading + 40 nA 0.016 % of reading + 35 nA 0.011 % of reading + 35 nA 0.02 % of reading + 0.11 µA 0.11 % of reading + 0.65 µA 0.025 % of reading + 0.4 µA 0.016 % of reading + 0.35 µA 0.011 % of reading + 0.35 µA 0.02 % of reading + 0.55 µA 0.11 % of reading + 5 µA 0.025 % of reading + 4 µA 0.016 % of reading + 3.5 µA 0.011 % of reading + 2.5 µA 0.02 % of reading + 3.5 µA 0.11 % of reading + 10 µA 0.025 % of reading + 35 µA 0.045 % of reading + 80 µA 0.7 % of reading + 0.16 mA	Fluke 5730A/03 Multiproduct Calibrator

**Electrical – DC/Low Frequency**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
AC Current – Source <sup>1</sup>	(2.2 to 3) A (10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.18 % of reading + 0.1 mA 0.06 % of reading + 0.1 mA 0.6 % of reading + 1 mA 2.5 % of reading + 5 mA	Fluke 5730A/03 Multiproduct Calibrator
AC Current – Source <sup>1</sup>	(3 to 11) A (45 to 100) Hz (0.1 to 1) kHz (1 to 5) kHz (11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.049 % of reading + 1.6 mA 0.079 % of reading + 1.6 mA 2.3 % of reading + 1.6 mA 0.095 % of reading + 3.9 mA 0.12 % of reading + 3.9 mA 2.3 % of reading + 3.9 mA	Fluke 5522A/11 Multiproduct Calibrator
AC Clamp-on Ammeters (Toroidal Type) Transformer Type Sensor <sup>1</sup>	(20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz	0.31 % of reading + 26 mA 0.84 % of reading + 47 mA 0.35 % of reading + 0.12 A 1.2 % of reading + 0.22 A	Fluke 5522A/11 Multiproduct Calibrator, 50-turn Coil
AC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor <sup>1</sup>	(20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz	0.58 % of reading + 0.25 A 1.1 % of reading + 0.25 A 0.6 % of reading + 0.9 A 1.3 % of reading + 0.92 A	Fluke 5522A/11 Multiproduct Calibrator, 50-turn Coil
AC Current – Measure <sup>1</sup>	Up to 200 $\mu$ A 1 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz (0.2 to 2) mA (1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz (2 to 20) mA (1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz	0.031 % of reading + 20 nA 0.071 % of reading + 20 nA 0.4 % of reading + 20 nA 0.031 % of reading + 0.2 $\mu$ A 0.03 % of reading + 0.2 $\mu$ A 0.071 % of reading + 0.2 $\mu$ A 0.4 % of reading + 0.2 $\mu$ A 0.031 % of reading + 2 $\mu$ A 0.03 % of reading + 2 $\mu$ A 0.071 % of reading + 2 $\mu$ A 0.4 % of reading + 2 $\mu$ A	Fluke 8508A opt 001 8.5 Digit Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(20 to 200) mA 1 Hz to 10 Hz 10 Hz to 10 kHz (10 to 30) kHz (0.2 to 2) A 10 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (2 to 20) A 10 Hz to 2 kHz (2 to 10) kHz	0.031 % of reading + 20 μA 0.029 % of reading + 20 μA 0.063 % of reading + 20 μA 0.062 % of reading + 0.2 mA 0.074 % of reading + 0.2 mA 0.3 % of reading + 0.2 mA 0.082 % of reading + 2 mA 0.25 % of reading + 2 mA	Fluke 8508A opt 001 8.5 Digit Multimeter
AC Current – Measure <sup>1</sup>	(1 to 10) mA (5 to 400) Hz 400 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (10 to 30) mA (5 to 400) Hz 400 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (30 to 300) mA (5 to 400) Hz 400 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (0.3 to 3) A (5 to 400) Hz 400 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (3 to 10) A (5 to 400) Hz 400 Hz to 20 kHz (20 to 50) kHz (10 to 20) A (5 to 400) Hz 400 Hz to 20 kHz (20 to 50) kHz	0.002 % of reading 0.003 % of reading 0.004 % of reading 0.006 % of reading 0.002 % of reading 0.003 % of reading 0.005 % of reading 0.007 % of reading 0.003 % of reading 0.004 % of reading 0.007 % of reading 0.01 % of reading 0.002 % of reading 0.003 % of reading 0.005 % of reading 0.01 % of reading 0.002 % of reading 0.003 % of reading 0.005 % of reading 0.007 % of reading 0.012 % of reading 0.018 % of reading	Fluke 5790A-03 AC Measurement Standard, Fluke A40 Current Shunts



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Power – Source <sup>1</sup> 330 $\mu$ W to 330 mA	11 $\mu$ W to 1.1 mW (1.1 to 110) mW 110 mW to 110 W (110 to 330) W	0.024 % of reading 0.027 % of reading 0.024 % of reading 0.018 % of reading	Fluke 5520A Multiproduct Calibrator
330 mA to 3 A	11 $\mu$ W to 110 mW 110 mW to 990 W 990 W to 3 kW	0.044 % of reading 0.053 % of reading 0.01 % of reading	
(3 to 20.5) A	99 mW to 0.99 W 0.99 W to 6.8 kW (6.8 to 20.5) kW	0.088 % of reading 0.07 % of reading 0.04 % of reading	
AC Power – Source <sup>1,2,8</sup> PF = 1			Fluke 5520A Multiproduct Calibrator
(3.3 to 9) mA	(10 to 65) Hz 110 $\mu$ W to 3 mW 3 mW to 9 W	0.13 % of reading 0.077 % of reading	
(9 to 33) mA	(10 to 65) W 300 $\mu$ W to 10 mW 10 mW to 33 W	0.089 % of reading 0.077 % of reading	
(33 to 90) mA	(10 to 65) Hz (1 to 30) mW 30 mW to 90 W	0.071 % of reading 0.057 % of reading	
(90 to 330) mA	(10 to 65) Hz (3 to 100) mW 100 mW to 300 W	0.089 % of reading 0.078 % of reading	
(330 to 900) mA	(10 to 65) Hz (11 to 300) mW 300 mW to 900 W	0.071 % of reading 0.081 % of reading	
900 mA to 2.2 A	(10 to 65) Hz (30 to 720) mW 0.72 W to 2 kW	0.089 % of reading 0.079 % of reading	
(2.2 to 4.5) A	(10 to 65) Hz 80 mW to 1.4 W 1.4 W to 4.5 kW	0.088 % of reading 0.18 % of reading	
(4.5 to 20.5) A	(10 to 65) Hz 150 mW to 20kW	0.17 % of reading	



**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure	Type B		Fluke 5522A/11 Multiproduct Calibrator
	(600 to 800) °C	0.35 °C	
	(800 to 1 000) °C	0.28 °C	
	(1 000 to 1 550) °C	0.24 °C	
	(1 550 to 1 820) °C	0.26 °C	
	Type C		
	(0 to 150) °C	0.24 °C	
	(150 to 650) °C	0.21 °C	
	(650 to 1000) °C	0.24 °C	
	(1 000 to 1 800) °C	0.39 °C	
	(1 800 to 2 316) °C	0.65 °C	
	Type E		
	(-250 to -100) °C	0.39 °C	
	(-100 to -25) °C	0.13 °C	
	(-25 to 350) °C	0.12 °C	
	(350 to 650) °C	0.13 °C	
	(650 to 1 000) °C	0.17 °C	
	Type J		
	(-210 to -100) °C	0.21 °C	
	(-100 to -30) °C	0.13 °C	
	(-30 to 150) °C	0.12 °C	
	(150 to 760) °C	0.14 °C	
	(760 to 1 200) °C	0.18 °C	
	Type K		
(-200 to -100) °C	0.26 °C		
(-100 to -25) °C	0.15 °C		
(-25 to 120) °C	0.13 °C		
(120 to 1 000) °C	0.21 °C		
(1 000 to 1 372) °C	0.31 °C		
Type L			
(-200 to -100) °C	0.29 °C		
(-100 to 800) °C	0.21 °C		
(800 to 900) °C	0.14 °C		
Type N			
(-200 to -100) °C	0.31 °C		
(-100 to -25) °C	0.18 °C		
(-25 to 120) °C	0.15 °C		
(120 to 410) °C	0.15 °C		
(410 to 1 300) °C	0.21 °C		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure	Type R (0 to 250) °C (250 to 400) °C (400 to 1 000) °C (1 000 to 1 767) °C Type S (0 to 250) °C (250 to 1 000) °C (1 000 to 1 400) °C (1 400 to 1 767) °C Type T (-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C Type U (-200 to 0) °C (0 to 600) °C	0.46 °C 0.29 °C 0.26 °C 0.32 °C 0.45 °C 0.3 °C 0.29 °C 0.36 °C 0.49 °C 0.19 °C 0.13 °C 0.12 °C 0.44 °C 0.21 °C	Fluke 5522A/11 Multiproduct Calibrator
Electrical Simulation of RTD Indicating Devices – Source 1	Pt 385, 100 Ω (-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C Pt 385, 200 Ω (-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.039 °C 0.039 °C 0.054 °C 0.07 °C 0.078 °C 0.093 °C 0.18 °C 0.031 °C 0.031 °C 0.031 °C 0.039 °C 0.093 °C 0.1 °C 0.11 °C 0.12 °C	Fluke 5522A/11 Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices – Source 1	Pt 385, 500 Ω		Fluke 5522A/11 Multiproduct Calibrator
	(-200 to -80) °C	0.031 °C	
	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.039 °C	
	(100 to 260) °C	0.047 °C	
	(260 to 300) °C	0.062 °C	
	(300 to 400) °C	0.062 °C	
	(400 to 600) °C	0.07 °C	
	(600 to 630) °C	0.085 °C	
	Pt 385, 1 kΩ		
	(-200 to -80) °C	0.023 °C	
	(-80 to 0) °C	0.023 °C	
	(0 to 100) °C	0.031 °C	
	(100 to 260) °C	0.039 °C	
	(260 to 300) °C	0.047 °C	
	(300 to 400) °C	0.054 °C	
	(400 to 600) °C	0.054 °C	
	(600 to 630) °C	0.18 °C	
	Pt 3916, 100 Ω		
	(-200 to -190) °C	0.19 °C	
	(-190 to -80) °C	0.031 °C	
	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.047 °C	
	(100 to 260) °C	0.054 °C	
	(260 to 300) °C	0.062 °C	
	(300 to 400) °C	0.07 °C	
	(400 to 600) °C	0.078 °C	
	(600 to 630) °C	0.18 °C	
	Pt 3926, 100 Ω		
	(-200 to -80) °C	0.039 °C	
(-80 to 0) °C	0.039 °C		
(0 to 100) °C	0.054 °C		
(100 to 300) °C	0.07 °C		
(300 to 400) °C	0.078 °C		
(400 to 630) °C	0.093 °C		
PtNi 385, 120 Ω			
(-80 to 0) °C	0.062 °C		
(0 to 100) °C	0.062 °C		
(100 to 260) °C	0.1 °C		
Cu 427, 10 Ω			
(-100 to 260) °C	0.23 °C		





ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes <sup>1,2</sup> Leveled Sine Wave – Generate into 50 Ω load	5 mVp-p to 5.5 Vp-p 50 kHz 100 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mVp-p to 3.5 Vp-p 600 MHz to 1.1 GHz	1.8 % of reading + 0.23 mV 2.8 % of reading + 0.23 mV 3.2 % of reading + 0.23 mV 4 % of reading + 0.23 mV 5.5 % of reading + 0.23 mV	Fluke 5522A/11 Multiproduct Calibrator
Bandwidth/Flatness – Measure (50 kHz Reference) into 50 Ω load	5 mVp-p to 5.5 Vp-p 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mVp-p to 3.5 Vp-p 600 MHz to 1.1 GHz	1.4 % of reading + 78 μV 1.8 % of reading + 78 μV 3.2 % of reading + 78 μV 4 % of reading + 78 μV	
Input Impedance – Measure into 50 Ω load into 1 MΩ load	(40 to 60) Ω (0.5 to 1.5) MΩ	0.082 % of reading 0.081 % of reading	
Input Capacitance – Measure	(5 to 50) pF	3.9 % of reading + 0.39 pF	
Wave Generator – Source Amplitude (Sine, Square, Triangle) into 50 Ω load into 1 MΩ load	10 Hz to 10 kHz 1.8 mVp-p to 2.5 Vp-p 1.8 mVp-p to 55 Vp-p	2.3 % of reading + 78 μV 2.3 % of reading + 78 μV	
Frequency	10 Hz to 10 kHz	0.001 9 % of reading + 12 mHz	
Pulse Characterization Rise Time – Measure	30 ps to 1 μs	21 ps	



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Absolute Power – Source <sup>1,3</sup>	(-35 to 14) dBm 100 MHz to 2.4 GHz (2.4 to 8) GHz (8 to 18) GHz (18 to 26.5) GHz	0.08 dB 0.11 dB 0.14 dB 0.17 dB	Fluke 96720A RF Reference Source, R&S NRP-Z55(x2) Power Sensor, Agilent 11667B Power Splitter, Sucoflex 102EA 40 GHz Test Cable
RF Absolute Power – Measure <sup>1</sup>	1 mW Reference, 50 MHz	0.03 % of reading	HP 8478B Power Sensor, HP 432A Power Meter
RF Absolute Power – Measure <sup>1,3</sup>	(-65 to -35) dBm (> 0.01 to ≤ 0.03) GHz (> 0.03 to ≤ 4.00) GHz (> 4.00 to ≤ 8.00) GHz (> 8.00 to ≤ 10.00) GHz (> 10.00 to ≤ 13.00) GHz (> 13.00 to ≤ 15.00) GHz (> 15.00 to ≤ 18.00) GHz	2.84 % of reading 1.9 % of reading 2.34 % of reading 2.44 % of reading 2.98 % of reading 3.48 % of reading 3.84 % of reading	HP 8484A Power Sensor, Agilent E4419B Power Meter
RF Absolute Power – Measure <sup>1,3</sup>	(-35 to 20) dBm DC to 100 MHz (> 0.1 to ≤ 2.4) GHz (> 2.4 to ≤ 8) GHz (> 8 to ≤ 12.4) GHz (> 12.4 to ≤ 18) GHz (> 18 to ≤ 26.5) GHz (> 26.5 to ≤ 33) GHz (> 33 to ≤ 40) GHz	0.04 dB 0.048 dB 0.054 dB 0.063 dB 0.082 dB 0.086 dB 0.11 dB 0.11 dB	Fluke 96720A RF Reference Source, R&S NRP-Z55(x2) Power Sensor
Tuned RF Absolute Power – Measure <sup>3</sup>	2.5 MHz to 26.5 GHz (-127 to -110) dBm (-110 to -90) dBm (-90 to -80) dBm (-80 to -50) dBm (-50 to -40) dBm (-40 to -10) dBm (-10 to 0) dBm (0 to 10) dBm	0.54 dB 0.39 dB 0.37 dB 0.34 dB 0.33 dB 0.31 dB 0.3 dB 0.3 dB	HP 8902A Measuring Receiver, HP 11722A Power Sensor, HP 11792A Microwave Converter, HP 11793A Microwave Converter



**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Amplitude Modulation – AM Depth Measure <sup>1</sup> Rate: 50 Hz to 10 kHz	150 kHz to 10 MHz (5 to 40) % Depth (40 to 99) % Depth	0.85 % Depth 2.3 % Depth	HP 8902A Measuring Receiver, HP 11722A Power Sensor, HP 11792A Microwave Converter, HP 11793A Microwave Converter
	Rate: (20 to 50) Hz 150 kHz to 10 MHz (5 to 40) % Depth (40 to 99) % Depth	1.3 % Depth 3.3 % Depth	
	Rate: 50 Hz to 50 kHz 10 MHz to 1.3 GHz (5 to 40) % Depth (40 to 99) % Depth	0.45 % Depth 1.3 % Depth	
	Rate: 50 Hz to 50 kHz (1.3 to 26.5) GHz (5 to 40) % Depth (40 to 99) % Depth	0.65 % Depth 1.8 % Depth	
	Rate: (20 to 50) Hz or (50 to 100) kHz 10 MHz to 26.5 GHz (5 to 40) % Depth (40 to 99) % Depth	1.3 % Depth 3.3 % Depth	
Frequency Modulation – Measure <sup>1,7</sup> Rate: 20 Hz to 10 kHz	250 kHz to 10 MHz Dev: ≤ 40 kHz pk	2 % of reading	HP 8902A Measuring Receiver, HP 11722A Power Sensor, HP 11792A Microwave Converter, HP 11793A Microwave Converter
	Rate: 50 Hz to 100 kHz 10 MHz to 1.3 GHz Dev: ≤ 400 kHz pk 10 MHz to 26.5 GHz Dev: ≤ 400 kHz pk	1 % of reading 1 % of reading	
	Rate: 20 Hz to 200 kHz 10 MHz to 1.3 GHz Dev: ≤ 400 kHz pk 10 MHz to 26.5 GHz Dev: ≤ 400 kHz pk	6 % of reading 6 % of reading	
	Phase Modulation – Measure <sup>1,7</sup> Rate: 200 Hz to 10 kHz	150 kHz to 10 MHz	
Rate: 200 Hz to 20 kHz	10 MHz to 26.5 GHz	4 % of reading	



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Power Range Accuracy – Measure	3 $\mu$ W to 100 mW	0.31 % of reading	HP 11683A Power Meter
Distortion – Measure	20 Hz to 20 kHz (-80 to 0) dB	1.2 dB	HP 8903B Audio Analyzer
	(20 to 100) kHz (-65 to 0) dB	2.4 dB	
RF Amplitude Frequency Response – Measure	9 kHz to 2.9 GHz (2.90 to 6.46) GHz (6.46 to 13) GHz (13 to 19.7) GHz (19.7 to 22) GHz	1.1 dB 1.5 dB 2.1 dB 3.1 dB 3.1 dB	HP 8562A Spectrum Analyzer
Leveled Sine Wave Output – Absolute Amplitude Accuracy Level <sup>3</sup>	10 Hz to 4 GHz (-130 to -94) dBm	1.1 dB	Fluke 96720A RF Reference Source, Fluke 96040A-50 Low Phase Noise Reference Source
	(-94 to -74) dBm	0.68 dB	
	(-74 to -17) dBm	0.34 dB	
	(-17 to 24) dBm	0.2 dB	
Leveled Sine Wave Output – Absolute Amplitude Accuracy Level <sup>3</sup>	10 Hz to 4 GHz (-120 to -100) dBm	1.1 dB	Fluke 96720A RF Reference Source, Fluke 96040A-75 Low Phase Noise Reference Source
	(-100 to -80) dBm	0.67 dB	
	(-80 to 18) dBm	0.34 dB	
S11/S22 Reflection Magnitude – Measure <sup>1,6</sup> (Linear)	50 MHz to 2 GHz $\leq 0.25$ lin	0.013	HP 8722ES Network Analyzer, Agilent 85056A Calibration Kit
	(> 0.25 to 0.5) lin	0.012	
	(> 0.5 to $\leq 0.7$ ) lin	0.012	
	(> 0.7 to $\leq 1$ ) lin	0.013	
	(2 to 8) GHz $\leq 0.25$ lin	0.013	
	(> 0.25 to 0.5) lin	0.012	
S11/S22 Reflection Magnitude – Measure <sup>1,6</sup> (Linear)	(> 0.5 to $\leq 0.7$ ) lin	0.012	HP 8722ES Network Analyzer, Agilent 85056A Calibration Kit
	(> 0.7 to $\leq 1$ ) lin	0.014	
	(20 to 40) GHz $\leq 0.25$ lin	0.022	
	(> 0.25 to 0.5) lin	0.022	
	(> 0.5 to $\leq 0.7$ ) lin	0.023	
	(> 0.7 to $\leq 1$ ) lin	0.026	



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
S11/S22 Reflection Phase – Measure <sup>1</sup> (Linear)	50 MHz to 2 GHz		HP 8722ES Network Analyzer, Agilent 85056A Calibration Kit
	≤ 0.25 lin	2.7 °	
	(> 0.25 to 0.5) lin	1.4 °	
	(> 0.5 to ≤ 0.7) lin	1 °	
	(> 0.7 to ≤ 1) lin	0.9 °	
	(2 to 8) GHz		
	≤ 0.25 lin	2.7 °	
	(> 0.25 to 0.5) lin	1.4 °	
	(> 0.5 to ≤ 0.7) lin	1.1 °	
	(> 0.7 to ≤ 1) lin	1.1 °	
	(8 to 20) GHz		
	≤ 0.25 lin	2.7 °	
	(> 0.25 to 0.5) lin	1.4 °	
	(> 0.5 to ≤ 0.7) lin	1.1 °	
(> 0.7 to ≤ 1) lin	1.1 °		
(20 to 40) GHz			
≤ 0.25 lin	4.9 °		
(> 0.25 to 0.5) lin	2.6 °		
(> 0.5 to ≤ 0.7) lin	2.1 °		
(> 0.7 to ≤ 1) lin	2 °		
S21/S12 Transmission Magnitude – Measure <sup>1</sup> (dB)	50 MHz to 2 GHz		HP 8722ES Network Analyzer, Agilent 85056A Calibration Kit
	(-90 to ≤ -80) dB	0.35 dB	
	(-80 to ≤ -70) dB	0.18 dB	
	(-70 to ≤ -60) dB	0.15 dB	
	(-60 to ≤ -50) dB	0.15 dB	
	(-50 to ≤ -40) dB	0.14 dB	
	(-40 to ≤ -30) dB	0.14 dB	
	(-30 to ≤ -20) dB	0.14 dB	
	(-20 to ≤ -10) dB	0.14 dB	
	(-10 to ≤ 0) dB	0.14 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
S21/S12 Transmission Magnitude – Measure <sup>1</sup> (dB)	(2 to 8) GHz		HP 8722ES Network Analyzer, Agilent 85056A Calibration Kit
	(-90 to ≤ -80) dB	0.36 dB	
	(-80 to ≤ -70) dB	0.19 dB	
	(-70 to ≤ -60) dB	0.17 dB	
	(-60 to ≤ -50) dB	0.17 dB	
	(-50 to ≤ -40) dB	0.16 dB	
	(-40 to ≤ -30) dB	0.16 dB	
	(-30 to ≤ -20) dB	0.16 dB	
	(-20 to ≤ -10) dB	0.16 dB	
	(-10 to ≤ 0) dB	0.16 dB	
	(8 to 20) GHz		
	(-90 to ≤ -80) dB	0.46 dB	
	(-80 to ≤ -70) dB	0.25 dB	
	(-70 to ≤ -60) dB	0.21 dB	
	(-60 to ≤ -50) dB	0.21 dB	
	(-50 to ≤ -40) dB	0.21 dB	
	(-40 to ≤ -30) dB	0.2 dB	
	(-30 to ≤ -20) dB	0.2 dB	
	(-20 to ≤ -10) dB	0.2 dB	
	(-10 to ≤ 0) dB	0.2 dB	
	(20 to 40) GHz		
	(-90 to ≤ -80) dB	0.84 dB	
	(-80 to ≤ -70) dB	0.44 dB	
	(-70 to ≤ -60) dB	0.37 dB	
	(-60 to ≤ -50) dB	0.36 dB	
	(-50 to ≤ -40) dB	0.36 dB	
	(-40 to ≤ -30) dB	0.36 dB	
(-30 to ≤ -20) dB	0.36 dB		
(-20 to ≤ -10) dB	0.36 dB		
(-10 to ≤ 0) dB	0.36 dB		

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
S21/S12 Transmission Phase – Measure <sup>1</sup>	50 MHz to 2 GHz		HP 8722ES Network Analyzer, Agilent 85056A Calibration Kit
	(-90 to ≤ -40) dB	180 °	
	(-40 to ≤ -30) dB	30 °	
	(-30 to ≤ -20) dB	9.2 °	
	(-20 to ≤ -10) dB	2.9 °	
	(-10 to ≤ 0) dB	1 °	
	(2 to 8) GHz		
	(-80 to ≤ -40) dB	180 °	
	(-40 to ≤ -30) dB	35 °	
	(-30 to ≤ -20) dB	10 °	
	(-20 to ≤ -10) dB	3.3 °	
	(-10 to ≤ 0) dB	1.2 °	
	(8 to 20) GHz		
	(-90 to ≤ -40) dB	180 °	
	(-40 to ≤ -30) dB	47 °	
	(-30 to ≤ -20) dB	13 °	
(-20 to ≤ -10) dB	4.2 °		
(-10 to ≤ 0) dB	1.5 °		
(20 to 40) GHz			
(-80 to ≤ -30) dB	180 °		
(-30 to ≤ -20) dB	24 °		
(-20 to ≤ -10) dB	7.4 °		
(-10 to ≤ 0) dB	2.6 °		

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Micrometers, Calipers (Outside, Inside, Depth and Step) <sup>1</sup>	(0.01 to 0.04) in (0.05 to 1) in (1 to 4) in (4 to 15) in (15 to 40) in	13 μin (13 + 1L) μin (9 + 4L) μin (12 + 4L) μin (16 + 4L) μin	Gage Blocks
Anvil Flatness <sup>1</sup>	Up to 1 in Diameter	14 μin	Optical Flats
Spindle Parallelism	Up to 1 in Diameter	14 μin	Optical Flats

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length – OD Measurement <sup>1,2</sup> (Single Axis)	Up to 1 in (1 to 4) in (4 to 15) in (15 to 20) in	(17 + 2L) μin (14 + 4L) μin (17 + 4L) μin (3 + 5L) μin	Universal Length Measuring Machine, Gage Blocks
Length – OD Measurement <sup>1,2</sup> (Single Axis)	(20 to 40) in	(64 + 4L) μin	Gage Amplifier/Check, Gage Blocks
Length – ID Measurement <sup>1,2</sup> (Single Axis)	Up to 1 in (1 to 4) in (4 to 7) in (7 to 14) in	(21 + 4L) μin (19 + 3L) μin (12 + 5L) μin (18 + 4L) μin	Universal Length Measuring Machine, Gage Blocks
Cylindrical Ring Gages <sup>2</sup> Inside Diameter	Up to 16 in	(14 + 4.6L) μin	Universal Length Measuring Machine, Master Ring Gages
Thread Ring Gages <sup>2</sup> Pitch Diameter	Up to 6 in	(22 + 4.7L) μin	Comparison to Set Plugs
Cylindrical Plug Gages <sup>2</sup> Outside Diameter	Up to 16 in	(15 + 4.5L) μin	Universal Length Measuring Machine, Gage Blocks
Thread Plug Gages Major Diameter (OD)	Up to 6 in	(16 + 3.7L) μin	Universal Length Measuring Machine, Thread Wires
Thread Plug Gages Pitch Diameter	Up to 6 in	(24 + 3L) μin	Universal Length Measuring Machine, Thread Wires
Surface Plates <sup>1,2</sup>  Overall Flatness	Up to 161 inDL	95 μin	In accordance with ASME B89.3.7 using Planekator
Local Area Flatness	Up to 0.001 in	29 μin	Repeat-O-Meter
Coating Thickness Gages (Eddy Current, Magnetic Induction, Fixed Point)	Up to 3 000 μm Up to 118 mils	(0.44 + 0.03L) μm (0.017 + 0.001L) mils	Universal Length Measuring Machine, Shims, Gage Blocks
Dial Indicators	Up to 1 in (1 to 4) in	(17 + 2L) μin (14 + 4L) μin	Universal Length Measuring Machine, Gage Blocks

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Air/Nitrogen Flow Meters	Up to 100 slpm	0.73 % of reading	CME FCS Laminar Flow Elements



**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Liquid Flow Meters	Up to 250 gpm	0.33 % of reading	FT-32 Turbine Flow System
Air Velocity Measuring Equipment	Up to 5 800 ft/min	1.7 % of reading	Alnor RVA Anemometer, Wind Tunnel
Low Pressure Gages	(0 to 2) inH <sub>2</sub> O	0.008 inH <sub>2</sub> O	Dwyer 1430 Microtector
Pneumatic Pressure Gages, Vacuum Gages <sup>4</sup>	-30 inHg to 1 000 psig	0.02 % of span	DHI PPC4EX-7M Pressure Controller
Pneumatic Pressure Gages, Vacuum Gages <sup>4</sup>	(0.1 to 1 000) psia	0.02 % of span + 0.007 psi	DHI PPC4EX-7M Pressure Controller
Pneumatic Pressure Gages	(0.2 to 718) psia (psig)	0.003 % of reading	Ruska 2465 Gas Piston Gauge
Hydraulic Pressure Gages	(100 to 50 000) psig	0.008 % of reading	DHI 5306 Liquid Piston Gauge
Pressure/Vacuum Gages <sup>1</sup> (Pneumatic and Hydraulic)	(0 to 900) mmHg (15 to 30) psig (30 to 300) psig (300 to 500) psig (500 to 3 000) psig (3 000 to 10 000) psig	0.23 mmHg 0.08 psi 0.15 psi 0.13 psi 1.5 psi 8 psi	Comparison to Meriam DAI0900 Fluke 700PD5 Fluke 700P27 Meriam DGI0500-11-1 Meriam DGI3000-11-1 Fluke 700P31
Scales and Balances <sup>1,5</sup>	Up to 500 mg (0.5 to 5) g (5 to 10) g (10 to 20) g (20 to 100) g (100 to 210) g	12 µg 40 µg 59 µg 89 µg 0.000 31 % of reading 0.000 41 % of reading	ASTM E617 Class 1 Weights and internal calibration procedure utilized for the calibration of the weighing system.
Scales and Balances <sup>1,5</sup>	100 g to 10 kg	0.000 59 % of reading	ASTM E617 Class 2 Weights and internal calibration procedure utilized for the calibration of the weighing system.
Scales and Balances <sup>1,5</sup>	Up to 7 g (7 to 453) g 453 g to 2.2 kg (2.2 to 454) kg	0.12 % of reading 0.024 % of reading 0.019 % of reading 0.012 % of reading	NIST Class F Weights and internal calibration procedure utilized for the calibration of the weighing system.



ANSI National Accreditation Board

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales and Balances <sup>1,5</sup>	Up to 30 lb (30 to 40) lb (40 to 70) lb (70 to 90) lb	0.012 % of reading 0.01 % of reading 0.009 % of reading 0.008 % of reading	ASTM E617 Class 6 Weights and internal calibration procedure utilized for the calibration of the weighing system.
Scales and Balances <sup>1,5</sup>	Up to 0.5 lb (0.5 to 1) lb (1 to 5) lb (5 to 1 000) lb	0.12 % of reading 0.024 % of reading 0.019 % of reading 0.012 % of reading	NIST Class F Weights and internal calibration procedure utilized for the calibration of the weighing system.
Force Measuring Equipment – Compression <sup>1</sup>	Up to 10 lbf (10 to 25) lbf (25 to 50) lbf (50 to 150) lbf (150 to 250) lbf (250 to 500) lbf	0.001 5 lbf 0.002 1 lbf 0.008 3 lbf 0.023 lbf 0.065 lbf 0.13 lbf	NIST Class F Weights
Force Measuring Equipment – Compression <sup>1</sup>	(500 to 1 000) lbf	0.77 lbf	Optima OP-312 Load Cell
Force Measuring Equipment – Compression <sup>1</sup>	(1 000 to 5 000) lbf	2.8 lbf	
Force Measuring Equipment – Compression <sup>1</sup>	(5 000 to 20 000) lbf	8.1 lbf	
Force Measuring Equipment – Tension <sup>1</sup>	Up to 10 lbf (10 to 25) lbf (25 to 50) lbf (50 to 150) lbf (150 to 250) lbf (250 to 500) lbf	0.001 5 lbf 0.002 1 lbf 0.008 3 lbf 0.023 lbf 0.065 lbf 0.13 lbf	NIST Class F Weights
Force Measuring Equipment – Tension <sup>1</sup>	(500 to 1 000) lbf	0.77 lbf	Optima OP-312 Load Cell
Force Measuring Equipment – Tension <sup>1</sup>	(1 000 to 5 000) lbf	2.8 lbf	
Force Measuring Equipment – Tension <sup>1</sup>	(5 000 to 10 000) lbf	4.2 lbf	
Force Measuring Equipment – Tension <sup>1</sup>	(10 000 to 20 000) lbf	8.1 lbf	Transcell BSS-20K Load Cell
Force Measuring Equipment – Tension <sup>1</sup>	(20 000 to 50 000) lbf	23 lbf	Optima OP-351 Load Cell
Force Measuring Equipment – Tension <sup>1</sup>	(50 000 to 100 000) lbf	44 lbf	Rinstrum TLWS-100K Load Cell

### Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Torque Devices	2 ozf·in to 2 000 lbf·ft	0.31 % of reading	AKO TSD2050 Torque Master
Torque Transducers	20 ozf·in to 100 lbf·in 100 lbf·in to 125 lbf·ft (125 to 2 000) lbf·ft	0.05 % of reading 0.06 % of reading 0.08 % of reading	Torque Arms, Master Weights

### Photometry and Radiometry

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Illuminance – Lux Meters	(180 to 1 792) lux (1 792 to 17 000) lux	1.6 % of reading 1.7 % of reading	FEL 1000W Lamp with Power Supply Unit

### Thermodynamic


Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity – Source/Measure	(10 to 30) °C (10 to 90) %RH	1.3 %RH	Comparison to Vaisala HMI-41/HMP-46 Temp/Humidity Indicator/Probe
Temperature – Source/Measure <sup>1</sup>	(-196 to 0) °C (0 to 100) °C	0.042 °C 0.042 °C	Hart Scientific 1521 Handheld Thermometer, Hart Scientific 5618B, Hart Scientific 5627A RTD Probes
Temperature – Source/Measure <sup>1</sup>	(100 to 420) °C (420 to 960) °C	0.06 °C 0.064 °C	Hart Scientific 1523 Handheld Thermometer, Hart Scientific 5624A RTD Probe
Infrared Measuring Devices <sup>1</sup>	(35 to 200) °C (200 to 350) °C (350 to 500) °C	0.95 °C 1.6 °C 2.1 °C	Black Body Source (Flat Plate) $\epsilon = (0.1 \text{ to } 1)$ , $\lambda = (8 \text{ to } 14) \mu\text{m}$

**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Source/Measure <sup>1</sup>	1 MHz to 27 GHz	3.8 pHz/Hz	Fluke 910R GPS Rubidium Frequency Standard
Stopwatches, Timers	Up to 19.99 s/d	58 ms/d	Helmut Klein TM-4500 Timometer
Optical Rotational Speed – Source/Measure <sup>2</sup>	(1 to 100 000) rpm	0.012 rpm	Agilent 33250A Function Generator
AC Duty Cycle – Source <sup>1</sup> Square Wave: < 3.3 Vp-p Freq: 10 MHz to 100 kHz	(1 to 10) % Duty Cycle 10 μs to 100 s (10 to 49) % Duty Cycle 10 μs to 100 s 50 % Duty Cycle 10 μs to 100 s (51 to 90) % Duty Cycle 10 μs to 100 s (90 to 99) % Duty Cycle 10 μs to 100 s	0.039 % of reading + 78 ns 0.62 % of reading + 78 ns 0.001 6 % of reading + 78 ns 0.62 % of reading + 78 ns 0.039 % of reading + 78 ns	Fluke 55xxA Multiproduct Calibrator

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

- Notes:
- On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
  - $t$  = time in seconds;  $L$  = length in inches;  $DL$  = diagonal length in inches; rpm = revolutions per minute; PF = power factor.
  - CMC does not include the Mismatch value. It will be added in the Measurement Uncertainty reported on the Certificate of Calibration.
  - The span is user set on the unit. The minimum range for this unit is 10 psi.
  - The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The CMC presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
  - The Measurement Uncertainty presented here is a unitless measurement.
  - For this parameter, 1 Digit will be added to the Measurement Uncertainty (MU) at the time of calibration.
  - The uncertainties shown are for the most favorable conditions. There is an increase in uncertainty that corresponds to the laboratory's AC voltage and current uncertainties at different frequencies other than the ones shown. Power factors (PF) other than the one shown contribute to the power uncertainty. PF is related to the cosine of phase. Therefore, uncertainties track the laboratory's phase uncertainty closely at PF near one but are magnified heavily as PF approaches zero. The lab may also report reactive power, apparent power, and power factor under this accreditation. If needed, contact laboratory for more information regarding uncertainties at frequency and power factor combinations other than the ones shown
  - This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2489.21.



Jason Stine, Vice President