



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**Transcat – Ireland**  
**Unit 15, Ballytrasna Business Park,**  
**Little Island, Cork, Ireland T45 DA34**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 07 September 2025

Certificate Number: AC-2489.29



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

**Transcat – Ireland**  
Unit 15, Ballytrasna Business Park,  
Little Island, Cork, Ireland T45 DA34  
Kevin Nuss +353 021 4506572

### CALIBRATION

Valid to: **September 7, 2025**

Certificate Number: **AC-2489.29**

#### Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source <sup>1</sup>	Up to 120 $\mu$ A		Fluke 5560A Multiproduct Calibrator
	(3 to 45) Hz	0.02 % of reading + 7.8 nA	
	45 Hz to 1 kHz	0.02 % of reading + 7.8 nA	
	(1 to 5) kHz	0.02 % of reading + 7.8 nA	
	(5 to 10) kHz	0.12 % of reading + 32 nA	
	(10 to 30) kHz	0.39 % of reading + 0.78 $\mu$ A	
	(0.12 to 1.2) mA		
	(3 to 45) Hz	0.02 % of reading + 78 nA	
	45 Hz to 1 kHz	0.02 % of reading + 78 nA	
	(1 to 5) kHz	0.02 % of reading + 78 nA	
	(5 to 10) kHz	0.12 % of reading + 78 nA	
	(10 to 30) kHz	0.39 % of reading + 3.9 $\mu$ A	
	(1.2 to 12) mA		
	(3 to 45) Hz	0.02 % of reading + 0.78 $\mu$ A	
	45 Hz to 1 kHz	0.02 % of reading + 0.78 $\mu$ A	
	(1 to 5) kHz	0.02 % of reading + 0.78 $\mu$ A	
	(5 to 10) kHz	0.12 % of reading + 0.78 $\mu$ A	
	(10 to 30) kHz	0.39 % of reading + 7.8 $\mu$ A	
(12 to 120) mA			
(3 to 45) Hz	0.02 % of reading + 7.8 $\mu$ A		
45 Hz to 1 kHz	0.012 % of reading + 3.9 $\mu$ A		
(1 to 5) kHz	0.02 % of reading + 6.3 $\mu$ A		
(5 to 10) kHz	0.12 % of reading + 7.8 $\mu$ A		
(10 to 30) kHz	0.39 % of reading + 78 $\mu$ A		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source <sup>1</sup>	(0.12 to 1.2) A		Fluke 5560A Multiproduct Calibrator
	(3 to 45) Hz	0.02 % of reading + 78 µA	
	45 Hz to 1 kHz	0.02 % of reading + 39 µA	
	(1 to 5) kHz	0.02 % of reading + 63 µA	
	(5 to 10) kHz	0.12 % of reading + 0.24 mA	
	(10 to 30) kHz	0.39 % of reading + 0.24 mA	
	(1.2 to 3.1) A		
	(3 to 45) Hz	0.03 % of reading + 0.39 mA	
	45 Hz to 1 kHz	0.024 % of reading + 0.24 mA	
	(1 to 5) kHz	0.03 % of reading + 0.24 mA	
	(5 to 10) kHz	0.2 % of reading + 0.39 mA	
	(3.1 to 12) A		
	(3 to 45) Hz	0.03 % of reading + 0.78 mA	
	45 Hz to 1 kHz	0.024 % of reading + 0.39 mA	
	(1 to 5) kHz	0.03 % of reading + 0.63 mA	
(5 to 10) kHz	0.2 % of reading + 0.78 mA		
(12 to 30,2) A			
(3 to 45) Hz	0.078 % of reading + 7.8 mA		
45 Hz to 1 kHz	0.055 % of reading + 6.3 mA		
(1 to 5) kHz	0.39 % of reading + 6.3 mA		
AC Current – Measure <sup>1</sup>	(0.2 to 20) µA		Fluke 8588A 8.5 Digit Multimeter
	1 Hz to 2 kHz	0.2 % of reading + 2.5 nA	
	(2 to 10) kHz	0.2 % of reading + 2.5 nA	
	(10 to 30) kHz	0.2 % of reading + 2.5 nA	
	(20 to 200) µA		
	1 Hz to 2 kHz	0.028 % of reading + 5 nA	
	(2 to 10) kHz	0.053 % of reading + 5 nA	
	(10 to 30) kHz	0.074 % of reading + 5 nA	
	(30 to 100) kHz	0.041 % of reading + 10 nA	
	(0.2 to 2) mA		
	1 Hz to 2 kHz	0.028 % of reading + 50 nA	
	(2 to 10) kHz	0.053 % of reading + 50 nA	
	(10 to 30) kHz	0.074 % of reading + 50 nA	
	(30 to 100) kHz	0.41 % of reading + 0.1 µA	
	(2 to 20) mA		
1 Hz to 2 kHz	0.028 % of reading + 0.5 µA		
(2 to 10) kHz	0.053 % of reading + 0.5 µA		
(10 to 30) kHz	0.074 % of reading + 0.5 µA		
(30 to 100) kHz	0.41 % of reading + 1 µA		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(20 to 200) mA 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (0.2 to 2) A 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (2 to 20) A 10 Hz to 2 kHz (2 to 10) kHz (20 to 30) A 10 Hz to 2 kHz (2 to 10) kHz	0.028 % of reading + 5 $\mu$ A 0.052 % of reading + 5 $\mu$ A 0.074 % of reading + 5 $\mu$ A 0.03 % of reading + 0.1 mA 0.056 % of reading + 0.1 mA 0.08 % of reading + 0.1 mA 0.084 % of reading + 0.5 mA 0.086 % of reading + 0.5 mA 0.084 % of reading + 12 mA 0.12 % of reading + 12 mA	Fluke 8588A 8.5 Digit Multimeter
AC Clamp-on Ammeter (Toroidal Type) Transformer Type Sensor <sup>1</sup>	(20 to 60) A (45 to 440) Hz (60 to 155) A (45 to 440) Hz (155 to 600) A (45 to 440) Hz (600 to 1 500) A (45 to 440) Hz	0.5 % of reading + 2 mA 0.5 % of reading + 12 mA 0.5 % of reading + 20 mA 0.51 % of reading + 0.31 A	Fluke 5560A Multiproduct Calibrator Fluke 55XXA/COIL 50 50-turn Coil
AC Clamp-on Ammeter (Non-Toroidal Type) Hall Effect Sensor <sup>1</sup>	(20 to 60) A (45 to 440) Hz (60 to 155) A (45 to 440) Hz (155 to 600) A (45 to 440) Hz (600 to 1 500) A (45 to 440) Hz	0.53 % of reading + 2 mA 0.53 % of reading + 12 mA 0.53 % of reading + 20 mA 0.53 % of reading + 0.31 A	Fluke 5560A Multiproduct Calibrator Fluke 55XXA/COIL 50 50-turn Coil
AC Voltage – Source <sup>1</sup>	Up to 12 mV (3 to 5) Hz (5 to 10) Hz 10 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz	0.2 % of reading + 5.5 $\mu$ V 0.068 % of reading + 5.5 $\mu$ V 0.012 % of reading + 4.7 $\mu$ V 0.03 % of reading + 4.7 $\mu$ V 0.12 % of reading + 12 $\mu$ V 0.63 % of reading + 24 $\mu$ V 0.63 % of reading + 24 $\mu$ V	Fluke 5560A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup>	(12 to 120) mV		Fluke 5560A Multiproduct Calibrator
	(3 to 5) Hz	0.2 % of reading + 5.5 μV	
	(5 to 10) Hz	0.068 % of reading + 5.5 μV	
	10 Hz to 20 kHz	0.011 % of reading + 4.7 μV	
	(20 to 50) kHz	0.028 % of reading + 6.3 μV	
	(50 to 100) kHz	0.063 % of reading + 16 μV	
	(100 to 300) kHz	0.16 % of reading + 24 μV	
	(300 to 500) kHz	0.16 % of reading + 24 μV	
	(0.12 to 1.2) V		
	(3 to 5) Hz	0.2 % of reading + 59 μV	
	(5 to 10) Hz	0.068 % of reading + 55 μV	
	(10 to 40) Hz	0.011 % of reading + 47 μV	
	40 Hz to 20 kHz	0.011 % of reading + 6.3 μV	
	(20 to 50) kHz	0.024 % of reading + 11 μV	
	(50 to 100) kHz	0.055 % of reading + 32 μV	
	(100 to 300) kHz	0.15 % of reading + 63 μV	
	(300 to 500) kHz	0.15 % of reading + 63 μV	
	(1.2 to 12) V		
	(3 to 5) Hz	0.2 % of reading + 0.59 mV	
	(5 to 10) Hz	0.068 % of reading + 0.59 mV	
	(10 to 40) Hz	0.011 % of reading + 0.28 mV	
	40 Hz to 20 kHz	0.011 % of reading + 39 μV	
	(20 to 50) kHz	0.024 % of reading + 39 μV	
	(50 to 100) kHz	0.055 % of reading + 98 μV	
	(100 to 300) kHz	0.16 % of reading + 0.47 mV	
	(300 to 500) kHz	0.16 % of reading + 0.47 mV	
	(12 to 70) V		
	(3 to 5) Hz	0.2 % of reading + 5.9 mV	
(5 to 10) Hz	0.068 % of reading + 5.9 mV		
(10 to 40) Hz	0.011 % of reading + 2.8 mV		
40 Hz to 20 kHz	0.011 % of reading + 0.39 mV		
(20 to 50) kHz	0.024 % of reading + 0.39 mV		
(50 to 100) kHz	0.055 % of reading + 0.97 mV		
(100 to 300) kHz	0.14 % of reading + 16 mV		
(70 to 120) V			
(3 to 5) Hz	0.2 % of reading + 5.9 mV		
(5 to 10) Hz	0.068 % of reading + 5.9 mV		
(10 to 40) Hz	0.011 % of reading + 2.8 mV		
40 Hz to 20 kHz	0.011 % of reading + 0.39 mV		
(20 to 50) kHz	0.024 % of reading + 0.39 mV		
(50 to 100) kHz	0.055 % of reading + 0.97 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup>	(120 to 330) V (3 to 5) Hz (5 to 10) Hz 10 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (330 to 1 020) V (3 to 5) Hz (5 to 10) Hz 10 Hz to 10 kHz	0.2 % of reading + 59 mV 0.068 % of reading + 59 mV 0.011 % of reading + 6.3 mV 0.024 % of reading + 6.3 mV 0.12 % of reading + 9.7 mV 0.2 % of reading + 59 mV 0.068 % of reading + 59 mV 0.011 % of reading + 63 mV	Fluke 5560A Multiproduct Calibrator
AC Voltage – Measure <sup>1</sup>	(0.1 to 10) mV 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (10 to 100) mV 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (0.1 to 1) V 1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	0.029 % of reading + 1.1 μV 0.037 % of reading + 1.1 μV 0.038 % of reading + 1.1 μV 0.3 % of reading + 0.78 μV 1 % of reading + 3.9 μV 2 % of reading + 3.9 μV 0.008 9 % of reading + 0.5 μV 0.013 % of reading + 0.5 μV 0.023 % of reading + 1 μV 0.053 % of reading + 5 μV 0.21 % of reading + 31 μV 1 % of reading + 0.1 mV 1.5 % of reading + 0.5 mV 4.1 % of reading + 1 mV 8.4 % of reading + 1 mV 16 % of reading + 1 mV 0.007 7% of reading + 5 μV 0.012 % of reading + 5 μV 0.023 % of reading + 10 μV 0.053 % of reading + 50 μV 0.21 % of reading + 0.31 mV 1 % of reading + 1 mV 1.5 % of reading + 5 mV 4 % of reading + 10 mV 8.2 % of reading + 10 mV 15 % of reading + 10 mV	Fluke 8588A 8.5 Digit Multimeter



**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(1 to 10) V		Fluke 8588A 8.5 Digit Multimeter
	1 Hz to 2 kHz	0.007 6 % of reading + 50 μV	
	(2 to 10) kHz	0.012 % of reading + 50 μV	
	(10 to 30) kHz	0.023 % of reading + 0.1 mV	
	(30 to 100) kHz	0.053 % of reading + 0.5 mV	
	(100 to 300) kHz	0.21 % of reading + 3.1 mV	
	300 kHz to 1 MHz	1 % of reading + 10 mV	
	(1 to 2) MHz	1.5 % of reading + 50 mV	
	(2 to 4) MHz	4 % of reading + 0.1 V	
	(4 to 8) MHz	8.2 % of reading + 0.1 V	
	(8 to 10) MHz	15 % of reading + 0.1 V	
	(10 to 100) V		
	1 Hz to 2 kHz	0.009 % of reading + 0.5 mV	
	(2 to 10) kHz	0.011 % of reading + 0.5 mV	
	(10 to 30) kHz	0.023 % of reading + 1 mV	
(30 to 100) kHz	0.059 % of reading + 5 mV		
(100 to 300) kHz	0.37 % of reading + 47 mV		
300 kHz to 1 MHz	1 % of reading + 0.5 V		
(100 to 1 050) V			
1 Hz to 2 kHz	0.011 % of reading + 25 mV		
(2 to 10) kHz	0.011 % of reading + 25 mV		
(10 to 30) kHz	0.023 % of reading + 25 mV		
(30 to 100) kHz	0.059 % of reading + 0.1 V		
DC Current – Source <sup>1</sup>	Up to 120 μA	0.009 8 % of reading + 4.7 nA	Fluke 5560A Multiproduct Calibrator
	(0.12 to 1.2) mA	0.007 8 % of reading + 12 nA	
	(1.2 to 12) mA	0.007 8 % of reading + 63 nA	
	(12 to 120) mA	0.007 8 % of reading + 0.63 μA	
	(0.12 to 1.2) A	0.013 % of reading + 7.8 μA	
	(1.2 to 3.1) A	0.024 % of reading + 0.12 mA	
	(3.1 to 12) A	0.024 % of reading + 0.2 mA	
	(12 to 30.2) A	0.078 % of reading + 0.4 mA	
DC Current – Measure <sup>1</sup>	Up to 20 μA	0.002 9 % of reading + 0.4 nA	Fluke 8588A 8.5 Digit Multimeter
	(20 to 200) μA	0.001 % of reading + 0.39 nA	
	(0.2 to 2) mA	0.001 % of reading + 3.9 nA	
	(2 to 20) mA	0.001 5 % of reading + 39 nA	
	(20 to 200) mA	0.005 8 % of reading + 1 μA	
	(0.2 to 2) A	0.013 % of reading + 0.1 mA	
	(2 to 20) A	0.023 % of reading + 0.4 mA	
	(20 to 30) A	0.055 % of reading + 4.4 mA	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Clamp-on Ammeter (Non-Toroidal Type) Hall Effect Sensor <sup>1</sup>	(20 to 60) A (60 to 155) A (155 to 600) A (600 to 1 500) A	0.59 % of reading + 3.9 mA 0.59 % of reading + 6 mA 0.59 % of reading + 9.5 mA 0.6 % of reading + 20 mA	Fluke 5560A Multiproduct Calibrator Fluke 55XXA/COIL 50 50-turn Coil
DC Voltage – Source <sup>1</sup>	Up to 120 mV (0.12 to 1.2) V (1.2 to 12) V (12 to 120) V (120 to 1 020) V	0.000 93 % of reading + 0.62 μV 0.000 64 % of reading + 0.78 μV 0.000 62 % of reading + 7.8 μV 0.000 85 % of reading + 78 μV 0.000 86 % of reading + 0.78 mV	Fluke 5560A Multiproduct Calibrator
DC Voltage – Measure <sup>1</sup>	Up to 200 mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1050) V	0.000 77 % of reading + 0.2 μV 0.000 29 % of reading + 0.3 μV 0.000 29 % of reading + 0.47 μV 0.000 43 % of reading + 30 μV 0.000 44 % of reading + 0.5 mV	Fluke 8588A 6.5 Digit Multimeter
Resistance – Source <sup>1</sup> (Simulation)	Up to 12 Ω (12 to 120) Ω (0.12 to 1.2) kΩ (1.2 to 12) kΩ (12 to 120) kΩ (0.12 to 1.2) MΩ (1.2 to 12) MΩ (12 to 120) MΩ (0.12 to 1.2) GΩ	0.002 % of reading + 0.78 mΩ 0.002 % of reading + 0.78 mΩ 0.002 % of reading + 1.6 mΩ 0.002 % of reading + 16 mΩ 0.002 % of reading + 0.16 Ω 0.002 % of reading + 1.6 Ω 0.002 8 % of reading + 24 Ω 0.034 % of reading + 2 kΩ 0.32 % of reading + 78 kΩ	Fluke 5560A Multiproduct Calibrator (4-wire Mode)
Resistance – Measure <sup>1</sup>	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω (0.2 to 2) kΩ (2 to 20) kΩ (20 to 200) kΩ (0.2 to 2) MΩ (2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ	0.001 6 % of reading + 4 μΩ 0.001 % of reading + 14 μΩ 0.000 92 % of reading + 47 μΩ 0.000 91 % of reading + 0.47 mΩ 0.000 93 % of reading + 4.7 mΩ 0.000 93 % of reading + 47 mΩ 0.001 1 % of reading + 1 Ω 0.001 9 % of reading + 0.1 kΩ 0.012 % of reading + 10 kΩ 0.13 % of reading + 1 MΩ	Fluke 8588A 8.5 Digit Multimeter





ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Low Current Resistance – Measure <sup>1</sup>	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω (0.2 to 2) kΩ (2 to 20) kΩ (20 to 200) kΩ (0.2 to 2) MΩ (2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ	0.001 7 % of reading + 4 μΩ 0.001 % of reading + 14 μΩ 0.001 7 % of reading + 0.2 mΩ 0.001 8 % of reading + 2 mΩ 0.002 2 % of reading + 20 mΩ 0.002 2 % of reading + 62 mΩ 0.002 6 % of reading + 1 Ω 0.038 % if reading + 0.3 kΩ 0.013 % of reading + 10 kΩ 0.13 % of reading + 1 MΩ	Fluke 8588A 8.5 Digit Multimeter
High Voltage Resistance – Measure <sup>1</sup>	(2 to 20) MΩ (20 to 200) MΩ (0.2 to 2) GΩ (2 to 20) GΩ	0.001 7 % of reading + 10 Ω 0.006 8 % of reading + 0.1 kΩ 0.023 % of reading + 0.1 MΩ 0.13 % of reading + 10 MΩ	Fluke 8588A 8.5 Digit Multimeter
Capacitance – Source <sup>1</sup> (Simulation)	Up to 1.2 nF (20 to 40) Hz (40 to 100) Hz 100 Hz to 10 kHz (10 to 12) kHz (12 to 14) kHz (1.2 to 3) nF (10 to 150) Hz 150 Hz to 5 kHz (5 to 6) kHz (6 to 8) kHz (3 to 12) nF 10 Hz to 5 kHz (5 to 6) kHz (6 to 8) kHz (12 to 30) nF (20 to 200) Hz 200 Hz to 1.3 kHz (1.3 to 2.7) kHz (2.7 to 3.7) kHz (30 to 120) nF 10 Hz to 1.3 kHz (1.3 to 2.7) kHz (2.7 to 3.7) kHz (0.12 to 1.2) μF (2 to 310) Hz (310 to 800) Hz 800 Hz to 1.1 kHz	0.87 % of reading + 1.6 pF 0.49 % of reading + 1.6 pF 0.094 % of reading + 1.6 pF 0.48 % of reading + 1.6 pF 0.87 % of reading + 1.6 pF 0.49 % of reading + 3.9 pF 0.09 % of reading + 3.9 pF 0.48 % of reading + 3.9 pF 0.87 % of reading + 3.9 pF 0.09 % of reading + 3.9 pF 0.48 % of reading + 3.9 pF 0.87 % of reading + 3.9 pF 0.49 % of reading + 24 pF 0.1 % of reading + 24 pF 0.49 % of reading + 24 pF 0.88 % of reading + 24 pF 0.1 % of reading + 24 pF 0.49 % of reading + 24 pF 0.88 % of reading + 24 pF 0.1 % of reading + 0.24 nF 0.49 % of reading + 0.24 nF 0.88 % of reading + 0.24 nF	Fluke 5560A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source <sup>1</sup> (Simulation)	(1.2 to 12) $\mu$ F		Fluke 5560A Multiproduct Calibrator
	500 mHz to 110 Hz	0.1 % of reading + 2.4 nF	
	(110 to 250) Hz	0.49 % of reading + 2.4 nF	
	(250 to 350) Hz	0.88 % of reading + 2.4 nF	
	(12 to 120) $\mu$ F		
	(100 to 500) mHz	0.5 % of reading + 20 nF	
	500 mHz to 40 Hz	0.12 % of reading + 20 nF	
	(40 to 80) Hz	0.5 % of reading + 20 nF	
	(80 to 110) Hz	0.89 % of reading + 20 nF	
	(0.12 to 1.2) mF		
	100 mHz to 11 Hz	0.2 % of reading + 0.2 $\mu$ F	
	(11 to 18) Hz	0.58 % of reading + 0.2 $\mu$ F	
(18 to 25) Hz	1 % of reading + 0.2 $\mu$ F		
Capacitance – Measure <sup>1</sup>	(1.2 to 12) mF		Fluke 8588A 8.5 Digit Multimeter
	30 mHz to 4 Hz	0.19 % of reading + 2.4 $\mu$ F	
	(4 to 6) Hz	0.58 % of reading + 2.4 $\mu$ F	
	(6 to 8) Hz	1 % of reading + 2.4 $\mu$ F	
	(12 to 120) mF		
	10 mHz to 1.3 Hz	0.39 % of reading + 24 $\mu$ F	
	(1.3 to 1.7) Hz	0.78 % of reading + 24 $\mu$ F	
	(1.7 to 2.5) Hz	1.2 % of reading + 24 $\mu$ F	
	Up to 2 nF	0.19 % of reading + 1 pF	
	(2 to 20) nF	0.081 % of reading + 2 pF	
	(20 to 200) nF	0.049 % of reading + 10 pF	
	(0.2 to 2) $\mu$ F	0.041 % of reading + 0.1 nF	
(2 to 20) $\mu$ F	0.042 % of reading + 1 nF		
(20 to 200) $\mu$ F	0.061 % of reading + 10 nF		
(0.2 to 2) mF	0.061 % of reading + 0.1 $\mu$ F		
(2 to 20) mF	0.071 % of reading + 1 $\mu$ F		
(20 to 200) mF	0.072 % of reading + 10 $\mu$ F		
Inductance – Source <sup>1</sup> (Simulation)	Up to 120 $\mu$ H		Fluke 5560A Multiproduct Calibrator
	(490 to 550) Hz	0.93 % of reading + 0.16 $\mu$ H	
	550 Hz to 1 kHz	0.35 % of reading + 0.16 $\mu$ H	
	1 kHz	0.16 % of reading + 0.16 $\mu$ H	
	(1 to 13) kHz	0.35 % of reading + 0.16 $\mu$ H	
	(13 to 17) kHz	0.93 % of reading + 0.16 $\mu$ H	
	(0.12 to 1.2) mH		
	(260 to 330) Hz	0.87 % of reading + 0.78 $\mu$ H	
	330 Hz to 1 kHz	0.29 % of reading + 0.78 $\mu$ H	
	1 kHz	0.094 % of reading + 0.78 $\mu$ H	
	(1 to 1.6) kHz	0.29 % of reading + 0.78 $\mu$ H	
	(1.6 to 2.5) kHz	0.87 % of reading + 0.78 $\mu$ H	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Source <sup>1</sup> (Simulation)	(1.2 to 3.3) mH		Fluke 5560A Multiproduct Calibrator
	500 mHz to 110 Hz	0.3 % of reading + 7.8 μH	
	110 Hz	0.094 % of reading + 7.8 μH	
	(110 to 800) Hz	0.29 % of reading + 7.8 μH	
	(800 to 980) Hz	0.87 % of reading + 7.8 μH	
	(3.3 to 12) mH		
	500 mHz to 110 Hz	0.29 % of reading + 7.8 μH	
	110 Hz	0.093 % of reading + 7.8 μH	
	110 to 1 kHz	0.29 % of reading + 7.8 μH	
	(1 to 1.4) kHz	0.87 % of reading + 7.8 μH	
	(12 to 83) mH		
	100 mHz to 100 Hz	0.29 % of reading + 78 μH	
	100 Hz	0.093 % of reading + 78 μH	
	(100 to 180) Hz	0.29 % of reading + 78 μH	
	(180 to 230) Hz	0.87 % of reading + 78 μH	
	(83 to 120) mH		
	100 mHz to 100 Hz	0.29 % of reading + 78 μH	
	100 Hz	0.093 % of reading + 78 μH	
	(100 to 320) Hz	0.3 % of reading + 78 μH	
	320 Hz to 1 kHz	0.87 % of reading + 78 μH	
	(120 to 650) mH		
	50 mHz to 10 Hz	0.32 % of reading + 0.78 mH	
	10 Hz	0.14 % of reading + 0.78 mH	
	(10 to 30) Hz	0.32 % of reading + 0.78 mH	
	(30 to 55) Hz	0.9 % of reading + 0.78 mH	
	(0.65 to 1.2) H		
	50 mHz to 10 Hz	0.32 % of reading + 0.78 mH	
	10 Hz	0.14 % of reading + 0.78 mH	
	(10 to 100) Hz	0.32 % of reading + 0.78 mH	
	(100 to 170) Hz	0.9 % of reading + 0.78 mH	
(1.2 to 5.5) H			
10 mHz to 3 Hz	0.36 % of reading + 7.8 mH		
3 Hz	0.17 % of reading + 7.8 mH		
(3 to 8) Hz	0.36 % of reading + 7.8 mH		
(8 to 16) Hz	0.93 % of reading + 7.8 mH		
(5.5 to 12) H			
10 mHz to 3 Hz	0.36 % of reading + 7.8 mH		
3 Hz	0.17 % of reading + 7.8 mH		
(3 to 19) Hz	0.36 % of reading + 7.8 mH		
(19 to 37) Hz	0.93 % of reading + 7.8 mH		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Source <sup>1</sup> (Simulation)	(12 to 30) H 5 mHz to 2 Hz 2 Hz (2 to 4) Hz (4 to 9) Hz	0.39 % of reading + 78 mH 0.21 % of reading + 78 mH 0.39 % of reading + 78 mH 1 % of reading + 78 mH	Fluke 5560A Multiproduct Calibrator
	(30 to 120) H 5 mHz to 2 Hz 2 Hz (2 to 7) Hz (7 to 14) Hz	0.39 % of reading + 78 mH 0.21 % of reading + 78 mH 0.39 % of reading + 78 mH 1 % of reading + 78 mH	
Phase – Source <sup>1</sup>	(0 to 360) °		Fluke 5560A Multiproduct Calibrator
	(3 to 65) Hz	0.08 °	
	(65 to 500) Hz	0.19 °	
	500 Hz to 1 kHz	0.39 °	
	(1 to 5) kHz	1.9 °	
Electrical Simulation of RTD Indicating Devices – Source <sup>1</sup>	Pt 385, 100 Ω		Fluke 5560A Multiproduct Calibrator
	(-200 to -80) °C	0.039 °C	
	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.054 °C	
	(100 to 300) °C	0.07 °C	
	(300 to 400) °C	0.078 °C	
	(400 to 630) °C	0.093 °C	
	(630 to 800) °C	0.18 °C	
	Pt 385, 200 Ω		
	(-200 to -80) °C	0.031 °C	
	(-80 to 0) °C	0.031 °C	
	(0 to 100) °C	0.031 °C	
	(100 to 260) °C	0.039 °C	
	(260 to 300) °C	0.093 °C	
(300 to 400) °C	0.1 °C		
(400 to 600) °C	0.11 °C		
(600 to 630) °C	0.12 °C		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices – Source <sup>1</sup>	Pt 385, 500 Ω		Fluke 5560A Multiproduct Calibrator
	(-200 to -80) °C	0.031 °C	
	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.039 °C	
	(100 to 260) °C	0.047 °C	
	(260 to 300) °C	0.062 °C	
	(300 to 400) °C	0.062 °C	
	(400 to 600) °C	0.07 °C	
	(600 to 630) °C	0.085 °C	
	Pt 385, 1 000 Ω		
	(-200 to -80) °C	0.023 °C	
	(-80 to 0) °C	0.023 °C	
	(0 to 100) °C	0.031 °C	
	(100 to 260) °C	0.039 °C	
	(260 to 300) °C	0.047 °C	
	(300 to 400) °C	0.054 °C	
	(400 to 600) °C	0.054 °C	
	(600 to 630) °C	0.18 °C	
	Pt 3916, 100 Ω		
	(-200 to -190) °C	0.19 °C	
	(-190 to -80) °C	0.031 °C	
	(-80 to 0) °C	0.039 °C	
	(0 to 100) °C	0.047 °C	
	(100 to 260) °C	0.054 °C	
	(260 to 300) °C	0.062 °C	
	(300 to 400) °C	0.07 °C	
	(400 to 600) °C	0.078 °C	
	(600 to 630) °C	0.18 °C	
Pt 3926, 100 Ω			
(-200 to -80) °C	0.039 °C		
(-80 to 0) °C	0.039 °C		
(0 to 100) °C	0.054 °C		
(100 to 300) °C	0.07 °C		
(300 to 400) °C	0.078 °C		
(400 to 630) °C	0.093 °C		
Ni 672, 120 Ω			
(-80 to 0) °C	0.06 °C		
(0 to 100) °C	0.06 °C		
(100 to 260) °C	0.11 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices – Source <sup>1</sup>	Cu 427, 10 Ω (-80 to 260) °C	0.23 °C	Fluke 5560A Multiproduct Calibrator
	Cu 428, 50 Ω (-180 to 200) °C	0.31 °C	
	Cu 428, 100 Ω (-180 to 40) °C	0.31 °C	
	(40 to 200) °C	0.5 °C	
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure <sup>1</sup>	Type B (600 to 800) °C	0.34 °C	Fluke 5560A Multiproduct Calibrator
	(800 to 1 000) °C	0.26 °C	
	(1 000 to 1 550) °C	0.23 °C	
	(1 550 to 1 820) °C	0.26 °C	
	Type C (0 to 150) °C	0.19 °C	
	(150 to 650) °C	0.16 °C	
	(650 to 1 000) °C	0.2 °C	
	(1 000 to 1 800) °C	0.35 °C	
	(1 800 to 2 315) °C	0.61 °C	
	Type D (0 to 150) °C	0.19 °C	
	(150 to 650) °C	0.16 °C	
	(650 to 1 000) °C	0.2 °C	
	(1 000 to 1 800) °C	0.34 °C	
	(1 800 to 2 315) °C	0.61 °C	
	Type E (-250 to -150) °C	0.31 °C	
	(-150 to -25) °C	0.11 °C	
	(-25 to 350) °C	0.09 °C	
	(350 to 650) °C	0.12 °C	
	(650 to 1 000) °C	0.16 °C	
	Type G (0 to 150) °C	0.39 °C	
	(150 to 650) °C	0.26 °C	
	(650 to 1 000) °C	0.2 °C	
	(1 000 to 1 800) °C	0.33 °C	
	(1 800 to 2 315) °C	0.6 °C	
Type J (-210 to -100) °C	0.19 °C		
(-100 to -30) °C	0.1 °C		
(-30 to 150) °C	0.09 °C		
(150 to 760) °C	0.11 °C		
(760 to 1 200) °C	0.16 °C		



**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure <sup>1</sup>	Type K		Fluke 5560A Multiproduct Calibrator
	(-200 to -100) °C	0.22 °C	
	(-100 to -25) °C	0.1 °C	
	(-25 to 120) °C	0.09 °C	
	(120 to 1 000) °C	0.16 °C	
	(1 000 to 1 372) °C	0.27 °C	
	Type L		
	(-200 to -100) °C	0.24 °C	
	(-100 to 800) °C	0.16 °C	
	(800 to 900) °C	0.09 °C	
	Type N		
	(-200 to -100) °C	0.26 °C	
	(-100 to -25) °C	0.12 °C	
	(-25 to 120) °C	0.09 °C	
	(120 to 410) °C	0.09 °C	
	(410 to 1 300) °C	0.16 °C	
	Type R		
	(0 to 250) °C	0.4 °C	
	(250 to 400) °C	0.23 °C	
	(400 to 1 000) °C	0.21 °C	
	(1 000 to 1 767) °C	0.26 °C	
Type S			
(0 to 250) °C	0.33 °C		
(250 to 400) °C	0.24 °C		
(400 to 1 000) °C	0.25 °C		
(1 000 to 1 767) °C	0.32 °C		
Type T			
(-250 to -150) °C	0.47 °C		
(-150 to 0) °C	0.16 °C		
(0 to 120) °C	0.1 °C		
(120 to 400) °C	0.09 °C		
Type U			
(-200 to 0) °C	0.31 °C		
(0 to 600) °C	0.09 °C		
Type BP			
(0 to 1 000) °C	0.31 °C		
(1 000 to 2 000) °C	0.47 °C		
(2 000 to 2 500) °C	0.62 °C		
Type XK			
(-200 to 300) °C	0.16 °C		
(300 to 800) °C	0.23 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes <sup>1,4</sup>			
Amplitude – DC			
into 50 Ω load	(-6.6 to 6.6) V	0.22 % of reading + 31 μV	
into 1 MΩ load	(-130 to 130) V	0.12 % of reading + 31 μV	
Amplitude – Square Wave			
into 50 Ω load	1 mVp-p to 6.6 Vp-p 10 Hz to 10 kHz	0.22 % of reading + 31 μV	
into 1 MΩ load	1 mVp-p to 130 Vp-p 10 Hz to 1 kHz (1 to 10) kHz	0.14 % of reading + 31 μV 0.22 % of reading + 31 μV	
Time Markers			
into 50 Ω load	1 nS to 20 mS 50 ms 100 ms 200 ms 500 ms 1 s 2 s 5 s	0.000 22 % of reading 0.005 9 % of reading 0.009 8 % of reading 0.018 % of reading 0.041 % of reading 0.080 % of reading 0.16 % of reading 0.39 % of reading	
Rise Time			
into 50 Ω load	5 mVp-p to 2.5 Vp-p		
Rate: 1 kHz to 2 MHz	(200 to 300) ps	50 ps	
Rate: (2 to 10) MHz	(250 to 350) ps	50 ps	
Level Sine Wave			
into 50 Ω load	5 mVp-p to 5.5 Vp-p 50 kHz	1.8 % of reading + 0.23 mV	
	50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz	2.8 % of reading + 0.23 mV 3.2 % of reading + 0.23 mV 4 % of reading + 0.23 mV	
	5 mVp-p to 3.5 Vp-p 600 MHz to 1.1 GHz	5.5 % of reading + 0.23 mV	

Fluke 5560A/1G  
Multiproduct Calibrator  
with 1.1 GHz Scope Option

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes <sup>1</sup> Bandwidth Flatness (50 kHz Reference) into 50 Ω load	5 mVp-p to 5.5 Vp-p 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mVp-p to 3.5 Vp-p 600 MHz to 1.1 GHz	1.4 % of reading + 78 μV 1.8 % of reading + 78 μV 3.2 % of reading + 78 μV 4 % of reading + 78 μV	Fluke 5560A/1G Multiproduct Calibrator with 1.1 GHz Scope Option
Input Impedance Measure	(40 to 60) Ω (0.5 to 1.5) MΩ	0.082 % of reading 0.081 % of reading	
Input Capacitance Measure	(5 to 50) pF	3.9 % of reading + 0.39 pF	
Wave Generator (Sine, Square, Triangle) Amplitude into 50 Ω load	1.8 mVp-p to 2.5 Vp-p 10 Hz to 10 kHz	2.3 % of reading + 78 μV	
into 1 MΩ load	1.8 mVp-p to 55 Vp-p 10 Hz to 10 kHz	2.3 % of reading + 78 μV	
Frequency	10 Hz to 10 kHz	0.001 9 % of reading + 12 mHz	
AC Power – Source <sup>1,5</sup> (10 to 65) Hz Power Factor = 1	(3.3 to 9) mA (0.11 to 3) mW 3 mW to 9 W (9 to 33) mA (0.3 to 10) mW 10 mW to 33 W (33 to 90) mA (1 to 30) mW 30 mW to 90 W (90 to 330) mA (3 to 100) mW 100 mW to 300 W (0.33 to 0.9) A (11 to 300) mW 300 mW to 900 W	0.13 % of reading 0.077 % of reading 0.089 % of reading 0.077 % of reading 0.071 % of reading 0.057 % of reading 0.089 % of reading 0.078 % of reading 0.071 % of reading 0.081 % of reading	Fluke 5522A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power – Source <sup>1,5</sup> (10 to 65) Hz Power Factor = 1	(0.9 to 2.2) A	0.089 % of reading	Fluke 5522A Multiproduct Calibrator
	30 mW to 0.72 W	0.079 % of reading	
	720 mW to 2 kW	0.088 % of reading	
	(2.2 to 4.5) A	0.18 % of reading	
DC Power – Source <sup>1,5</sup>	80 mW to 1.4 W	0.17 % of reading	Fluke 5522A Multiproduct Calibrator
	1.4 W to 4.5 kW	0.17 % of reading	
	(4.5 to 20.5) A		
	(0.15 to 6.7) W		
DC Power – Source <sup>1,5</sup>	6.7 W to 20 kW		Fluke 5522A Multiproduct Calibrator
	(0.33 to 330) mA	0.024 % of reading	
	11 µW to 1.1 mW	0.027 % of reading	
	(1.1 to 110) mW	0.024 % of reading	
	(0.11 to 110) W	0.018 % of reading	
	(110 to 330) W		
	(0.33 to 3) A	0.044 % of reading	
	11 µW to 110 mW	0.053 % of reading	
(0.11 to 990) W	0.009 6 % of reading		
(0.99 to 3) kW			
DC Power – Source <sup>1,5</sup>	(3 to 20.5) A	0.088 % of reading	Fluke 5522A Multiproduct Calibrator
	(99 to 990) mW	0.07 % of reading	
	0.99 W to 6.8 kW	0.04 % of reading	
	(6.8 to 20.5) kW		

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Determination <sup>2,3</sup> (OIML Class E2, F1, F2)	1 mg	2 µg	Modified Double Substitution; Electronic Mass Comparators
	2 mg	2 µg	
	5 mg	2 µg	
	10 mg	2.7 µg	
	20 mg	3 µg	
	50 mg	4 µg	
	100 mg	5.3 µg	
	200 mg	6.7 µg	
	500 mg	8.3 µg	

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Determination <sup>2,3</sup> (OIML Class E2, F1, F2)	1 g	10 µg	Modified Double Substitution; Electronic Mass Comparators
	2 g	13 µg	
	5 g	17 µg	
	10 g	20 µg	
	20 g	27 µg	
	50 g	33 µg	
	100 g	53 µg	
	200 g	0.1 mg	
	500 g	0.27 mg	
	1 kg	0.53 mg	
	2 kg	1 mg	
	5 kg	2.7 mg	
	10 kg	5.3 mg	
Mass Determination <sup>2,3</sup> (OIML Class F1, F2)	20 kg	33 mg	Modified Single Substitution; Electronic Mass Comparators
Mass Determination <sup>2,3</sup> (OIML Class M1, M2, M3)	1 mg	2 µg	Modified Single Substitution; Electronic Mass Comparators
	2 mg	2 µg	
	5 mg	2 µg	
	10 mg	2.7 µg	
	20 mg	3 µg	
	50 mg	4 µg	
	100 mg	5.3 µg	
	200 mg	6.7 µg	
	500 mg	8.3 µg	
	1 g	10 µg	
	2 g	13 µg	
	5 g	17 µg	
	10 g	20 µg	
	20 g	27 µg	
	50 g	33 µg	
	100 g	1 mg	
	200 g	2 mg	
	500 g	5 mg	
	1 kg	10 mg	
	2 kg	20 mg	
	5 kg	50 mg	
	10 kg	0.1 g	
	20 kg	0.2 g	
25 kg	0.25 g		


**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Source <sup>1</sup>	10 mHz to 2 MHz	0.000 19 % of reading	Fluke 5560A Multiproduct Calibrator
Frequency – Measure <sup>1</sup>	10 Hz to 100 MHz	0.008 % of reading	Fluke 8588A 8.5 Digit Multimeter
AC Duty Cycle – Source <sup>1</sup> Square Wave: < 3.3 Vp-p Freq.: 10 mHz to 100 kHz	(10 to 49) % Duty Cycle 10 μs to 100 s 50 % Duty Cycle 10 μs to 100 s (51 to 90) % Duty Cycle 10 μs to 100 s	0.039 % of reading + 78 ns 0.016 % of reading + 78 ns 0.039 % of reading + 78 ns	Fluke 5560A Multiproduct Calibrator

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. Laboratory maintains the ability to generate quantities between the reported uncertainties. This reported uncertainty may be larger than the cardinal points listed within this parameter.
3. Newton and Pound weights are mathematically converted to equivalent kilograms.
4. The stated uncertainty is the laboratory’s ability to source a fast rise pulse that is approximately 250 ps. In the typical application of measuring rise time of an oscilloscope, this value is one of the contributing factors, but other factors are derived from the DUT.
5. The uncertainties shown are for the most favorable conditions. There is an increase in uncertainty that corresponds to the laboratory’s AC voltage and current uncertainties at different frequencies other than the ones shown. Power factors (PF) other than the one shown contribute to the power uncertainty. PF is related to the cosine of phase. Therefore, uncertainties track the laboratory’s phase uncertainty closely at PF near one but are magnified heavily as PF approaches zero. The lab may also report reactive power, apparent power, and power factor under this accreditation. If needed, contact the laboratory for more information regarding uncertainties at frequency and power factor combinations other than the ones shown.
6. Unless otherwise specified in the Reference Standard, Method, and/or Equipment column, the calibration procedure or method used during the calibration of these parameters were developed internally and were validated according to ISO/IEC 17025:2017 by the CAB.
7. The legal entity for this client is Transcat, Inc.
8. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2489.29.



Jason Stine, Vice President