



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Transcat, Inc.
35 Vantage Point Drive
Rochester, NY 14624

Fulfills the requirements of

ISO/IEC 17025:2017

and national standards

ANSI/NCSL Z540-1-1994 (R2002) AND
ANSI/NCSL Z540.3-2006 (R2013)

In the fields of

CALIBRATION AND DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 07 September 2027
Certificate Number: AC-2489



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

ANSI/NCSL Z540.3-2006 (R2013)

Transcat, Inc.

35 Vantage Point Drive

Rochester, NY 14624

Len Gaffney 585-352-9720

CALIBRATION AND DIMENSIONAL MEASUREMENT

ISO/IEC 17025 Accreditation Granted: **04 April 2025**

Certificate Number: **AC-2489** Certificate Expiry Date: **07 September 2027**

CALIBRATION

Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH – Source ¹	4 pH 7 pH 10 pH	0.013 pH 0.013 pH 0.012 pH	Comparison to Accredited pH Solutions
Conductivity – Source	5 µS 10 µS 100 µS 1 000 µS 10 000 µS 100 000 µS 150 000 µS	0.3 µS 0.3 µS 0.88 µS 4.4 µS 46 µS 420 µS 710 µS	Comparison to Accredited Conductivity Solutions

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Sine Wave Flatness – Measure ¹	Up to 3 V 10 Hz 10 Hz to 1 MHz (1 to 10) MHz (10 to 30) MHz (30 to 50) MHz (50 to 80) MHz (80 to 100) MHz	0.2 % of reading 0.11 % of reading 0.21 % of reading 0.32 % of reading 0.36 % of reading 0.48 % of reading 0.53 % of reading	Comparison to Thermal Voltage Converter, Keysight 3458A, Opt. 002 8.5 Digit Multimeter
DC Current – Source ¹	(0 to 220) μ A (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A	0.004 1 % of reading + 6 nA 0.003 5 % of reading + 7 nA 0.003 5 % of reading + 40 nA 0.004 5 % of reading + 0.7 μ A 0.008 1 % of reading + 12 μ A	Comparison to Fluke 5730A Multiproduct Calibrator
DC Current – Source ¹	(2.2 to 11) A	0.036 % of reading + 0.48 mA	Comparison to Fluke 5720A Multiproduct Calibrator, Fluke 5725 Amplifier
DC Current – Source ¹	(11 to 100) A	0.012 % of reading + 5 mA	Comparison to Fluke 5520A Multiproduct Calibrator, Fluke 52120A Transconductance Amplifier
DC Current – Measure ¹	(0 to 100) nA 100 nA to 1 μ A (1 to 10) μ A (10 to 100) μ A 100 μ A to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1 A	0.003 6 % of reading + 40 pA 0.002 9 % of reading + 40 pA 0.002 9 % of reading + 0.1 nA 0.002 3 % of reading + 0.8 nA 0.002 4 % of reading + 5 nA 0.002 6 % of reading + 50 nA 0.003 8 % of reading + 0.5 μ A 0.011 % of reading + 10 μ A	Comparison to Keysight 3458A, Opt. 002 8.5 Digit Multimeter
DC Current – Measure ¹	(1 to 3) A (3 to 10) A	0.14 % of reading 0.18 % of reading + 0.8 mA	Comparison to Fluke 8846A 6.5 Digit Multimeter
DC Current – Measure ¹	(1 to 10) A	0.047 % of reading	Comparison to L&N 4361 Current Shunt, Keysight 3458A, Opt. 002 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure ¹	Up to 100 A	0.012 % of reading + 0.58 mA	Comparison to Ohms Labs CS-100 Precision Shunt, Keysight 3458A, Opt. 002 8.5 Digit Multimeter
DC Clamp-on Ammeter (Non-Toroidal Type) Hall Effect Sensor ¹	(20 to 150) A (150 to 1 000) A	0.5 % of reading + 0.14 A 0.52 % of reading + 0.5 A	Comparison to Fluke 5520A Multiproduct Calibrator, Fluke 5500A/Coil 50-turn Coil
DC Clamp-on Ammeter ¹	(1 000 to 5 000) A	0.58 % of reading	Comparison to Fluke 52120A Transconductance Amplifier, Fluke 5520A Multiproduct Calibrator, 3 kA or 6 kA Coil
AC Current – Source ¹	Up to 220 μ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.025 % of reading + 16 nA 0.02 % of reading + 10 nA 0.016 % of reading + 8 nA 0.03 % of reading + 12 nA 0.11 % of reading + 65 nA 0.025 % of reading + 40 nA 0.016 % of reading + 35 nA 0.012 % of reading + 35 nA 0.02 % of reading + 0.11 μ A 0.11 % of reading + 0.65 μ A 0.025 % of reading + 0.4 μ A 0.022 % of reading + 0.35 μ A 0.012 % of reading + 0.35 μ A 0.023 % of reading + 0.55 μ A 0.11 % of reading + 5 μ A	Comparison to Fluke 5730A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ¹	(22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.025 % of reading + 4 μ A 0.022 % of reading + 3.5 μ A 0.014 % of reading + 2.5 μ A 0.022 % of reading + 3.5 μ A 0.11 % of reading + 10 μ A 0.03 % of reading + 35 μ A 0.05 % of reading + 80 μ A 0.7 % of reading + 0.16 mA	Comparison to Fluke 5730A Multiproduct Calibrator
AC Current – Source ¹	(2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.046 % of reading + 0.85 mA 0.095 % of reading + 0.19 mA 0.36 % of reading + 0.8 mA	Comparison to Fluke 5730A Multiproduct Calibrator, Fluke 5725 Amplifier
AC Current – Source ¹	(11 to 20.5) A (10 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.097 % of reading + 3.9 mA 0.12 % of reading + 3.9 mA 2.3 % of reading + 3.9 mA	Comparison to Fluke 5520A Multiproduct Calibrator
AC Current – Source ¹	(20.5 to 40) A (10 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.13 % of reading + 11 mA 0.17 % of reading + 11 mA 2.3 % of reading + 11 mA	Comparison to (2) Fluke 5520A Multiproduct Calibrators in Parallel configuration.
AC Current – Source ¹	(10 to 65) Hz (20 to 33) A (33 to 40) A (40 to 60) A (60 to 100) A (65 to 300) Hz (20 to 33) A (33 to 60) A (60 to 100) A 300 Hz to 1 kHz (20 to 33) A (33 to 60) A (60 to 100) A	0.032 % of reading + 24 mA 0.04 % of reading + 36 mA 0.04 % of reading + 36 mA 0.037 % of reading + 36 mA 0.044 % of reading + 36 mA 0.052 % of reading + 45 mA 0.037 % of reading + 45 mA 0.12 % of reading + 0.12 A 0.12 % of reading + 0.12 A 0.12 % of reading + 0.12 A	Comparison to Fluke 52120A Transconductance Amplifier, Fluke 5520A Multiproduct Calibrator
AC Current – Source ¹ Extended Frequency Ranges	(10 to 30) kHz (29 to 330) μ A (0.33 to 3.3) mA (3.3 to 33) mA (33 to 330) mA	1.2 % of reading + 0.31 μ A 0.78 % of reading + 0.47 μ A 0.31 % of reading + 3.1 μ A 0.31 % of reading + 0.16 mA	Comparison to Fluke 5520A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Clamp-on Ammeters (Toroidal Type) Transformer Type Sensor ¹	(20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz	0.3 % of reading + 26 mA 0.83 % of reading + 47 mA 0.35 % of reading + 0.12 A 1.1 % of reading + 0.22 A	Comparison to Fluke 5520A Multiproduct Calibrator, Fluke 5500A/COIL 50-turn Coil
AC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor ¹	(20 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 000) A (45 to 65) Hz (65 to 440) Hz	0.57 % of reading + 0.26 A 1 % of reading + 0.29 A 0.6 % of reading + 0.9 A 1.3 % of reading + 0.92 A	Comparison to Fluke 5520A Multiproduct Calibrator, Fluke 5500A/COIL 50-turn Coil
AC Clamp-on Ammeters (Non-Toroidal Type) Hall Effect Sensor ¹	(1 to 6) kA (10 to 300) Hz (1 to 2) kA (300 to 440) Hz (2 to 6) kA (300 to 440) Hz	0.6 % of reading 0.8 % of reading 0.66 % of reading	Comparison to Fluke 52120A Transconductance Amplifier, Fluke 5520A Multiproduct Calibrator, 3 kA or 6 kA Coil
AC Current – Measure ¹	Up to 100 μ A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 1 kHz (0.1 to 1) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (1 to 10) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (10 to 100) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.46 % of reading + 35 nA 0.17 % of reading + 35 nA 0.072 % of reading + 35 nA 0.072 % of reading + 35 nA 0.46 % of reading + 0.23 μ A 0.17 % of reading + 0.23 μ A 0.07 % of reading + 0.23 μ A 0.038 % of reading + 0.23 μ A 0.46 % of reading + 2.3 μ A 0.17 % of reading + 2.3 μ A 0.071 % of reading + 2.3 μ A 0.038 % of reading + 2.3 μ A 0.48 % of reading + 23 μ A 0.17 % of reading + 23 μ A 0.071 % of reading + 23 μ A 0.037 % of reading + 23 μ A	Comparison to Keysight 3458A, Opt. 002 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ¹	100 mA to 1 A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.46 % of reading + 0.23 mA 0.19 % of reading + 0.23 mA 0.097 % of reading + 0.23 mA 0.12 % of reading + 0.23 mA	Comparison to Keysight 3458A, Opt. 002 8.5 Digit Multimeter
AC Current – Measure ¹	(1 to 3) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz (5 to 10) kHz (3 to 10) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz (5 to 10) kHz	1.3 % of reading + 2 mA 0.41 % of reading + 2 mA 0.18 % of reading + 2 mA 0.41 % of reading + 24 mA 1.1 % of reading + 13 mA 0.41 % of reading + 7 mA 0.18 % of reading + 7 mA 0.42 % of reading + 81 mA	Comparison to Fluke 8846A 6.5 Digit Multimeter
AC Current – Measure ¹	Up to 60 Hz Up to 10 A (10 to 20) A (20 to 40) A (40 to 60) A (60 to 100) A	0.019 % of reading + 2.3 mA 0.02 % of reading + 2.3 mA 0.029 % of reading + 2.3 mA 0.03 % of reading + 2.3 mA 0.025 % of reading + 2.3 mA	Comparison to Current Shunt, Keysight 3458A, Opt. 002 8.5 Digit Multimeter
AC Current – Measure ¹	(60 to 100) Hz Up to 5 A (5 to 20) A (20 to 60) A (60 to 100) A 100 Hz to 1 kHz Up to 100 A	0.023 % of reading + 2.3 mA 0.024 % of reading + 2.3 mA 0.032 % of reading + 2.3 mA 0.037 % of reading + 2.3 mA 0.12 % of reading + 2.3 mA	Comparison to Ohms Labs CS-100 Current Shunt, Keysight 3458A, Opt. 002 8.5 Digit Multimeter
DC Resistance – Source/Measure ¹	Up to 10 Ω (10 to 100) Ω (0.1 to 1) k Ω (1 to 10) k Ω (10 to 100) k Ω (0.1 to 1) M Ω (1 to 10) M Ω (10 to 100) M Ω (0.1 to 1) G Ω	0.001 8 % of reading + 58 $\mu\Omega$ 0.001 5 % of reading + 0.58 m Ω 0.001 3 % of reading + 0.58 m Ω 0.001 2 % of reading + 5.8 m Ω 0.001 3 % of reading + 58 m Ω 0.002 1 % of reading + 2.3 Ω 0.006 2 % of reading + 0.12 k Ω 0.059 % of reading + 1.2 k Ω 0.82 % of reading + 12 k Ω	Comparison to Keysight 3458A, Opt. 002 8.5 Digit Multimeter, Decade Resistor
DC Resistance – Source ¹ (Fixed Artifact)	1 m Ω	0.12 $\mu\Omega$	Comparison to Ohms Labs CS-100 Current Shunt

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Resistance – Source ¹ (Fixed Artifact)	1 Ω	12 $\mu\Omega$	Comparison to Leeds & Northrup 4020-B Standard Resistor
DC Resistance – Source ¹ (Variable Artifact)	10 m Ω 0.1 Ω	1 $\mu\Omega$ 10 $\mu\Omega$	Comparison to Guideline 9200 Multiple Standard Resistor
DC Resistance – Source ¹ (Variable Artifact)	1 G Ω 10 G Ω	5.8 M Ω 0.12 G Ω	Comparison to Biddle Megadek, High Resistance Decade Box
DC Resistance – Source ¹ Up to 10 kV (Variable Artifact)	(100 to 200) k Ω (300 to 700) k Ω (0.8 to 1) M Ω (1 to 10) M Ω (20 to 100) M Ω (200 to 900) M Ω (1 to 9) G Ω (10 to 90) G Ω (0.1 to 1) T Ω	0.035 % of reading 0.035 % of reading 0.035 % of reading 0.035 % of reading + 1.2 $\mu\Omega/V$ 0.12 % of reading + 1.2 $\mu\Omega/V$ 0.3 % of reading + 1.2 $\mu\Omega/V$ 0.59 % of reading + 1.2 $\mu\Omega/V$ 1.2 % of reading + 2.3 $\mu\Omega/V$ 1.2 % of reading + 5.8 $\mu\Omega/V$	Comparison to IET High Resistance Decade Substituter
DC Resistance – Source ¹ (Fixed Artifact)	(0.5 to 5) kV 1 T Ω	6 G Ω	Comparison to IET Standard Resistor
DC Resistance – Source ¹ (Fixed Artifact)	500 V 10 T Ω 1 kV 10 T Ω 2.5 kV 10 T Ω 5 kV 10 T Ω	86 G Ω 0.15 T Ω 0.1 T Ω 60 G Ω	Comparison to IET Standard Resistors
DC Voltage – Source ¹	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V	0.000 78 % of reading + 0.4 μV 0.000 5 % of reading + 0.7 μV 0.000 35 % of reading + 2.5 μV 0.000 35 % of reading + 4 μV 0.000 5 % of reading + 40 μV	Comparison to Fluke 5730A Multiproduct Calibrator
DC Voltage – Source ¹	(220 to 1 100) V	0.000 65 % of reading + 0.4 mV	Comparison to Fluke 5730A Multiproduct Calibrator, Fluke 5725A Amplifier

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Measure ¹	(0 to 100) mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 500) V (500 to 800) V (800 to 1 000) V	0.000 8 % of reading + 0.58 μ V 0.000 5 % of reading + 0.58 μ V 0.000 5 % of reading + 0.58 μ V 0.000 8 % of reading + 35 μ V 0.001 5 % of reading + 0.12 mV 0.001 8 % of reading + 0.12 mV 0.002 1 % of reading + 0.12 mV	Comparison to Keysight 3458A, Opt. 002 8.5 Digit Multimeter
DC High Voltage – Measure ¹	(1 to 10) kV	0.039 % of reading + 0.12 V	Comparison to Vitretek 4700 Digital HV Meter
DC High Voltage – Measure ¹	(10 to 20) kV (20 to 35) kV	0.036 % of reading + 1.2 V 0.064 % of reading + 1.2 V	Comparison to Vitretek 4700 Digital HV Meter, Vitretek HVL-35 High Voltage Probe
DC High Voltage – Measure ¹	(15 to 30) kV (30 to 45) kV (45 to 70) kV	0.039 % of reading + 1.2 V 0.049 % of reading + 1.2 V 0.088 % of reading + 1.2 V	Comparison to Vitretek 4700 Digital HV Meter, Vitretek HVL-70 High Voltage Probe
DC High Voltage – Measure ¹	(25 to 100) kV	0.17 % of reading + 1.2 V	Comparison to Vitretek 4700 Digital HV Meter, Vitretek HVL-100 High Voltage Probe
AC Voltage – Source ¹	Up to 2.2 mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.024 % of reading + 4 μ V 0.009 1 % of reading + 4 μ V 0.008 % of reading + 4 μ V 0.025 % of reading + 4 μ V 0.05 % of reading + 5 μ V 0.11 % of reading + 10 μ V 0.14 % of reading + 20 μ V 0.27 % of reading + 20 μ V	Comparison to Fluke 5730A Multiproduct Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(2.2 to 22) mV		Comparison to Fluke 5730A Multiproduct Calibrator
	(10 to 20) Hz	0.024 % of reading + 4 μ V	
	(20 to 40) Hz	0.009 % of reading + 4 μ V	
	40 Hz to 20 kHz	0.008 % of reading + 4 μ V	
	(20 to 50) kHz	0.02 % of reading + 4 μ V	
	(50 to 100) kHz	0.05 % of reading + 5 μ V	
	(100 to 300) kHz	0.11 % of reading + 10 μ V	
	(300 to 500) kHz	0.14 % of reading + 20 μ V	
	500 kHz to 1 MHz	0.27 % of reading + 20 μ V	
	(22 to 220) mV		
	(10 to 20) Hz	0.024 % of reading + 12 μ V	
	(20 to 40) Hz	0.009 3 % of reading + 7 μ V	
	40 Hz to 20 kHz	0.005 9 % of reading + 7 μ V	
	(20 o 50) kHz	0.012 % of reading + 7 μ V	
	(50 to 100) kHz	0.031 % of reading + 17 μ V	
	(100 to 300) kHz	0.066 % of reading + 20 μ V	
	(300 to 500) kHz	0.14 % of reading + 25 μ V	
	500 kHz to 1 MHz	0.27 % of reading + 45 μ V	
	(0.22 to 2.2) V		
	(10 to 20) Hz	0.024 % of reading + 40 μ V	
	(20 to 40) Hz	0.009 4 % of reading + 15 μ V	
	40 Hz to 20 kHz	0.004 6 % of reading + 8 μ V	
	(20 o 50) kHz	0.006 9 % of reading + 10 μ V	
	(50 to 100) kHz	0.008 6 % of reading + 30 μ V	
	(100 to 300) kHz	0.034 % of reading + 80 μ V	
	(300 to 500) kHz	0.1 % of reading + 0.2 mV	
	500 kHz to 1 MHz	0.17 % of reading + 0.3 mV	
	(2.2 to 22) V		
	(10 to 20) Hz	0.024 % of reading + 0.4 mV	
	(20 to 40) Hz	0.009 3 % of reading + 0.15 mV	
	40 Hz to 20 kHz	0.004 7 % of reading + 50 μ V	
	(20 o 50) kHz	0.007 1 % of reading + 0.1 mV	
	(50 to 100) kHz	0.008 4 % of reading + 0.2 mV	
	(100 to 300) kHz	0.025 % of reading + 0.6 mV	
	(300 to 500) kHz	0.1 % of reading + 2 mV	
	500 kHz to 1 MHz	0.15 % of reading + 3.2 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(22 to 220) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.024 % of reading + 4 mV 0.009 4 % of reading + 1.5 mV 0.005 5 % of reading + 0.6 mV 0.008 1 % of reading + 1 mV 0.015 % of reading + 2.5 mV 0.09 % of reading + 16 mV 0.44 % of reading + 40 mV 0.8 % of reading + 80 mV	Comparison to Fluke 5730A Multiproduct Calibrator
AC Voltage – Source ¹	(220 to 250) V (15 to 50) Hz (220 to 750) V (30 to 50) kHz (50 to 100) kHz (220 to 1 100) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	0.03 % of reading + 11 mV 0.06 % of reading 0.23 % of reading 0.009 3 % of reading 0.007 3 % of reading 0.06 % of reading	Comparison to Fluke 5730A Multiproduct Calibrator, Fluke 5725A Amplifier
AC Voltage – Measure ¹	Up to 10 mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 4) MHz (10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	0.04 % of reading + 3.5 μ V 0.03 % of reading + 1.2 μ V 0.04 % of reading + 1.2 μ V 0.15 % of reading + 1.2 μ V 0.59 % of reading + 1.2 μ V 4.6 % of reading + 2.3 μ V 1.5 % of reading + 5.8 μ V 8.1 % of reading + 8.1 μ V 0.013 % of reading + 4.6 μ V 0.009 7 % of reading + 2.3 μ V 0.017 % of reading + 2.3 μ V 0.038 % of reading + 2.3 μ V 0.093 % of reading + 2.3 μ V 0.36 % of reading + 12 μ V 1.2 % of reading + 12 μ V 1.8 % of reading + 12 μ V 4.7 % of reading + 81 μ V 4.7 % of reading + 92 μ V 17 % of reading + 0.12 mV	Comparison to Keysight 3458A, Opt. 002 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ¹	(0.1 to 1) V		Comparison to Keysight 3458A, Opt. 002 8.5 Digit Multimeter
	(1 to 40) Hz	0.008 8 % of reading + 46 μ V	
	40 Hz to 1 kHz	0.008 3 % of reading + 23 μ V	
	(1 to 20) kHz	0.017 % of reading + 23 μ V	
	(20 to 50) kHz	0.036 % of reading + 23 μ V	
	(50 to 100) kHz	0.093 % of reading + 23 μ V	
	(100 to 300) kHz	0.35 % of reading + 0.12 mV	
	300 kHz to 1 MHz	1.2 % of reading + 0.12 mV	
	(1 to 2) MHz	1.8 % of reading + 0.12 mV	
	(2 to 4) MHz	4.6 % of reading + 0.81 mV	
	(4 to 8) MHz	4.6 % of reading + 0.92 mV	
	(8 to 10) MHz	17 % of reading + 1.2 mV	
	(1 to 10) V		
	(1 to 40) Hz	0.009 5 % of reading + 0.46 mV	
	40 Hz to 1 kHz	0.023 % of reading + 0.23 mV	
	(1 to 20) kHz	0.017 % of reading + 0.23 mV	
	(20 to 50) kHz	0.036 % of reading + 0.23 mV	
	(50 to 100) kHz	0.093 % of reading + 0.23 mV	
	(100 to 300) kHz	0.35 % of reading + 1.2 mV	
	300 kHz to 1 MHz	1.2 % of reading + 1.2 mV	
	(1 to 2) MHz	1.8 % of reading + 1.2 mV	
	(2 to 4) MHz	4.6 % of reading + 8.1 mV	
	(4 to 8) MHz	4.6 % of reading + 9.2 mV	
	(8 to 10) MHz	17 % of reading + 12 mV	
	(10 to 100) V		
	(1 to 40) Hz	0.024 % of reading + 4.6 mV	
	40 Hz to 1 kHz	0.024 % of reading + 2.3 mV	
	(1 to 20) kHz	0.024 % of reading + 2.3 mV	
	(20 to 50) kHz	0.041 % of reading + 2.3 mV	
	(50 to 100) kHz	0.14 % of reading + 2.3 mV	
	(100 to 300) kHz	0.46 % of reading + 12 mV	
	300 kHz to 1 MHz	1.7 % of reading + 12 mV	
	(100 to 700) V		
	(1 to 40) Hz	0.048 % of reading + 46 mV	
	40 Hz to 1 kHz	0.048 % of reading + 23 mV	
	(1 to 20) kHz	0.071 % of reading + 23 mV	
	(20 to 50) kHz	0.19 % of reading + 23 mV	
	(50 to 100) kHz	0.35 % of reading + 23 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ¹	Up to 1 mV 100 kHz to 1 MHz (1 to 3) MHz (3 to 10) MHz (10 to 20) MHz	2 % of reading + 2.4 μ V 3.8 % of reading + 2.4 μ V 10 % of reading + 2.4 μ V 25 % of reading + 2.4 μ V	Comparison to Rohde & Schwarz URE3 RMS Voltmeter
	(1 to 3) mV 100 kHz to 1 MHz (1 to 3) MHz (3 to 10) MHz (10 to 20) MHz	1 % of reading + 2 μ V 3.8 % of reading + 2 μ V 11 % of reading + 2 μ V 25 % of reading + 2 μ V	
	(3 to 100) mV 100 kHz to 1 MHz (1 to 3) MHz (3 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.98 % of reading + 3 μ V 1.9 % of reading + 3 μ V 3.2 % of reading + 3 μ V 7.6 % of reading + 3 μ V 16 % of reading + 3 μ V	
AC High Voltage – Measure ¹	(0.7 to 10) kV 60 Hz	0.15 % of reading + 0.21 V	Comparison to Vitretek 4700 Digital HV Meter
AC High Voltage – Measure ¹	(10 to 20) kV 60 Hz	0.07 % of reading + 1.2 V	Comparison to Vitretek 4700 Digital HV Meter, Vitretek HVP-35 High Voltage Probe
	(20 to 35) kV 60 Hz	0.08 % of reading + 1.2 V	
AC High Voltage – Measure ¹	(12.5 to 25) kV 60 Hz	0.094 % of reading + 1.3 V	Comparison to Vitretek 4700 Digital HV Meter, Vitretek HVL-70 High Voltage Probe
	(25 to 37.5) kV 60 Hz	0.096 % of reading + 1.3 V	
	(37.5 to 50) kV 60 Hz	0.1 % of reading + 1.3 V	
AC High Voltage – Measure ¹	25 kV to 75 kV 60 Hz	0.17 % of reading + 1.4 V	Comparison to Vitretek 4700 Digital HV Meter, Vitretek HVL-100 High Voltage Probe

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source ^{1,2} (Simulation)	220 pF to 399.99 pF 0.4 nF to 1.0999 nF 1.1 nF to 3.2999 nF 3.3 nF to 10.9999 nF 11 nF to 32.9999 nF 33 nF to 109.99 nF 110 nF to 329.999 nF 0.33 μF to 1.0999 μF 1.1 μF to 3.2999 μF 3.3 μF to 10.9999 μF 11 μF to 32.9999 μF 33 μF to 109.999 μF 110 μF to 329.999 μF 0.33 mF to 1.099999 mF 1.1 mF to 3.29999 mF 3.3 mF to 10.9999 mF 11 mF to 32.9999 mF 33 mF to 110 mF	0.45 % of reading + 7.8 pF 0.4 % of reading + 7.8 pF 0.4 % of reading + 0.0078 nF 0.22 % of reading + 0.0078 nF 0.22 % of reading + 0.078 nF 0.22 % of reading + 0.078 nF 0.22 % of reading + 0.23 nF 0.21 % of reading + 0.78 nF 0.22 % of reading + 2.3 nF 0.21 % of reading + 7.8 nF 0.32 % of reading + 23 nF 0.38 % of reading + 0.078 μF 0.35 % of reading + 0.23 μF 0.35 % of reading + 0.78 μF 0.35 % of reading + 2.3 μF 0.35 % of reading + 7.8 μF 0.58 % of reading + 23 μF 0.85 % of reading + 78 μF	Comparison to Fluke 5522A Multifunction Calibrator
Capacitance – Source ^{1,2} (Fixed Artifacts)	100 Hz to 1 kHz (0.1 to 0.7) nF (0.7 to 600) nF 600 nF to 1.4 μF	0.1 % of reading + 0.54 pF 0.16 % of reading + 0.1 pF 0.045 % of reading + 0.5 nF	Comparison to Arco SS32 Precision Standard Capacitor Set
Capacitance – Source ^{1,2} (Variable Artifact)	1 kHz 100 pF to 1 nF (1.1 to 10.1) nF (10.1 to 100.1) nF 100.1 nF to 1.1111 μF	0.61 fF/pF + 1.3 fF 0.17 pF/nF + 12 pF 0.57 pF/nF + 4.6 pF 0.61 pF/nF + 2.1 pF	Comparison to GenRad 1423A Decade Capacitor
Capacitance – Measure ²	0.1 pF 100 kHz 1 MHz 1 pF 10 kHz 100 kHz 1 MHz 10 pF 1 kHz 10 kHz 100 kHz 1 MHz	1.4 % of reading 1.8 % of reading 1.4 % of reading 0.37 % of reading 0.44 % of reading 1.4 % of reading 0.28 % of reading 0.28 % of reading 0.3 % of reading	Comparison to Agilent E4980AL LCR Meter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Measure ²	100 pF		Comparison to Agilent E4980AL Comparison to LCR Meter
	100 Hz	2.1 % of reading	
	1 kHz	0.23 % of reading	
	10 kHz	0.18 % of reading	
	100 kHz	0.21 % of reading	
	1 MHz	0.23 % of reading	
	1 nF		
	20 Hz	1.8 % of reading	
	100 Hz	0.3 % of reading	
	1 kHz	0.1 % of reading	
	10 kHz	0.1 % of reading	
	100 kHz	0.1 % of reading	
	1 MHz	0.14 % of reading	
	10 nF		
	20 Hz	0.31 % of reading	
	100 Hz	0.12 % of reading	
	1 kHz	0.1 % of reading	
	10 kHz	0.1 % of reading	
	100 kHz	0.1 % of reading	
	1 MHz	0.25 % of reading	
	100 nF		
	20 Hz	0.16 % of reading	
	100 Hz	0.1 % of reading	
	1 kHz	0.1 % of reading	
	10 kHz	0.1 % of reading	
	100 kHz	0.18 % of reading	
	1 MHz	0.33 % of reading	
	1 µF		
	20 Hz	0.15 % of reading	
	100 Hz	0.1 % of reading	
	1 kHz	0.1 % of reading	
	10 kHz	0.18 % of reading	
	100 kHz	0.25 % of reading	
	1 MHz	0.79 % of reading	
	10 µF		
	20 Hz	0.15 % of reading	
	100 Hz	0.01 % of reading	
	1 kHz	0.16 % of reading	
	10 kHz	0.28 % of reading	
	100 kHz	0.73 % of reading	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Measure ²	100 μ F		Comparison to Agilent E4980AL Comparison to LCR Meter
	20 Hz	0.16 % of reading	
	100 Hz	0.17 % of reading	
	1 kHz	0.29 % of reading	
	10 kHz	0.8 % of reading	
Impedance – Measure ^{1,2} (AC Resistance)	0.1 Ω		Comparison to Agilent E4980AL LCR Meter
	1 kHz	2 % of reading	
	10 kHz	1.2 % of reading	
	100 kHz	1.1 % of reading	
	1 MHz	1.2 % of reading	
	1 Ω		
	20 Hz	0.67 % of reading	
	100 Hz	0.45 % of reading	
	1 kHz	0.36 % of reading	
	10 kHz	0.33 % of reading	
	100 kHz	0.31 % of reading	
	1 MHz	0.38 % of reading	
	10 Ω		
	20 Hz	0.29 % of reading	
	100 Hz	0.2 % of reading	
	1 kHz	0.17 % of reading	
	10 kHz	0.18 % of reading	
	100 kHz	0.18 % of reading	
	1 MHz	0.31 % of reading	
	100 Ω		
	20 Hz	0.16 % of reading	
	100 Hz	0.1 % of reading	
	1 kHz	0.1 % of reading	
	10 kHz	0.12 % of reading	
	100 kHz	0.12 % of reading	
	1 MHz	0.2 % of reading	
	1 k Ω		
	20 Hz	0.15 % of reading	
	100 Hz	0.1 % of reading	
	1 kHz	0.1 % of reading	
	10 kHz	0.1 % of reading	
	100 kHz	0.1 % of reading	
	1 MHz	0.14 % of reading	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Impedance – Measure ^{1,2} (AC Resistance)	10 k Ω		Comparison to Agilent E4980AL LCR Meter
	20 Hz	0.15 % of reading	
	100 Hz	0.1 % of reading	
	1 kHz	0.1 % of reading	
	10 kHz	0.1 % of reading	
	100 kHz	0.1 % of reading	
	1 MHz	0.29 % of reading	
	100 k Ω		
	20 Hz	0.17 % of reading	
	100 Hz	0.1 % of reading	
	1 kHz	0.1 % of reading	
	10 kHz	0.17 % of reading	
	100 kHz	0.28 % of reading	
	1 MHz	0.38 % of reading	
Inductance – Source ^{1,2} (Fixed Artifact)	100 mH		Comparison to Standard Inductor
	1 kHz	0.14 mH	
Inductance – Measure ²	1 μ H		Comparison to Agilent E4980AL LCR Meter
	10 kHz	1.6 % of reading	
	100 kHz	0.36 % of reading	
	1 MHz	0.27 % of reading	
	10 μ H		
	10 kHz	0.37 % of reading	
	100 kHz	0.2 % of reading	
	1 MHz	0.2 % of reading	
	100 μ H		
	1 kHz	0.41 % of reading	
	10 kHz	0.2 % of reading	
	100 kHz	0.12 % of reading	
	1 MHz	0.14 % of reading	
	1 mH		
	100 Hz	0.56 % of reading	
	1 kHz	0.19 % of reading	
	10 kHz	0.12 % of reading	
	100 kHz	0.1 % of reading	
	1 MHz	0.23 % of reading	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Measure ^{1,2}	10 mH		Comparison to Agilent E4980AL LCR Meter
	20 Hz	0.86 % of reading	
	100 Hz	0.22 % of reading	
	1 kHz	0.1 % of reading	
	10 kHz	0.1 % of reading	
	100 kHz	0.11 % of reading	
	1 MHz	0.35 % of reading	
	100 mH		
	20 Hz	0.28 % of reading	
	100 Hz	0.11 % of reading	
	1 kHz	0.1 % of reading	
	10 kHz	0.1 % of reading	
	100 kHz	0.21 % of reading	
	1 MHz	0.88 % of reading	
	10 H		
	20 Hz	0.15 % of reading	
	100 Hz	0.1 % of reading	
	1 kHz	0.11 % of reading	
	10 kHz	0.21 % of reading	
	100 kHz	0.69 % of reading	
	100 H		
	20 Hz	0.15 % of reading	
	100 Hz	0.11 % of reading	
	1 kHz	0.15 % of reading	
	10 kHz	0.62 % of reading	
Current Harmonics – Source ^{1,3}	Carrier Range to 0.25 A Harmonic: Up to 75 mA (16 to 850) Hz	61 μ A/A + 21 μ A	Comparison to Fluke 6105A Electrical Power Quality Calibrator
	850 Hz to 6.5 kHz	0.46 mA/A + 22 μ A	
	Carrier Range to 0.5 A Harmonic: Up to 0.15 A (16 to 850) Hz	70 μ A/A + 21 μ A	
	850 Hz to 6.5 kHz	0.46 mA/A + 23 μ A	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Current Harmonics – Source ^{1,3}	Carrier Range to 1 A Harmonic: Up to 0.3 A (16 to 850) Hz 850 Hz to 6.5 kHz	70 μ A/A + 29 μ A 0.46 mA/A + 29 μ A	Comparison to Fluke 6105A Electrical Power Quality Calibrator
	Carrier Range to 2 A Harmonic: Up to 0.6 A (16 to 850) Hz 850 Hz to 6.5 kHz	70 μ A/A + 0.1 mA 0.46 mA/A + 0.1 mA	
	Carrier Range to 5 A Harmonic: Up to 1.5 A (16 to 850) Hz 850 Hz to 6.5 kHz	70 μ A/A + 0.1 mA 0.46 A/A + 0.1 mA	
	Carrier Range to 10 A Harmonic: Up to 3 A (16 to 850) Hz 850 Hz to 6.5 kHz	74 μ A/A + 0.29 mA 0.46 mA/A + 0.29 mA	
	Carrier Range to 20 A Harmonic: Up to 6 A (16 to 850) Hz 850 Hz to 6.5 kHz	75 μ A/A + 0.45 mA 0.46 mA/A + 0.45 mA	
Voltage Harmonics – Source ^{1,3}	Carrier Range to 23 V Harmonic: Up to 6.9 V (16 to 850) Hz 850 Hz to 6.5 kHz	58 μ V/V + 1 mV 0.52 mV/V + 1 mV	Comparison to Fluke 6105A Electrical Power Quality Calibrator
	Carrier Range to 45 V Harmonic: Up to 13.5 V (16 to 850) Hz 850 Hz to 6.5 kHz	67 μ V/V + 2 mV 0.52 mV/V + 2 mV	
Voltage Harmonics – Source ^{1,3}	Carrier Range to 90 V Harmonic: Up to 27 V (16 to 850) Hz 850 Hz to 6.5 kHz	69 μ V/V + 2 mV 0.52 mV/V + 2 mV	Comparison to Fluke 6105A Electrical Power Quality Calibrator
	Carrier Range to 135 V Harmonic: Up to 40.5 V (16 to 850) Hz 850 Hz to 6.5 kHz	69 μ V/V + 2 mV 0.52 mV/V + 2 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Voltage Harmonics – Source ^{1,3}	Carrier Range to 180 V Harmonic: Up to 54 V (16 to 850) Hz 850 Hz to 6.5 kHz	69 μ V/V + 6 mV 0.52 mV/V + 6 mV	Comparison to Fluke 6105A Electrical Power Quality Calibrator
	Carrier Range to 360 V Harmonic: Up to 108 V (16 to 850) Hz 850 Hz to 6.5 kHz	69 μ V/V + 13 mV 0.52 V/V + 13 mV	
	Carrier Range to 650 V Harmonic: Up to 195 V (16 to 850) Hz 850 Hz to 6.5 kHz	70 μ V/V + 22 mV 0.52 mV/V + 22 mV	
	Carrier Range to 1 008 V Harmonic: Up to 302 V (16 to 850) Hz 850 Hz to 6.5 kHz	70 μ V/V + 33 mV 0.52 mV/V + 33 mV	
DC Power – Source ¹ (0.33 to 330) mA	11 μ W to 1.1 mW 1.1 mW to 0.11 W (0.11 to 110) W (110 to 330) W	0.024 % of reading 0.027 % of reading 0.024 % of reading 0.018 % of reading	Comparison to Fluke 5520A Multiproduct Calibrator
(0.33 to 3) A	11 μ W to 110 mW (0.11 to 990) W (0.99 to 3) kW	0.044 % of reading 0.053 % of reading 0.009 6 % of reading	
(3 to 20.5) A	99 mW to 0.99 W 0.99 W to 6.8 kW (6.8 to 20.5) kW	0.088 % of reading 0.07 % of reading 0.04 % of reading	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power – Source ^{1,4} PF = 1			
(3.3 to 9) mA	(10 to 65) Hz (0.11 mW to 3) mW 3 mW to 9 W	0.13 % of reading 0.077 % of reading	Comparison to Fluke 5520A Multiproduct Calibrator
(9 to 33) mA	(10 to 65) W (0.3 to 10) mW 10 mW to 33 W	0.089 % of reading 0.077 % of reading	
(33 to 90) mA	(10 to 65) Hz (1 to 30) mW 30 mW to 90 W	0.071 % of reading 0.057 % of reading	
(90 to 330) mA	(10 to 65) Hz (3 to 100) mW 100 mW to 300 W	0.089 % of reading 0.078 % of reading	
(0.33 to 0.9) A	(10 to 65) Hz (11 to 300) mW (0.3 to 900) W	0.071 % of reading 0.081 % of reading	
(0.9 to 2.2) A	(10 to 65) Hz (30 to 720) mW 0.72 W to 2 kW	0.089 % of reading 0.079 % of reading	
(2.2 to 4.5) A	(10 to 65) Hz 80 mW to 1.4 W 1.4 W to 4.5 kW	0.088 % of reading 0.18 % of reading	
(4.5 to 20.5) A	(10 to 65) Hz 150 mW to 20 kW	0.17 % of reading	
AC Power – Source ^{1,4} PF = 1			Comparison to Fluke 6105A Electrical Power Quality Calibrator
(0.5 to 20) A	(16 to 850) Hz 23 W to 13 kW	0.024 % of reading	
Phase – Source ¹	(0 to 90)° (10 to 65) Hz (65 to 500) Hz 500 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.11° 0.2° 0.39° 1.9° 3.9° 7.8°	Comparison to Fluke 5520A Multiproduct Calibrator
Total Harmonic Distortion	5 Hz to 500 kHz 500 kHz to 1 MHz	18 % of reading + 0.13 % THD 29 % of reading + 0.7 % THD	Comparison to Kron-Hite 6900B Distortion Analyzer

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹	Type B		Comparison to Ectron 1140A Thermocouple Calibrator/Simulator
	(250 to 350) °C	1.2 °C	
	(350 to 445) °C	0.9 °C	
	(445 to 580) °C	0.71 °C	
	(580 to 750) °C	0.55 °C	
	(750 to 1 000) °C	0.45 °C	
	(1 000 to 1 820) °C	0.35 °C	
	Type C		
	(0 to 250) °C	0.21 °C	
	(250 to 1 000) °C	0.17 °C	
	(1 000 to 1 500) °C	0.19 °C	
	(1 500 to 1 800) °C	0.22 °C	
	(1 800 to 2 000) °C	0.24 °C	
	(2 000 to 2 250) °C	0.30 °C	
	(2 250 to 2 315) °C	0.33 °C	
	Type E		
	(-270 to -245) °C	1.6 °C	
	(-245 to -195) °C	0.24 °C	
	(-195 to -155) °C	0.12 °C	
	(-155 to -90) °C	0.095 °C	
	(-90 to 0) °C	0.08 °C	
	(0 to 15) °C	0.076 °C	
	(15 to 890) °C	0.064 °C	
	(890 to 1 000) °C	0.074 °C	
	Type J		
	(-210 to -180) °C	0.15 °C	
	(-180 to -120) °C	0.12 °C	
	(-120 to -50) °C	0.093 °C	
	(-50 to 990) °C	0.08 °C	
	(990 to 1 200) °C	0.094 °C	
	Type K		
	(-270 to -255) °C	2.5 °C	
	(-255 to -195) °C	0.85 °C	
	(-195 to -115) °C	0.16 °C	
	(-115 to -55) °C	0.12 °C	
	(-55 to 1 000) °C	0.087 °C	
	(1 000 to 1 372) °C	0.096 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Measure/Source ¹	Type N		Comparison to Ectron 1140A Thermocouple Calibrator/Simulator
	(-270 to -260) °C	5.4 °C	
	(-260 to -200) °C	1.5 °C	
	(-200 to -140) °C	0.29 °C	
	(-140 to -70) °C	0.18 °C	
	(-70 to 25) °C	0.14 °C	
	(-25 to 160) °C	0.12 °C	
	(160 to 1 300) °C	0.11 °C	
	Type R		
	(-50 to -30) °C	0.8 °C	
	(-30 to 45) °C	0.69 °C	
	(45 to 160) °C	0.49 °C	
	(160 to 380) °C	0.35 °C	
	(380 to 775) °C	0.3 °C	
	(775 to 1 768) °C	0.26 °C	
	Type S		
	(-50 to -30) °C	0.76 °C	
	(-30 to 45) °C	0.68 °C	
	(45 to 105) °C	0.49 °C	
	(105 to 310) °C	0.41 °C	
	(310 to 615) °C	0.35 °C	
	(615 to 1 768) °C	0.31 °C	
	Type T		
	(-270 to -255) °C	1.9 °C	
	(-255 to -240) °C	0.6 °C	
	(-240 to -210) °C	0.36 °C	
	(-210 to -150) °C	0.22 °C	
	(-150 to -40) °C	0.15 °C	
	(-40 to 100) °C	0.095 °C	
	(100 to 400) °C	0.08 °C	
Oscilloscopes ^{1,8} Amplitude – DC into 50 Ω load into 1 MΩ load	(-5 to 5) V (-200 to 200) V	0.023 % of reading + 19 μV 0.023 % of reading + 19 μV	Comparison to Fluke 9500B Oscilloscope Calibrator, Fluke 9500B/3200 Oscilloscope Calibrator, Fluke 9530 3.2 GHz Active Head, Fluke 9550 Active Head w/ 25 ps Capability

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes ^{1,8} Amplitude – Square Wave Rate: 10 Hz to 10 kHz into 50 Ω load	40 μ Vp-p to 1 mVp-p 1 mVp-p to 5 Vp-p	0.78 % of reading + 7.8 μ V 0.078 % of reading + 7.8 μ V	Comparison to Fluke 9500B Oscilloscope Calibrator, Fluke 9500B/3200 Oscilloscope Calibrator, Fluke 9530 3.2 GHz Active Head, Fluke 9550 Active Head w/ 25 ps Capability
into 1 M Ω load	40 μ Vp-p to 1 mVp-p 1 mVp-p to 200 Vp-p	0.78 % of reading + 7.8 μ V 0.078 % of reading + 7.8 μ V	
Time Markers 100 mVp-p to 1 Vp-p into 50 Ω load			
Square Wave	9.009 1 ns to 83 μ s 83 μ s to 55s	0.19 μ s/s 2.3 μ s/s	
Sine Wave	450.5 ps to 9.009 ns	0.19 μ s/s	
Pulse	900.91 ns to 83 μ s 83 μ s to 55s	0.19 μ s/s 2.3 μ s/s	
Triangle Wave	900.91 ns to 83 μ s 83 μ s to 55s	0.19 μ s/s 2.3 μ s/s	
Rise Time into 50 Ω load Rate: 10 Hz to 2 MHz	5 mVp-p to 3 Vp-p 500 ps (nominal) 150 ps (nominal)	290 ps 31 ps	
Rate: 10 Hz to 1 MHz	425 mVp-p to 575 mVp-p 25 ps (nominal)	6.8 ps	
Leveled Sine Wave 50 kHz Reference into 50 Ω load	50 kHz to 10 MHz	1.2 % of reading	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes ^{1,8} Input Impedance Measure	(10 to 40) Ω (40 to 90) Ω (90 to 150) Ω (50 to 800) k Ω (0.8 to 1.2) M Ω (1.2 to 12) M Ω	0.39 % of reading 0.083 % of reading 0.39 % of reading 0.39 % of reading 0.083 % of reading 0.39 % of reading	Comparison to Fluke 9500B Oscilloscope Calibrator, Fluke 9500B/3200 Oscilloscope Calibrator, Fluke 9530 3.2 GHz Active Head, Fluke 9550 Active Head w/ 25 ps Capability
Input Capacitance Measure	(1 to 35) pF (35 to 95) pF	1.6 % of reading + 0.19 pF 2.3 % of reading + 0.19 pF	
Rise Time – Source ^{1,8} (45 to 55) kHz	25 Vp-p (300 to 400) ps	64 ps	Comparison to Tektronix PG509 Pulse Generator
Rise Time – Measure ¹	≥ 350 ps	28 ps	Comparison to Agilent DSO6102 Oscilloscope
Bandwidth Flatness Measure ¹ into VSWR (1.2:1) (wrt Reference Frequency)	5 mVp-p to 5 Vp-p 100 Hz to 300 MHz (300 to 550) MHz 5 mVp-p to 3 Vp-p 550 MHz to 1.1 GHz (1.1 to 2.5) GHz 5 mVp-p to 2 Vp-p (2.5 to 3.2) GHz	1.6 % of reading 1.9 % of reading 2.7 % of reading 3.1 % of reading 3.1 % of reading	Comparison to Fluke 9500B/3200 Oscilloscope Calibrator, Fluke 9530 3.2 GHz Active Head

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Harmonic Distortion – Measure (dBc) into VSWR $\leq 1.3:1$	(-80 to 0) dB 9 kHz to 3 GHz (3 to 6.8) GHz (6.8 to 13.2) GHz	0.02 dB/dB + 1.1 dB 0.02 dB/dB + 3.1 dB 0.02 dB/dB + 3.7 dB	Comparison to Agilent E4405B Opt. 1DS, 1DR, 1D5 Spectrum Analyzer
Frequency Response to RF Source	100 kHz to 4.2 GHz (-30 to -20) dB (-20 to -10) dB (-10 to 0) dB (0 to 10) dB (10 to 20) dB	1.9 % of reading 1.9 % of reading 1.9 % of reading 1.9 % of reading 1.9 % of reading	Comparison to Agilent 438A Power Meter, Agilent 8482A Power Sensor

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Angle Measuring Devices (Protractors, Inclinometers, Squares, Angle Gages, etc.)	0.005 6" to 5°	3.2"	Comparison to 5 in Sine Bar, Gage Blocks
	(5 to 20)°	6.1"	
	(20 to 35)°	11"	
	(35 to 45)°	15"	
	(45 to 60)°	25"	
	90°	15"	Master Square
Calipers ^{1,5,9} (Dial, Digital, Vernier)	Up to 40 in	$(51 + 2.9L) \mu\text{in}$	Comparison to Gage Blocks
	Up to 1 000 mm	$(1.3 + 1.7L) \mu\text{m}$	
Micrometers ^{1,5,9} (Inside, Outside)	Up to 36 in	$(14 + 6.2L) \mu\text{in}$	Comparison to Gage Blocks
	Up to 900 mm	$(14 + 6.2L) \mu\text{m}$	
Micrometer Anvil Flatness ^{1,5,9} (Inside, Outside)	Flatness up to 1 in	6.4 μin	Comparison to Optical Flats
	Flatness up to 25 mm	0.16 μm	
Micrometer Anvil Parallelism ^{1,5,9} (Inside, Outside)	Parallelism to 1 in	13 μin	Comparison to Optical Parallels
	Parallelism > 1 in	$(11 + 0.64L) \mu\text{in}$	
	Parallelism to 25 mm	0.33 μm	
	Parallelism > 25 mm	$(0.27 + 0.000 65L) \mu\text{m}$	
Micrometer ^{1,5,9} (Depth, Step)	Up to 36 in	$(42 + 3.8L) \mu\text{in}$	Comparison to Gage Blocks
	Up to 900 mm	$(1.1 + 3.5L) \mu\text{m}$	
Micrometer Flatness ¹ (Depth, Step)	Up to 1.5 in	6.4 μin	Comparison to Optical Flats
	Up to 37.5 mm	0.16 μm	
Bore Gages ⁵	(0.125 to 1) in (1 to 6) in	33 μin $(26 + 7L) \mu\text{in}$	Comparison to Characterized Cylindrical Rings
Indicators ^{1,5} (Dial, Digital, Test)	Up to 1 in (1 to 6) in	$(10 + 2L) \mu\text{in}$ $(5 + 5L) \mu\text{in}$	Comparison to Gage Blocks, Surface Plates
Linear Indicators ^{5,9}	Up to 10 in	$(10 + 2.2L) \mu\text{in}$	Comparison to Universal Length Measuring Machine
	Up to 250 mm	$(0.26 + 2.2L) \mu\text{m}$	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Single Axis Length – Outside ⁵	Up to 1 in (1 to 7) in (7 to 12) in (12 to 24) in (24 to 48) in	$(6 + 1L) \mu\text{in}$ $(4 + 3.5L) \mu\text{in}$ $(4L) \mu\text{in}$ $(53 + 3L) \mu\text{in}$ $(8.2 + 1.67L) \mu\text{in}$	Comparison to Universal Length Measuring Machine P&W U304393 Digital & Laser Measuring Machine
Single Axis Length – Inside ⁵	(0.04 to 1) in (1 to 2.5) in (2.5 to 10) in (10 to 12) in	$(9 + 1L) \mu\text{in}$ $(9 + 3L) \mu\text{in}$ $(14 + 3L) \mu\text{in}$ $(25 + 3L) \mu\text{in}$	Comparison to Universal Length Measuring Machine
Chamfer/Countersink Gages ^{1,5}	Up to 0.825 in (0.825 to 1.51) in (1.51 to 2.51) in	$58 \mu\text{in}$ $(72 + 12L) \mu\text{in}$ $(80 + 23L) \mu\text{in}$	Comparison to Characterized Cylindrical Ring Gages
Height Gages ^{1,5,9}	Up to 24 in Up to 600 mm	$(21 + 3.2L) \mu\text{in}$ $(0.55 + 1.8L) \mu\text{m}$	Comparison to Gage Blocks, Grade AA Surface Plate
Height Standards, Height Masters, Risers ⁵	Up to 12 in	$(28 + 3.2L) \mu\text{in}$	Comparison to Gage Amplifier, Probe, Grade AA Surface Plate
Squareness	Up to 18 in	$9 \mu\text{in/in}$	Comparison to Master Square, Gage Amplifier, Probe, Grade AA Surface Plate
Parallelism Physical Size Up to (3 x 12) in (3 x 12) in to (3 x 18) in (6 x 18) in to (6 x 30) in	Up to 0.05 in Up to 0.05 in Up to 0.05 in	$47 \mu\text{in}$ $52 \mu\text{in}$ $42 \mu\text{in}$	Comparison to Gage Amplifier, Probe, Grade AA Surface Plate
Straightness Physical Size Up to 12 in (12 to 18) in (18 to 30) in	Up to 0.05 in Up to 0.05 in Up to 0.05 in	$35 \mu\text{in}$ $42 \mu\text{in}$ $46 \mu\text{in}$	Comparison to Gage Amplifier, Probe, Grade AA Surface Plate

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Flatness			Comparison to
Physical Size			Optical Flat
Up to 4 in Diameter	Up to 250 μin	4.4 μin	
Up to (3 x 12) in	Up to 0.05 in	49 μin	Gage Amplifier, Probe,
(3 x 12) in to (3 x 18) in	Up to 0.05 in	54 μin	Grade AA Surface Plate
(6 x 18) in to (6 x 30) in	Up to 0.05 in	54 μin	
Up to (3 x 12) in	Up to 1 in	138 μin	Multi-axis Vision System
Optical Comparator ^{1,5}			Comparison to
Length	(0 to 6) in	(185 + 13L) μin	Calibration Grids
Squareness	(0.04 to 1) in	(213 + 1L) μin	Calibration Grids
Magnification	10X to 50X	(295 + 20L) μin	Magnification Checker
Length Measuring Equipment ⁵			Comparison to
Linear Displacement	Up to 12 ft	(1 + 2.1L) μin	Laser Interferometer
Optical Reference Plane	Up to 6 in	51 μin	Comparison to
	(6 to 12) in	74 μin	Glass Scale, Calibration Grid
Thread Wires	(2 to 120) TPI (0.008 33 to 0.5) in	12 μin	Direct Measure using Measuring Machine
Cylindrical Plug Gages ⁵			Comparison to
Outside Diameter	(0.01 to 4) in	(8 + 3.3L) μin	Universal Length Measuring Machine,
	(> 4 to 10) in	(10 + 3.3L) μin	Gage Blocks
	(0.25 to 100) mm	(0.2 + 3L) μm	
	(> 100 to 250) mm	(0.24 + 3.3L) μm	
Pin Gages			Non-contact Measure using
Outside Diameter	(0.003 to 1) in	30 μin	Laser Micrometer
Cylindrical Rings ⁵			Comparison to
Inside Diameter	(0.04 to < 0.25) in	(12 + 1.3L) μin	Universal Length Measuring Machine,
	(0.25 to < 4) in	(10 + 1.3L) μin	Master Ring Gages
	(4 to 14) in	(11 + 1.3L) μin	(Non-commercial Cal)
Cylindrical Rings ⁵			Comparison to
Inside Diameter	(0.04 to < 0.25) in	(15 + 1.9L) μin	Universal Length Measuring Machine,
	(0.25 to < 4) in	(13 + 1.9L) μin	Working Reference Rings
	(4 to 14) in	(19 + 1.9L) μin	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Laser Micrometers ^{1,5}	Up to 0.1 in (0.1 to 0.4) in (0.4 to 1) in	13 µin 8 µin (12 + 5L) µin	Comparison to Characterized Master Pin Gages
Measuring Tapes, Rulers ⁵	Up to 1 ft (1 to 3) ft (3 to 1 000) ft	(760 + 0.5L) µin (740 + 2L) µin (23L) µin	Direct Measure using Vision System
Laser Distance Meters	Up to 1 m	0.31 mm	Direct Measure using Vision System
Surface Plates ^{1,4}			In accordance with ASME B89.3.7 using Electronic Level System
Overall Flatness	(17 to 168) inDL	(24 + √DL) µin	
Local Area Flatness (Repeat Readings)	Up to 0.001 in	31 µin	Repeat-o-Meter w/ Supramess
Thread Plug Gages ⁵ Pitch Diameter 60° Thread	Up to 1 in (> 1 to 4) in (> 4 to 10) in Up to 25 mm (> 25 to 100) mm (> 100 to 250) mm	(78 + 2.4L) µin (78 + 3.2L) µin (78 + 3.5L) µin (2 + 2.4L) µm (2 + 2.8L) µm (2 + 2.8L) µm	Comparison to Universal Length Measuring Machine, Master Thread Wires
Thread Plug Gages ⁵ Major Diameter	Up to 1 in (> 1 to 4) in (> 4 to 10) in Up to 25 mm (> 25 to 100) mm (> 100 to 250) mm	(2.4 + 32L) µin (3.2 + 32L) µin (3.5 + 32L) µin (2.4 + 0.78L) µm (2.8 + 0.78L) µm (2.8 + 0.78L) µm	Comparison to Universal Length Measuring Machine, Gage Blocks
Step Height	Up to 1 in	32 µin	Comparison to Gage Amplifier, Probe, Gage Blocks
Tapered Thread Plug Pitch Diameter	Up to 3 in	90 µin	Comparison to Universal Length Measuring Machine, Master Thread Wires
Standoff	Up to 1 in	31 µin	Gage Amplifier, Probe
Tapered Thread Ring Gage	Up to 3 in	90 µin	Comparison to Master Plug

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thread Ring Gage Inner Pitch Diameter	Up to 1 in (1 to 4) in (4 to 7) in	79 μ in 80 μ in 83 μ in	Comparison to Master Plug
Wire Crimpers/Dies ^{1,5} Die Diameter	(0.011 to 0.625) in	(34 + 4L) μ in	Comparison to Pin Gage Set
Crimp Height	(0.001 to 0.8) in	180 μ in	Crimp Height Micrometer

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Air Velocity Pitot Tube, Constant Temperature Anemometers	(0.3 to 2) m/s (2 to 30) m/s	0.76 % of reading + 0.02 m/s 0.6 % of reading	Comparison to Laser Doppler Velocimeter, Westenberg Westi-box Wind Tunnel System
	(0.3 to 2) m/s (2 to 60) m/s	0.06 m/s 1.4 % of reading	Westenberg Westi-box Wind Tunnel System
Air Velocity Vane Anemometers	(0.3 to 2) m/s (2 to 30) m/s (30 to 60) m/s	0.87 % of reading + 0.04 m/s 0.95 % of reading 2.4 % of reading	Comparison to Laser Doppler Velocimeter, Westenberg Westi-box Wind Tunnel System
Shore Hardness Test Blocks Type A	(15 to 95) duro	1.8 duro	Comparison to Type A Digital Durometer
Type D	(15 to 95) duro	1.6 duro	Type D Digital Durometer

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Durometers (Type A, B, C, D, E, DO, O) Indenter Dimensions Length Angle Radius Spring Force	Up to 1 in Up to 40° 50 µin to 1 in Up to 60 duro (60 to 100) duro	84 µin 0.006° 120 µin 0.73 duro 0.94 duro	Direct Verification per ASTM D2240 using Multi-axis Vision System Duro Calibrator
Balances and Scales ^{1,6} (SI)	Up to 500 mg (0.5 to 5) g (5 to 20) g 20 g to 10 kg (10 to 25) kg (25 to 35) kg	12 µg 40 µg 90 µg 0.000 31 % of reading 0.000 3 % of reading 0.000 25 % of reading	ASTM E617 Class 1 weights and internal calibration procedure utilized for the calibration of the weighing system.
Balances and Scales ^{1,6} (SI)	Up to 3 mg (3 to 10) mg (10 to 50) mg (50 to 100) mg (100 to 300) mg	30 µg 40 µg 50 µg 60 µg 70 µg	ASTM E617 Class 3 weights and internal calibration procedure utilized for the calibration of the weighing system.
Balances and Scales ^{1,6} (SI)	(300 to 500) mg (0.5 to 3) g (3 to 5) g (5 to 10) g (10 to 30) g (30 to 50) g (50 to 100) g	0.1 mg 0.15 mg 0.2 mg 0.3 mg 0.5 mg 0.7 mg 1.2 mg	ASTM E617 Class 3 weights and internal calibration procedure utilized for the calibration of the weighing system.
Balances and Scales ^{1,6} (Avoirdupois)	Up to 0.5 lb (0.5 to 2) lb (2 to 10) lb (10 to 1 150) lb	0.024 % of reading 0.018 % of reading 0.013 % of reading 0.012 % of reading	NIST Class F weights and internal calibration procedure utilized for the calibration of the weighing system.
Mass Determination	Up to 1.2 kg	0.002 8 mg/g + 6.8 mg	Direct Measure using Mettler Toledo XS1203S Precision Balance
Force (Tension and Compression)	Up to 400 lbf	0.012 % of reading	Comparison to Deadweight

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Force			
Compression	(2.5 to 100) lbf (100 to 500) lbf (500 to 2 000) lbf (10 000 to 25 000) lbf (25 000 to 100 000) lbf	0.03 % of reading + 0.02 lbf 0.02 % of reading + 0.05 lbf 0.02 % of reading + 0.15 lbf 0.04 % of reading + 1.2 lbf 0.04 % of reading + 9 lbf	Comparison to Morehouse PCM/UCM with Precision Load Cells
Tension	(2.5 to 100) lbf (100 to 500) lbf (500 to 2 000) lbf (10 000 to 25 000) lbf (25 000 to 100 000) lbf	0.03 % of reading + 0.03 lbf 0.02 % of reading + 0.06 lbf 0.02 % of reading + 0.19 lbf 0.04 % of reading + 1.2 lbf 0.04 % of reading + 9 lbf	
Pneumatic Absolute Pressure Devices	Up to 14.7 psia (14.7 to 39.7) psia (39.7 to 514.7) psia	0.002 5 psi 0.000 7 % of reading + 0.002 3 psi 0.006 5 % of reading	Comparison to Ruska 7250xi Pressure Controller/Calibrator
Pneumatic Absolute Pressure Devices	Up to 30 psia (30 to 1 000) psia	0.002 4 psi 0.007 % of reading + 0.000 48 psi	Comparison to DHI PPC4 Pressure Controller
Hydraulic Absolute Pressure Devices	(500 to 1 600) psia (> 1 600 to 16 000) psia	0.011 % of reading + 0.07 psi 0.015 % of reading + 0.01 psi	Comparison to Fluke P3125-3 Deadweight Tester, Barometer
Pneumatic Gauge Pressure Devices ¹	(0 to 25) psig (25 to 500) psig	0.001 4 % of reading + 0.001 3 psi 0.006 7 % of reading	Comparison to Ruska 7250xi Pressure Controller/Calibrator
Pneumatic Gauge Pressure Devices ¹	(500 to 1 000) psig	0.006 2 % of reading + 0.000 1 psi	Comparison to DHI PPC4 Pressure Controller
Pneumatic Vacuum Devices ¹	(-14.7 to 0) psig	0.000 64 % of reading + 0.001 3 psi	Comparison to Ruska 7250xi Pressure Controller/Calibrator
Pneumatic Compound Devices ¹	(-60 to -6) inH ₂ O (-6 to 6) inH ₂ O (6 to 60) inH ₂ O (-14.7 to -7.5) psi (-7.5 to 7.5) psi (7.5 to 30) psi	0.007 % of reading + 0.000 005 inH ₂ O 0.002 1 % of reading + 0.000 3 inH ₂ O 0.007 % of reading + 0.000 005 inH ₂ O 0.006 5 % of reading 0.001 5 % of reading + 0.000 38 psi 0.006 5 % of reading	Comparison to Ruska 7252i Pressure Controller/Calibrator

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Hydraulic Gauge Pressure Devices ¹	(500 to 1 600) psig (> 1 600 to 16 000) psig	0.1 psi 0.006 1 % of reading	Comparison to Fluke P3125-3 Deadweight Tester
Torque Tools ¹	(1 to 10) lbf·in 9 lbf·in to 800 lbf·ft	0.37 % of reading + 0.028 lbf·in 1 % of reading	Comparison to Torque Calibration System
Pneumatic Torque Tools	0.2 lbf·in to 18 lbf·in 18 lbf·in to 70 lbf·in 8 lbf·in to 130 lbf·in	0.55 lbf·in 1.1 lbf·in 0.89 lbf·in	Comparison to Imada Torque Tester
Torque Transducers	2.5 ozf·in to 12.5 lbf·in (12 to 150) lbf·in (12.5 to 250) lbf·ft (250 to 2 000) lbf·ft	0.085 % of reading 0.052 % of reading 0.048 % of reading 0.053 % of reading	Comparison to Master Weights, Butterfly Wheels, 40 in Torque Arm
Torque Angle (Fixed Points)	45° and 225° 90° and 270° 135° and 315° 180° and 360°	0.67° 0.67° 0.67° 0.67°	Comparison to Torque Angle Fixture

Thermodynamics

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity – Source	(15.56 to 25) °F 0 %RH	0.003 %RH	Comparison to Thunder Scientific 3900 Humidity Generator, Liquid Nitrogen (N ₂), Vaisala DMT-152 Dew Point Transmitter
Humidity – Source	(-10 to 15) °C (10 to 75) %RH (75 to 95) %RH (15 to 35) °C (10 to 95) %RH (35 to 70) °C (10 to 50) %RH (50 to 70) %RH (70 to 95) %RH	0.5 %RH 0.65 %RH 0.5 %RH 0.5 %RH 0.7 %RH 0.85 %RH	Comparison to Thunder Scientific 2500 Humidity Generator

Thermodynamics

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity – Measure ¹	(-20 to 15) °C Up to 95 %RH (15 to 25) °C (10 to 90) %RH (90 to 95) %RH (25 to 40) °C Up to 95 %RH	0.008 4 % of reading + 1.2 %RH 1.3 %RH 1.9 %RH 0.008 4 % of reading + 1.2 %RH	Direct Measure using Vaisala MI70/HMP76B Temperature/Humidity Indicator/Probe
Dew/Frost Point – Source ¹	(-85 to -80) °C (-80 to -70) °C (-70 to -60) °C (-60 to -50) °C (-50 to -40) °C (-40 to -10) °C	1.2 °C 0.61 °C 0.4 °C 0.36 °C 0.25 °C 0.15 °C	Comparison to Thunder Scientific 3900 Humidity Generator
Dew/Frost Point – Source ¹	(-10 to 20) °C (20 to 30) °C (30 to 40) °C (40 to 65) °C	0.13 °C 0.15 °C 0.16 °C 0.17 °C	Comparison to Thunder Scientific 2500 Humidity Generator
Temperature – Source ¹ (Temperature Measuring Devices)	(-75 to -70) °C (-70 to 100) °C (100 to 200) °C (200 to 400) °C (400 to 600) °C	0.039 °C 0.039 °C 0.051 °C 0.074 °C 0.092 °C	Comparison to AccuMac AM1760 SPRT, Black Stack, Hart Bath or Metrology Well
Temperature – Source ¹ (Temperature Measuring Devices)	(600 to 800) °C (800 to 1 000) °C (1 000 to 1 200) °C	1.2 °C 1.5 °C 3.6 °C	Comparison to Fluke 5649 Type R Thermocouple Probe, HP 3458 8.5 Digit Multimeter, Furnace
Temperature – Measure ¹	(-195 to 0) °C (0 to 420) °C (420 to 660) °C	0.015 °C 0.03 °C 0.039 °C	Direct Measure using AccuMac AM1760 SPRT, Black Stack
Temperature – Measure ¹	(600 to 800) °C (800 to 1 000) °C (1 000 to 1 450) °C	0.55 °C 0.76 °C 2.9 °C	Comparison to Fluke 5649 Type R Thermocouple Probe, HP 3458 8.5 Digit Multimeter

Thermodynamics

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Infrared Temperature – Source (Infrared Temperature Measuring Instruments)	(-15 to 0) °C (0 to 50) °C (50 to 100) °C (100 to 120) °C (120 to 200) °C (200 to 350) °C (350 to 500) °C (500 to 600) °C (600 to 700) °C (700 to 800) °C (800 to 900) °C (900 to 1 000) °C	0.44 °C 0.44 °C 0.48 °C 0.59 °C 0.64 °C 1.1 °C 1.8 °C 4.4 °C 4.7 °C 5.1 °C 5.4 °C 5.8 °C	Comparison to Blackbody Source (Flat Plate) $\lambda = (8 \text{ to } 14) \mu\text{m}$, $\varepsilon = (0.9 \text{ to } 1)$
Surface Probe Measuring Equipment	(20 to 100) °C	0.68 °C	Comparison to Fluke 3125 Surface Probe Calibrator
Surface Probe Measuring Equipment	(100 to 300) °C	1.2 °C	Comparison to Pyromation RAT185 RTD, Fluke 1502 Thermometer Readout
Centrifuge Chamber ¹ Temperature	(-40 to 100) °C	1.2 °C	Comparison to Altek 422 Thermocouple Calibrator, Type T Thermocouple Probe

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Reference	10 MHz	1.6 pHz/Hz	Comparison to Pendulum FTR-210R GPS Disciplined Oscillator
Period – Source	(1 to 100) s	58 ns/s	Comparison to Keysight 33250A Function/Arbitrary Waveform Generator, Pendulum FTR-210R GPS Disciplined Oscillator

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Period – Measure	(1 to 10) s (10 to 100) s	17 μ s 53 μ s	Comparison to HP 53132A Universal Counter, Pendulum FTR-210R GPS Disciplined Oscillator
Frequency – Source	1 Hz to 80 MHz	58 nHz/Hz	Comparison to Keysight 33250A Function/Arbitrary Waveform Generator, Pendulum FTR-210R GPS Disciplined Oscillator
Frequency – Measure	1 Hz to 1 kHz 1 kHz to 10 MHz (10 to 225) MHz	1.5 nHz/Hz + 5.5 μ Hz 1.3 nHz/Hz + 4 μ Hz 1.2 nHz/Hz	Comparison to HP 53132A Universal Counter, Pendulum FTR-210R GPS Disciplined Oscillator
AC Duty Cycle – Source ¹ Square Wave: < 3.3 Vp-p Freq: 0.1 Hz to 100 kHz	(1 to 10) % Duty Cycle 10 μ s to 100 s (10 to 49) % Duty Cycle 10 μ s to 100 s 50 % Duty Cycle 10 μ s to 100 s (51 to 90) % Duty Cycle 10 μ s to 100 s (90 to 99) % Duty Cycle 10 μ s to 100 s	0.039 % of reading + 78 ns 0.62 % of reading + 78 ns 0.001 6 % of reading + 78 ns 0.62 % of reading + 78 ns 0.039 % of reading + 78 ns	Comparison to Fluke 5522A Multiproduct Calibrator
Stopwatches/Timers	Up to 19.99 s/d	59 ms/d	Direct Measure using Vibrograf 4500 Timometer
Time – Measure ¹	Up to 24 hr	1.1 s/d + 0.2 s	Comparison to Stopwatch
Non-Contact Rate of Rotation (Including Centrifuges) ^{1,2}	(5 to 99.999) rpm (100 to 999.99) rpm (1 000 to 9 999.9) rpm (10 000 to 99 999) rpm (100 000 to 200 000) rpm	0.012 % of reading + 0.001 2 rpm 0.012 % of reading + 0.012 rpm 0.012 % of reading + 0.12 rpm 0.014 % of reading + 1.2 rpm 0.014 % of reading + 12 rpm	Comparison to Optical Tachometer

DIMENSIONAL MEASUREMENT

1 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Geometric Measurement of Fixtures, Gauges, Dies and Molds ⁵ (Length – Single Axis) Outside Dimensions	Up to 1 in (1 to 7) in (7 to 12) in	$(6 + 1L) \mu\text{in}$ $(4 + 3.5L) \mu\text{in}$ $(4L) \mu\text{in}$	Universal Length Measuring Machine utilized as the reference standard for 1-D Length Measurements.
Inside Dimensions	(0.04 to 0.125) in (0.125 to 0.25) in (0.25 to 1) in (1 to 2.5) in (2.5 to 10) in (10 to 14) in	11 μin 11 μin 11 μin 17 μin $(18 + 3L) \mu\text{in}$ $(38 + 3L) \mu\text{in}$	

2 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Geometric Measurement of Fixtures, Gauges, Dies and Molds ⁵ X-Y Axis	Up to 16 in	$(97 + 5.8L) \mu\text{in}$	Multi-axis Vision System utilized as the reference standard for 2-D Length Measurements.
Angles ⁵	Up to 360° Up to 12 in	$(\frac{20}{L(\text{in})} + 1.2)''$	Multi-axis Vision System utilized as the reference standard for Angle Measurements.
Radius ⁵	Up to 8 in	$(97 + 7L) \mu\text{in}$	Multi-axis Vision System utilized as the reference standard for Radius Measurements.


3 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional Measurement – 3D ⁵ (Volumetric)	X = Up to 12 in Y = Up to 12 in Z = Up to 10 in	(191 + 5.8L) μin	Multi-axis Vision System utilized as the reference standard for 3-D Measurements.

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

- On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
- As frequency deviates from the listed values, uncertainty may be higher than stated. If needed, contact the laboratory for more information regarding uncertainties at frequency and range combinations other than the ones shown.
- Carrier range is the output amplitude capability of the fundamental wave. Up to 10 harmonics can be placed on the fundamental wave up to the amplitude of the value shown as the harmonic range. The uncertainties shown are for both the fundamental and harmonic amplitudes within the frequency range shown.
- The uncertainties shown are for the most favorable conditions. There is an increase in uncertainty that corresponds to the laboratory's AC voltage and current uncertainties at different frequencies other than the ones shown. Power factors (PF) other than the one shown contribute to the power uncertainty. PF is related to the cosine of phase. Therefore, uncertainties track the laboratory's phase uncertainty closely at PF near one but are magnified heavily as PF approaches zero. The lab may also report reactive power, apparent power, and power factor under this accreditation. If needed, contact the laboratory for more information regarding uncertainties at frequency and power factor combinations other than the ones shown.
- " = arc-second; L = length in inches or millimeter; DL = diagonal length in inches; rpm = revolutions per minute.
- The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The CMC presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
- Uncertainties are for cardinal point temperatures, measurement between cardinal temperate points are available with greater uncertainties.
- The stated uncertainty is the laboratory's ability to source a fast rise pulse that is approximately 500 ps, 350 ps, 125 ps, and 25 ps. In the typical application of measuring rise time of an oscilloscope, this value is one of the contributing factors, but other factors are derived from the DUT. The known source rise time is mathematically removed from the total measured rise time measured on the DUT.
- 0.6R will be added to the Measurement Uncertainty at the time of calibration (where R = resolution of the device under calibration).
- Unless otherwise specified in the far-right hand column, the calibration procedure/method being utilized by the laboratory was written and validated internally.



Jason Stine, Vice President