

Pneumatic Absolute Pressure Devices ¹	(0 to 14.7) psia (14.7 to 19.65) psia (19.65 to 31.2) psia (31.2 to 48) psia (48 to 62.55) psia (62.55 to 145) psia (145 to 179.7) psia (179.7 to 514.7) psia (514.7 to 1014.7) psia	0.0024 psi 0.0021 psi 0.0024 psi 0.0082% + 0.0020 psi 0.0059 psi 0.0082% + 0.0021 psi 0.014 psi 0.0082% + 0.0021 psi 0.0082% + 0.0061 psi	Comparison to Wika Pressure Controller
Pneumatic Absolute Pressure Devices ¹	(0 to 1.014) bara (1.014 to 1.355) bara (1.355 to 2.151) bara (2.151 to 3.309) bara (3.309 to 4.313) bara (4.313 to 9.997) bara (9.997 to 12.39) bara (12.39 to 35.487) bara (35.487 to 69.961) bara	0.00017 bar 0.00014 bar 0.00017 bar 0.0082% + 0.00014 bar 0.00041 bar 0.0082% + 0.00014 bar 0.00097 bar 0.0082% + 0.00014 bar 0.0082% + 0.00042 bar	Comparison to Wika Pressure Controller
Pneumatic Pressure Devices ¹	(-14.7 to 0) psig (0 to 1) psig (1 to 4.95) psig (4.95 to 16.5) psig (16.5 to 145) psig (145 to 165) psig (165 to 500) psig (500 to 1000) psig	0.0012 psi 0.000090 psi 0.00057 psi 0.0082% + 0.000058 psi 0.0082% + 0.00058 psi 0.014 psi 0.0082% + 0.00058 psi 0.0082% + 0.0058 psi	Comparison to Wika Pressure Controller
Pneumatic Pressure Devices ¹	(-1 to 0) bar (0 to 0.069) bar (1 to 0.341) bar (4.95 to 1.137) bar (1.137 to 9.997) bar (9.997 to 11.376) bar (11.376 to 34.473) bar (34.473 to 68.947) bar	0.083 mbar 0.0062 mbar 0.039 mbar 0.0082% + 0.0040 mbar 0.0082% + 0.040 mbar 0.97 mbar 0.0082% + 0.040 mbar 0.0082% + 0.040 mbar	Comparison to Wika Pressure Controller
Hydraulic Pressure Devices ¹	(0.1 to 1.2) MPa (1.2 to 3.5) MPa	0.0001 MPa 0.0084% of reading	Comparison to Hydraulic Low Piston Fluke Deadweight Tester
Hydraulic Pressure Devices ¹	(2 to 12) MPa (12 to 120) MPa	0.001 MPa 0.0083% of reading	Comparison to Hydraulic High Piston Fluke Deadweight Tester